

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 200

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	2	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WDXDH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	2	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXDH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	2	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPERATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 200

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	2	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WXDXH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	2	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WXDXH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		093154 TBO01	2	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25"; LOW PROFILE; TUNING FORK BASE DESIGN; SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 200

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1004559	4	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; NOTES: 2X PER BED	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		1004558	2	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1014992	2	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	2	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 200

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE


END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, SEMI-PRIVATE PAT ROOM
ROOM NUMBER: 201

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 202

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	2	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WXDXH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	2	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	2	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WXDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 202

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	2	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	2	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		093154 TBO01	2	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25"; LOW PROFILE; TUNING FORK BASE DESIGN; SGL THERMOFOIL TOP TO INCLUDE: (WDXDH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 202

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1004559	4	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; NOTES: 2X PER BED	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		1004558	2	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1014992	2	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	2	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, SEMI-PRIVATE
ROOM NUMBER: 202

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE


END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, SEMI-PRIVATE PAT ROOM
ROOM NUMBER: 203

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 204

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	1	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WXDXH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WXDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 204

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WXDXH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WXDXH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		093154 TBO01	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25"; LOW PROFILE; TUNING FORK BASE DESIGN; SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 204

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WXDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 204

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE


END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, PRIVATE
ROOM NUMBER: 205

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 206

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	1	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WDXDH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXDH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 206

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		093154 TBO01	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25"; LOW PROFILE; TUNING FORK BASE DESIGN; SGL THERMOFOIL TOP TO INCLUDE: (WDXDH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 206

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WXDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 206

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE


END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, PRIVATE
ROOM NUMBER: 207

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PAIENT ROOM, PRIVATE
ROOM NUMBER: 208

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	1	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WXDXH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WXDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPERATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 208

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		093154 TBO01	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25"; LOW PROFILE; TUNING FORK BASE DESIGN; SGL THERMOFOIL TOP TO INCLUDE: (WDXDH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, PRIVATE
ROOM NUMBER: 208

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WXDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PAITENT ROOM, PRIVATE
ROOM NUMBER: 208

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, PRIVATE
ROOM NUMBER: 209

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, ACCESSIBLE ISOLATION
ROOM NUMBER: 210

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1006009 BED01	1	BED, MED-SURG 14.5 MINIMUM HEIGHT;500# CAP;WI-FI;SMART BED READY;36" WIDE FRAME;VERBAL ALERTS;IN-BED SCALE;5TH WHEEL STEERING TO (WDXDH) IN: 40.00 x 98.70 x 32.50 LBS: 563.00 MM: 1016.00 x 2506.98 x 825.50 KG: 255.38 LOCATION: FLOOR NOTES: DISTRICT WILL HAVE 10 NEW BEDS BY TIME OF OPENING; REUSE AS EXISTING DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 10 WATS: 1200 ELECT NOTE: 10A	HILL-ROM CENTRELLA SMART +	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXDH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, ACCESSIBLE ISOLATION
ROOM NUMBER: 210

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WXDXH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5	No image available	1015783	1	ALLOWANCE, INSTALLATION INSTALLATION FOR PATIENT MONITORING SYSTEMS; TO INCLUDE: (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025.	PHILIPS MEDICAL SYSTEMS 866389 PATIENT INFORMATION CENTER IX	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
6		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WXDXH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, ACCESSIBLE ISOLATION
ROOM NUMBER: 210

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		093154 TBO01	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25";LOW PROFILE;TUNING FORK BASE DESIGN;SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; NOTES: 2X PER BED	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
9		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT ROOM, ACCESSIBLE ISOLATION
ROOM NUMBER: 210

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ANTEROOM
ROOM NUMBER: 211

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1013193 PPR01	8	RESPIRATOR, POWERED AIR PURIFYING (PAPR) POWERED AIR PURIFYING RESPIRATOR (PAPR). FEATURES: COMPLETE PAPR SYSTEM FOR USE IN RUGGED ENVIRONMENTS SUCH AS INDUSTRIAL, (WXDXH) IN: 8.00 x 3.50 x 8.25 LBS: 3.00 MM: 203.20 x 88.90 x 209.55 KG: 1.37 LOCATION: SPACE SPEC COMMENTS: DIMENSIONS AND WEIGHT (WITH BATTERY) REFLECTS VERSAFLO UNIT ONLY. SEE DATASHEET FOR OTHER COMPONENTS WEIGHT. NOTES: STORED WITHIN CABINET VOLTS: 120 HERTZ: 60 AMPS: 1.5 WATS: 180 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: ELECTRICAL REFLECTS SINGLE STATION BATTERY CHARGER.	3M HEALTH CARE TR-600-HIK	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: TOILET, PRIVATE ISOLATION
ROOM NUMBER: 212

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: SOILED HOLDING ROOM
ROOM NUMBER: 213

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100858 TRK01	1	TRUCK/BIN, LINEN 500LB; 16 CUFT CAPACITY; MANUAL UNLOADING; BLUE (WDXH) IN: 30.38 x 44.00 x 35.50 LBS: 60.50 MM: 771.66 x 1117.60 x 901.70 KG: 27.45 LOCATION: FLOOR NOTES: INCLUDES LID	TOTER INC. MMC16-90701 CUBE HEAVY DUTY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		100859 WST02	1	CONTAINER, BIOHAZARD RED; 64 GALLON; 2 WHEEL; HINGED LID; 2-10"WHEELS; 2-4"CASTERS; PLASTIC (WDXH) IN: 24.25 x 32.00 x 44.25 LBS: 45.30 MM: 615.95 x 812.80 x 1123.95 KG: 20.55 LOCATION: FLOOR	TOTER INC. RMC64-00RED RMC64	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1015414 TRK02	1	TRUCK, TILT, WASTE/LINEN 16 CUFT MANUAL CUBE TRUCK; INCLUDES LID (WDXH) IN: 30.40 x 44.00 x 35.50 LBS: 60.00 MM: 772.16 x 1117.60 x 901.70 KG: 27.22 LOCATION: FLOOR NOTES: BLACK IN COLOR; INCLUDES LID	TOTER INC. MMC16-10836 BLACK	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ED EQUIPMENT STORAGE
ROOM NUMBER: 214

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		107260 SCP03	1	SCOPE, LARYNGOSCOPE 7" SCREEN SIZE;FOR DIFFICULT INTUBATION. (WXDXH) IN: 23.00 x 23.00 x 55.00 LBS: 24.00 MM: 584.20 x 584.20 x 1397.00 KG: 10.89 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY UNIT WITH ACCESSORIES. MAKE/MODEL MAY CHANGE DURING PROCUREMENT DATA NOTE: HDMI;USB 2.0;VIDEO OUTPUT TO EXTERNAL MONITOR VOLTS: 120 HERTZ: 60 AMPS: .5 ELECT NOTE: POWER CORD ADAPTER;RECHARGEABLE LI-ION BATTERIES	KARL STORZ ENDOSCOPY-AMERICA, INC. C-MAC/OPTIONS 8403ZXK	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		1005310 ILL01	1	ILLUMINATOR, VEIN VEIN ILLUMINATOR W/ STAND; LED DISPLAY; TIME OUT SET, REMEMBERS LAST SETTING; NORMAL/INVERSE; SET AUTO TIMEOUT; (WXDXH) IN: 22.00 x 22.00 x 39.00 LBS: 27.00 MM: 558.80 x 558.80 x 990.60 KG: 12.25 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 AMPS: 0.4 ELECT NOTE: RECHARGEABLE INTEGRATED 3.6V, 3100MAH LI-ION BATTERY	ACCUVEIN AV500/HF580 AV500/HF580	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
3		058850 CRC01	1	CART, CAST STAINLESS;8 DRAWER;S.S. FOLDING SIDE SHELF; DRAWERS: (1) 3"W/BLACK VINYL MAT, (5) 3"W/ 4 ADJ DIVIDERS PER DRAWER, (2) (WXDXH) IN: 38.00 x 21.00 x 77.00 LBS: 202.00 MM: 965.20 x 533.40 x 1955.80 KG: 91.63 LOCATION: FLOOR	HARLOFF COMPANY, INC. DELUXE 6035	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1005648 LMS01	1	LAMP, SLIT LED; DIGITAL CAMERA; ELECTRIC AUTO TABLE TOP; IMAGENET 6 TO INCLUDE: (WXDXH) IN: 21.00 x 19.00 x 60.86 LBS: 78.83 MM: 533.40 x 482.60 x 1545.85 KG: 35.76 LOCATION: FLOOR NOTES: INCLUDES MOBILE TABLE VOLTS: 120 HERTZ: 60 WATS: 110 ELECT NOTE: ELECTRICAL SHOWN FOR SLIT LAMP ELECTRICAL FOR TABLE: 120V;60HZ;220VA	TOPCON INSTRUMENT CORP. OF AMERICA SL-D701	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ED EQUIPMENT STORAGE
ROOM NUMBER: 214

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		071749 RCK03	1	RACK, CRUTCH FOR CRUTCHES,CANES,WALKERS & ROLLS;ALL STEEL GREY POWDER COAT (WDXH) IN: 33.00 x 12.00 x 4.00 LBS: 13.00 MM: 838.20 x 304.80 x 101.60 KG: 5.90 LOCATION: WALL	IDEAL RR31	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		1002317	2	WALKER ALUMINUM FOLDING WALKER WITH WHEELS; PUSH-BUTTON, DUAL- RELEASE MECHANISMS FOR EASY-FOLDING; 350 LB WEIGHT CAPACITY (WDXH) IN: 24.00 x 17.50 x 39.00 LBS: 12.00 MM: 609.60 x 444.50 x 990.60 KG: 5.45 LOCATION: FLOOR	MCKESSON 1073638 146-10210-4	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1004200 WRMI01	1	WARMER, INFANT ELEVATING BASE;PATIENT PROBE;IN-BED SCALE;SINGLE DRAWER;INSTRUMENT SHELF 12"X12" TO INCLUDE: (WDXH) IN: 37.00 x 47.00 x 86.00 LBS: 220.00 MM: 939.80 x 1193.80 x 2184.40 KG: 99.80 LOCATION: FLOOR DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RS-232 PORT VOLTS: 115 HERTZ: 60 AMPS: 6.6 WATS: 520	GE HEALTHCARE M1112198 PANDA IRES	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
8		084955 TCRT01	1	CART, TELEMEDICINE MOBILE TELE-COMMUNICATION SYSTEM;FULLY INTEGRATED 24" HD LCD DISPLAY;INTEGRATED CAMERA & MICROPHONE (WDXH) IN: 23.00 x 27.00 x 71.00 LBS: 230.00 MM: 584.20 x 685.80 x 1803.40 KG: 104.33 LOCATION: FLOOR NOTES: NEED TO CONFIRM EXISTING MAKE/MODEL DATA HARDWIRED: YES WIRELESS: NO VOLTS: 125 HERTZ: 60 ELECT NOTE: 45 AH LITHIUM IRON BATTERY RECHARGED BY HOSPITAL GRADE PLUG ONLY	AVIZIA - CA750	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
9		059223	1	PUMP, COMPRESSION LEG & FOOT COMPRESSION; COMPRESSION SLEEVE CONTAINS 3 BLADDERS TO DELIVER CIRCUMFERENTIAL COMPRESSION ALL-IN-ONE (WDXH) IN: 7.70 x 7.30 x 6.80 LBS: 5.00 MM: 195.58 x 185.42 x 172.72 KG: 2.27 LOCATION: COUNTERTOP	KENDALL HEALTHCARE, COVIDIEN 29525 700 SERIES	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ED EQUIPMENT STORAGE
ROOM NUMBER: 214

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 WATS: 50 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 50VA;13' POWER CORD WITH HOSPITAL GRADE, REGION-SPECIFIC CORDAGE/PLUG OUTLET MUST BE GROUNDED RECHARGEABLE LITHIUM ION BATTERY 10.8V, 2200MAH				
10		106100 IUS01	1	IMAGING, ULTRASOUND SCANNER 0-99ML BLADDER VOLUME RANGE;3D ULTRASOUND;LCD COLOR DISPLAY;HAND-HELD PROBE;2 LITHIUM ION BATTERIES;BATTERY (WXDXH) IN: 21.00 x 21.00 x 52.00 LBS: 15.00 MM: 533.40 x 533.40 x 1320.80 KG: 6.81 LOCATION: FLOOR SPEC COMMENTS: ; CHANGE CODE: C CHANGE NOTES: 2025.05.12 MOVE FROM RM 224 TO RM 214 PER SHD REVI DATA NOTE: MICRO USB;USB PORT;SD CARD PORT VOLTS: 120 HERTZ: 60 AMPS: 1.4 ELECT NOTE: ELECTRICAL SHOWN IS FOR THE CHARGING UNIT BATTERY OPERATED POWER ADAPTER: 24V DC; 0-2.71A OUTPUT	VERATHON MEDICAL 0270-0870 BLADDERSCAN PRIME PLUS	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
11		1023026 DUV01*	1	DISINFECTOR, ROOM, MOBILE TELESCOPING ROOM STERILIZATION UNIT; 55W UV-C LAMPS; (WXDXH) IN: 16.00 x 13.50 x 46.75 LBS: 44.00 MM: 406.40 x 342.90 x 1187.45 KG: 19.96 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 110 HERTZ: 60	FIRST RESPONDER TECHNOLOGIES STERIZAP UV-C ST E500UVST	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: MULTI-PURPOSE ROOM
ROOM NUMBER: 215

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: CLEAN SUPPLY ROOM
ROOM NUMBER: 216

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1001188 CBW01	1	CABINET, WARMING 2.5 CUBIC FT; 6 BLANKET CAPACITY; COUNTERTOP MODEL; STACKABLE; NO (WXDXH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP SPEC COMMENTS: CLEARANCES: TOP/LEFT/RIGHT: 1", BACK 3" NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO LIVES ON SHELF IN ROOM ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 5 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 0.6KW;15A;125V PLUG;HOSPITAL GRADE PLUG	ENTHERMICS MEDICAL SYSTEMS TITAN SERIES EC250	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		105040 CBW02	1	CABINET, WARMING, WIPES FOR USE WITH COMFORT PERSONAL CLEANSING BRAND;ACCOMMODATES 12 INDIVIDUAL PACKAGES;125 DEGREES F TARGET PRODUCT (WXDXH) IN: 14.69 x 15.25 x 16.88 LBS: 30.00 MM: 373.13 x 387.35 x 428.76 KG: 13.61 LOCATION: COUNTERTOP ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 7 DEDICATED CIRCUIT: YES ELECT NOTE: SINGLE PHASE;DEDICATED CIRCUIT RECOMMENDED	SAGE PRODUCTS INC. 7937 12-SLOT	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015649 CTSS01*	1	CART, SUPPLY, STAINLESS, 36 INCH TWO SOLID STAINLESS STEEL SHELVES (WXDXH) IN: 24.00 x 36.00 x 27.00 LBS: 60.00 MM: 609.60 x 914.40 x 685.80 KG: 27.22 LOCATION: FLOOR NOTES: 2024.09.24 CART UNDER 31"H TO FIT UNDER SHELF CHANGE CODE: C CHANGE NOTES: SPEC UPDATED	INTERMETRO INDUSTRIES CORP 2436FNS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: CLEAN SUPPLY ROOM
ROOM NUMBER: 216

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		066552 PAN01	4	LOUVER PANEL, WALL MOUNTED STORAGE PANEL; 16 GAUGE STEEL W/ GRAY ENAMEL FINISH LOUVERED FOR BINS; 250LB LOAD CAPACITY ONCE FASTENED SECURELY TO (WXDXH) IN: 48.00 x 1.00 x 19.00 LBS: 18.00 MM: 1219.20 x 25.40 x 482.60 KG: 8.17 LOCATION: WALL ANCHOR DETAIL REQ: YES	QUANTUM STORAGE SYSTEMS QLP-4819 ULTRA HANGING SYSTEMS	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5	No image available	506580	1	ALLOWANCE, BIN (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		065434 SHW03	2	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WXDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
7		084882 CRK02	2	CART, CYLINDER, MEDICAL GAS 4 CAPACITY TILT BACK CART; E/D SIZE; W/HEAVY DUTY 5" CASTERS (WXDXH) IN: 11.00 x 11.00 x 39.00 LBS: 21.00 MM: 279.40 x 279.40 x 990.60 KG: 9.53 LOCATION: FLOOR NOTES: QTY 1 RED; QTY 1 GREEN	W.T. FARLEY, INC. - CR-DC04E	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		1015713	1	CART, SUPPLY THREE WIRE ADJ HT "TUB" SHELVES (TOP SHELF 5" DEEP, BOTTOM SHELVES 1" DEEP); CHROME PLATED STEEL;3" CASTERS;PUSH HANDLE (WXDXH) IN: 30.00 x 18.00 x 30.00 LBS: 50.00 MM: 762.00 x 457.20 x 762.00 KG: 22.68 LOCATION: FLOOR	LUXOR CORPORATION LICWT2918 LARGE WIRE TUB	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: SUPERVISOR OFFICE
ROOM NUMBER: 217

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015782 MTCR02	1	MONITOR, CENTRAL STATION, REMOTE MIRROR STATION; 23" DISPLAY; 120V AC (WXDXH) IN: 17.00 x 8.20 x 16.00 LBS: 30.00 MM: 431.80 x 208.28 x 406.40 KG: 13.61 LOCATION: COUNTERTOP SPEC COMMENTS: OVERALL DIMENSIONS INCLUDE CPU, UPS, AND RECORDER/ALARM SPEAKER MODULE SEE CUT SHEET FOR SPECIFIC DIMENSIONS. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: MINIMUM ETHERNET 10/100 T, CAT 5E OR HIGHER VOLTS: 115 HERTZ: 60 AMPS: 10.8 DEDICATED CIRCUIT: YES ELECT NOTE: REQUIRES A 15AMP QUAD OUTLET WITHIN 3' OF UPS/CPU.	PHILIPS MEDICAL SYSTEMS MXU0295 DISPLAY: 23 BTU: 1595 BTU/HR	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A MECH NOTE: REFRIGERANT: R600A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
3		055770 OVM01	1	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WXDXH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: SUPERVISOR OFFICE
ROOM NUMBER: 217

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1015592	1	SHREDDER, PAPER SHREDS 10 SHEETS PER PASS INTO 5/23" X 1 9/16" CROSS-CUT PARTICLES; SAFELY SHREDS STAPLES, PAPER CLIPS, AND PLASTIC CREDIT (WXDXH) IN: 9.19 x 14.50 x 15.88 LBS: 15.00 MM: 233.43 x 368.30 x 403.36 KG: 6.81 LOCATION: FLOOR VOLTS: 120 HERTZ: 60	FELLOWES, INC. 4400301 POWERSHRED LX65	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - NULL	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: MED ROOM
ROOM NUMBER: 218

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		104140 MED01	1	DISPENSER, MEDICATION, AUTOMATED 6 DRAWER; LCD MONITOR; KEYBOARD; BARCODE SCANNER; BIOMETRIC ACCESS SYSTEM; SECURE HALF AND FULL HEIGHT CUBIE POCKETS AND (WXDXH) IN: 23.00 x 27.00 x 55.00 LBS: 166.00 MM: 584.20 x 685.80 x 1397.00 KG: 75.30 LOCATION: FLOOR NOTES: ADD OVER CURRENT CONDITION ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 3 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 1A NOM,3 AMPS MAX LOAD CURRENT EMERGENCY POWER REQUIRED	BD CAREFUSION PYXIS MEDSTATION ES 323 6 DRAWER MAIN	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
2		089012 MED02	1	DISPENSER, MEDICATION, AUTOMATED 7-DRAWER AUXILIARY;18-GAUGE SHEET METAL;REMOVABLE, LOCKING REAR PANEL (WXDXH) IN: 23.00 x 27.00 x 47.00 LBS: 133.00 MM: 584.20 x 685.80 x 1193.80 KG: 60.33 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: CONNECTS TO MAIN MED DISPENSER VIA CABLE ELECT NOTE: POWERED BY MAIN VIA CABLE CONNECTION	BD CAREFUSION PYXIS 7-DRAWER AUXILIARY MEDSTATION ES 7-DRAWER AUXILIARY	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
3		105662 MED04	1	DISPENSER, MEDICATION, AUTOMATED DOUBLE COLUMN AUXILIARY;8 DOORS (WXDXH) IN: 52.00 x 28.00 x 80.00 LBS: 470.50 MM: 1320.80 x 711.20 x 2032.00 KG: 213.42 LOCATION: FLOOR NOTES: 1 IS AN ADD OVER CURRENT CONDITION ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 WATS: 65 ELECT NOTE: 120V;60HZ;65W;EMERGENCY POWER;DATA REQUIRED CONNECTS TO MAIN UNIT VIA CABLE	BD CAREFUSION PYXIS DOUBLE COLUMN AUX MEDSTATION ES	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: MED ROOM
ROOM NUMBER: 218

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		105843 REF04	1	REFRIGERATOR, UNDERCOUNTER, MEDICAL GRADE 5.3 CF;SOLID FIELD REVERSIBLE DOOR;TEMP RANGE 2-10 DEG C;7" I.C3 COLOR TOUCHSCREEN INTERFACE;AUTO DEFROST;R600A ISOBUTANE (WXDXH) IN: 25.00 x 29.00 x 34.00 LBS: 174.00 MM: 635.00 x 736.60 x 863.60 KG: 78.93 LOCATION: FLOOR NOTES: 2024-09 SHD CONFIRMED REMOT CONNECTIONS NOT REQUIRED ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CAHNGED FROM OF/OI TO OF/VI DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ45 ETHERNET VOLTS: 115 HERTZ: 60 AMPS: 1.4 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 8-10' POWER CORD;12V RECHARGEABLE LEAD ACID BATTERY FOR DISPLAY & ALARM BACKUP	HELMER LABS, INC. ILR105-GX I SERIES ADA BTU: 460 MECH NOTE: REAR ACCESS PORT FOR EXT MONITORING PROBES	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
5		079188 SHW06	1	SHELVING, WIRE 5 SHELVES;STATIONARY;CHROME;SHELF INLAY TO INCLUDE: (WXDXH) IN: 24.00 x 24.00 x 74.00 LBS: 61.00 MM: 609.60 x 609.60 x 1879.60 KG: 27.67 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
6		1020123 MBOX01	1	BOX, RETURN, DRUGS LOCKABLE; WALL MOUNTED; BEIGE METAL ANTI-THEFT BAFFLE; BOTTOM IS (WXDXH) IN: 12.00 x 5.20 x 6.25 LBS: 9.00 MM: 304.80 x 132.08 x 158.75 KG: 4.09 LOCATION: WALL NOTES: 2024-09 SHD CONFIRMED WALL MOUNTED CHANGE CODE: C CHANGE NOTES: CHANGE FROM OF/OI COUNTER MNT TO OF/CI WALL MNT	HEALTH CARE LOGISTICS 3783 LOCKABLE RETURN DRUG BOX	OF/CI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: MED ROOM
ROOM NUMBER: 218

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		506580	1	ALLOWANCE, BIN (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		1015778 RAL01	1	RAIL SYSTEM RAIL BIN SYSTEM; PRE-DRILLED IN 1" INCREMENTS; WALL-MOUNTING HARDWARE INCLUDED (WDXH) IN: 31.00 x 1.00 x 3.00 LBS: 1.50 MM: 787.40 x 25.40 x 76.20 KG: 0.69 LOCATION: WALL	VOLLRATH COMPANY 91954 TRAEX 31"	OF/CI	NEW	GROUP: 1 FIXED
9		1020550 DSPP04B	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL CONTROLLED SUBSTANCE WASTE MANAGEMENT SYSTEM. COMPLIANCE WITH REGULATORY GUIDELINES. REDUCES OPPORTUNITIES FOR DRUG (WDXH) IN: 15.50 x 9.50 x 12.50 LBS: 17.00 MM: 393.70 x 241.30 x 317.50 KG: 7.72 LOCATION: SPACE SPEC COMMENTS: WALL/COUNTER BRACKET INSTALLED SEPARATELY BY CONTRACTOR CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY 0085-013-000 SMART SINK	OF/OI	NEW	GROUP: 1 FIXED
10		1020549 DSPP04A	1	BRACKET CACTUS SMART SINK MOUNTING BRACKET. TO INCLUDE: (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: WALL SPEC COMMENTS: BRACKET INSTALLED BY CONTRACTOR, SMART SINK SENT TO SITE FOR OWNER INSTALL CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY CACTUS SMART SINK MOUNTING BRACKET 0085-000-003	OF/CI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: MED ROOM
ROOM NUMBER: 218

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		087350		1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL 8 GAL;SLIDE LID;10/CS;WHITE (WDXDH) IN: 15.50 x 11.00 x 17.75 LBS: 4.00 MM: 393.70 x 279.40 x 450.85 KG: 1.82 LOCATION: FLOOR	KENDALL HEALTHCARE, COVIDIEN 8851 PHARMA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		1015402		1	SHARPS, DISPOSAL, FLOOR BIN 8 GAL SHARPS WITH SLIDING LID (WDXDH) IN: 10.50 x 15.50 x 17.25 LBS: 1.00 MM: 266.70 x 393.70 x 438.15 KG: 0.46 LOCATION: FLOOR NOTES: 2024-0924 CHANGED FROM NEW TO INFO PER SHD	COVIDIEN 8980S 8 GAL	OF/OI	INFO	GROUP: 3 MINOR MOVABLE
13		087035 DSP01		1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXDH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
14		1004090		1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WDXDH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01		1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXDH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ALCOVE
ROOM NUMBER: 219

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		053947 CRI01	1	CART, ISOLATION 4 DRAWER;YELLOW;BASIC CART W/DRAWER; CONFIGURATION OF 1 EA.3" (WXDXH) IN: 33.50 x 25.00 x 46.25 LBS: 165.00 MM: 850.90 x 635.00 x 1174.75 KG: 74.85 LOCATION: FLOOR	ARMSTRONG MEDICAL INDUSTRIES AKL-Y-4 TWO TONE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		1001245 MVS02	1	MONITOR, VITAL SIGNS MOBILE STAND;10.1" LCD TOUCHSCREEN DISPLAY;EWS (EARLY WARNING SYSTEM); MASIMO RAINBOW SET OPTIONS (SPO2, RRA, SPHB);THERMAL (WXDXH) IN: 21.00 x 21.00 x 49.85 LBS: 39.00 MM: 533.40 x 533.40 x 1266.19 KG: 17.70 LOCATION: FLOOR NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: CONNECT DIRECTLY INTO EMR OR HIS VIA HL-7 VOLTS: 120 HERTZ: 60 WATS: 60 ELECT NOTE: RECHARGEABLE LITHIUM ION BATTERY 11.1V 7800MAH	PHILIPS MEDICAL SYSTEMS 863380/989803176601 EARLYVUE VS30 W/ROLLSTAND	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NUTRITION ALCOVE
ROOM NUMBER: 220

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		091485 ICE01	1	ICE MAKER W/WATER DISPENSER 12LB STORAGE;260LB PROD/24 HRS;SS EXTERIOR;EXTERNAL AIR FILTER;AGION ANTIMICROBIAL;R-134A REFRIGERANT;H2 NUGGET STYLE ICE (WXDXH) IN: 17.00 x 25.00 x 35.00 LBS: 183.00 MM: 431.80 x 635.00 x 889.00 KG: 83.01 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 7 ELECT NOTE: SINGLE PHASE;15A MAX FUSE SEE VENDOR TECHNICAL INFORMATION	SCOTSMAN ICE SYSTEMS HID312A-1 MERIDIAN AIR COOL	OF/CI	NEW	GROUP: 1 FIXED
2		1021297	1	FILTER VALVE-IN-HEAD WATER FILTER SYSTEM. INSTALLS ON THE 3/4-INCH MAIN WATER LINE AT THE POINT OF ENTRY. FILTER HOUSING DESIGNED (WXDXH) IN: 5.00 x 5.00 x 13.25 LBS: 0.50 MM: 127.00 x 127.00 x 336.55 KG: 0.23 LOCATION: WALL CHANGE CODE: A CHANGE NOTES: ADDED PER SENECA, HGA HAS ALREADY ACCOUNTED FOR FI	CULLIGAN WATER SYSTEMS HF-360	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NUTRITION ALCOVE
ROOM NUMBER: 220

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		055770 OVM01	1	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WXDXH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP NOTES: 2023-0606: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		083455 COF01	1	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WXDXH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED
6		1006012 COF04	1	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WXDXH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NUTRITION ALCOVE
ROOM NUMBER: 220

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1014420		1	WASTE RECEPTACLE 13GAL;OPEN TOP;BLUE (WDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 2026699 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NURSE STATION
ROOM NUMBER: 221

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015781 MTC02	1	MONITOR, CENTRAL STATION PIIC IX SURVEILLANCE CENTRAL MONITORING; 23" TOUCH; CPU TO INCLUDE: (WXDXH) IN: 16.00 x 8.20 x 16.00 LBS: 30.00 MM: 406.40 x 208.28 x 406.40 KG: 13.61 LOCATION: COUNTERTOP NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 ELECT NOTE: CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. EMERGENCY POWER	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		056002 CRC04	1	CART, RESUSCITATION, CARDIAC CART, RESUSCITATION, CARDIAC; 6 DRAWERS;4-3", 1-6",1-12",PVC BUMPER;LOCKING BREAKAWAY BAR, TWIN (WXDXH) IN: 43.00 x 25.00 x 47.00 LBS: 165.00 MM: 1092.20 x 635.00 x 1193.80 KG: 74.85 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE;	ARMSTRONG MEDICAL INDUSTRIES AR-6 PREMIER W/ER PK	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		082533 DEF01	1	DEFIBRILLATOR R SERIES ALS DEFIBRILLATOR;AED W/MANUAL OVERRIDE;GUIDELINES 2010 COMPATIBLE;CODE READINESS TESTING SYSTEM;RECTILINEAR (WXDXH) IN: 10.50 x 12.50 x 8.20 LBS: 15.20 MM: 266.70 x 317.50 x 208.28 KG: 6.90 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: OPTIONAL WIFI: 802.11B USB PORT 2.0 COMPATIBLE VOLTS: 120 HERTZ: 60 ELECT NOTE: SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER EMERGENCY POWER	ZOLL MEDICAL CORPORATION 30720005201310012 R SERIES ALS	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NURSE STATION
ROOM NUMBER: 221

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		094808 ASP01	1	PUMP, SUCTION/ASPIRATOR, GENERAL, PORTABLE PORTABLE; RECHARGEABLE BATTERY; LED DISPLAY; HEPA FILTER; BEMIS DISPOSABLE CANISTER SYSTEM; HIGH FLOW; HIGH VACUUM; (WXDXH) IN: 13.00 x 6.30 x 12.40 LBS: 8.90 MM: 330.20 x 160.02 x 314.96 KG: 4.04 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 ELECT NOTE: ELECTRICAL SHOWN FOR BATTERY CHARGER	LAERDAL MEDICAL CORP. LSU 4000 78002001	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		1023001 TMP02	1	THERMOMETER, ELECTRONIC, WALL MOUNTED TEMPORAL ARTERY PROFESSIONAL THERMOMETER; KEYLESS WALL MOUNT W/HIDDEN RELEASE LEVER; NO CABLE (WXDXH) IN: 3.23 x 0.00 x 6.20 LBS: 0.00 MM: 82.05 x 0.00 x 157.48 KG: 0.00 LOCATION: WALL NOTES: 2025.05.12 UPDATED MODEL FROM 5000 TO 5000S WITH NO CABLE KEYLESS WALL MOUNT PER SHD CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: 9-VOLT BATTERY	EXERGEN CORPORATION TAT-5000S 124275S/134305	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
6		105759 GLU01	1	LAB ANALYZER, GLUCOSE TOUCH SCREEN;QR BARCODE SCANNER;1,000 TEST STORAGE; 4,000 USER STORAGE; 500 TEST STORAGE;NUMERIC OR PASS/FAIL (WXDXH) IN: 3.10 x 5.80 x 1.18 LBS: 0.49 MM: 78.74 x 147.32 x 29.98 KG: 0.23 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: DATA REQUIRED: ETHERNET NETWORK DROP OR 802.11 A/B/G/N WIRELESS ACCESS. SEE CUTSHEET FOR FURTHER DETAILS. VOLTS: 120 HERTZ: 60 AMPS: .6 ELECT NOTE: STATSTRIP: 3.7V LI POLYMER RECHARGEABLE BATTERY ELECTRICAL INFO IS FOR DOCKING STATION	NOVA BIOMEDICAL 44993 STATSTRIP GLU 2	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
7		1005565 BIN01	1	BIN, SHREDDING SHREDDER COLLECTION BIN (WXDXH) IN: 19.00 x 16.00 x 27.00 LBS: 35.00 MM: 482.60 x 406.40 x 685.80 KG: 15.88 LOCATION: FLOOR SPEC COMMENTS: LEASED ITEM	RAPID SHRED SHORT CONSOLE	OF/OI	LEASED	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NURSE STATION
ROOM NUMBER: 221

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1015757 TELE02	1	MONITOR, TELEMETRY 1.4 GHZ;SMART HOPPING; CO1 ENHANCED ARRHYTHMIA; S02 ECG + FAST SPO2 ENABLED (WDXDH) IN: 3.00 x 2.00 x 5.00 LBS: 1.00 MM: 76.20 x 50.80 x 127.00 KG: 0.46 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: WLAN ELECT NOTE: THREE AA BATTERIES	PHILIPS MEDICAL SYSTEMS 865350 MX40	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
9		1015757 TELE02	1	MONITOR, TELEMETRY 1.4 GHZ;SMART HOPPING; CO1 ENHANCED ARRHYTHMIA; S02 ECG + FAST SPO2 ENABLED (WDXDH) IN: 3.00 x 2.00 x 5.00 LBS: 1.00 MM: 76.20 x 50.80 x 127.00 KG: 0.46 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: WLAN ELECT NOTE: THREE AA BATTERIES	PHILIPS MEDICAL SYSTEMS 865350 MX40	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
10		1015757 TELE02	1	MONITOR, TELEMETRY 1.4 GHZ;SMART HOPPING; CO1 ENHANCED ARRHYTHMIA; S02 ECG + FAST SPO2 ENABLED (WDXDH) IN: 3.00 x 2.00 x 5.00 LBS: 1.00 MM: 76.20 x 50.80 x 127.00 KG: 0.46 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: WLAN ELECT NOTE: THREE AA BATTERIES	PHILIPS MEDICAL SYSTEMS 865350 MX40	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: NURSE STATION
ROOM NUMBER: 221

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		1015757 TELE02	1	MONITOR, TELEMETRY 1.4 GHZ;SMART HOPPING; CO1 ENHANCED ARRHYTHMIA; S02 ECG + FAST SPO2 ENABLED (WDXH) IN: 3.00 x 2.00 x 5.00 LBS: 1.00 MM: 76.20 x 50.80 x 127.00 KG: 0.46 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: WLAN ELECT NOTE: THREE AA BATTERIES	PHILIPS MEDICAL SYSTEMS 865350 MX40	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: PATIENT, TOILET
ROOM NUMBER: 222

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: STAFF TOILET
ROOM NUMBER: 223

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ACUTE EQUIPMENT STORAGE
ROOM NUMBER: 224

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1023032 LFT05*	1	LIFT, PATIENT, MOBILE 500# CAP; POWER BASE; AUDIBLE ALERT; SIX-POINT CRADLE W/SPRING CLIPS; INCLUDES BATTERY AND EXTERNAL CHARGER STAND (WXDXH) IN: 45.80 x 52.00 x 76.30 LBS: 156.50 MM: 1163.32 x 1320.80 x 1938.02 KG: 70.99 LOCATION: FLOOR SPEC COMMENTS: EXTERNAL WIDTH WHEN OPEN: 45.8" EXTERNAL WIDTH WHEN CLOSED: 31.6" NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW ELECT NOTE: RECHARGEABLE BATTERY	DRIVE MEDICAL FLP500 LEVANTAR	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		074071 LFT02	1	LIFTER, PATIENT MOBILE;400LB CAP;ACTIVE STANDING/TRANSFER AID; FLIP DOWN SWIVEL (WXDXH) IN: 34.65 x 36.22 x 41.50 LBS: 65.00 MM: 880.11 x 919.99 x 1054.10 KG: 29.49 LOCATION: FLOOR SPEC COMMENTS: SPACE REQUIREMENT: 87" X 79" CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW	ARJOHUNTLEIGH NTB2000 SARA STEDY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ACUTE EQUIPMENT STORAGE
ROOM NUMBER: 224

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		077986 SCL02	1	SCALE, STAND-ON 1000LB CAPACITY;HANDRAIL;DIGITAL;MOBILE;LCD SCREEN;BMI;MANUAL TARE; ZERO;AUTO-OFF;AUTO-HOLD;INCL 2 WHEELS FOR EASE OF (WXDXH) IN: 27.75 x 30.00 x 49.40 LBS: 82.00 MM: 704.85 x 762.00 x 1254.76 KG: 37.20 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW VOLTS: 120 HERTZ: 60 ELECT NOTE: 120-240VAC, 50-60HZ 9V DC ADAPTER OR QTY (6) AA BATTERIES	HEALTH O METER, INC. 2101KL/ACC	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
5		075704 LHT03	1	LIGHT, EXAM/TREATMENT PROCEDURE LIGHT WITH MOBILE STAND;3 WHITE LEDS;TOUCHLESS ON/OFF DIMMING OPTION (WXDXH) IN: 21.00 x 21.00 x 47.00 LBS: 17.35 MM: 533.40 x 533.40 x 1193.80 KG: 7.87 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 AMPS: 0.65 ELECT NOTE: 3-WIRE GROUNDED CORD SET, 8'	WELCH ALLYN, INC. 44600 GS 600	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
6		066897	2	CHAIR, COMMODE, BARIATRIC BARIATRIC; 850 LBS CAP.; DROP ARM; 28"SEAT; STEEL; POWDER (WXDXH) IN: 29.00 x 19.50 x 25.50 LBS: 15.00 MM: 736.60 x 495.30 x 647.70 KG: 6.81 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. MDS89668XW GUARDIAN	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
7		1002317	2	WALKER ALUMINUM FOLDING WALKER WITH WHEELS; PUSH-BUTTON, DUAL- RELEASE MECHANISMS FOR EASY-FOLDING; 350 LB WEIGHT CAPACITY (WXDXH) IN: 24.00 x 17.50 x 39.00 LBS: 12.00 MM: 609.60 x 444.50 x 990.60 KG: 5.45 LOCATION: FLOOR	MCKESSON 1073638 146-10210-4	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
8		102251	1	CHAIR, SHOWER 500LB WT CAP;22" X 17" SEAT;WHITE ALUMINUM FRAME; (4) REAR LOCKING CASTERS;BLUE PADDED SEAT/ARMS;HT (WXDXH) IN: 26.00 x 36.00 x 42.00 LBS: 36.80 MM: 660.40 x 914.40 x 1066.80 KG: 16.70 LOCATION: FLOOR	DRIVE MEDICAL NRS185008 ALUM REHAB BARI	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ACUTE EQUIPMENT STORAGE
ROOM NUMBER: 224

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		051488 IVP02	2	I.V. POLE STAINLESS STEEL; 4 HOOK; 5 CASTER BASE; HAND OPERATED (WXDXH) IN: 21.25 x 21.25 x 118.00 LBS: 20.00 MM: 539.75 x 539.75 x 2997.20 KG: 9.08 LOCATION: FLOOR SPEC COMMENTS: HEIGHT ADJUSTABLE FROM 69" TO 118" CHANGE CODE: C CHANGE NOTES: 2025.05.12 MOVE FROM RM 200 TO RM 224	PEDIGO PRODUCTS, INC. P-1572-4 FOUR HOOK	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
10		051488 IVP02	2	I.V. POLE STAINLESS STEEL; 4 HOOK; 5 CASTER BASE; HAND OPERATED (WXDXH) IN: 21.25 x 21.25 x 118.00 LBS: 20.00 MM: 539.75 x 539.75 x 2997.20 KG: 9.08 LOCATION: FLOOR SPEC COMMENTS: HEIGHT ADJUSTABLE FROM 69" TO 118" CHANGE CODE: C CHANGE NOTES: 2025.05.12 MOVE FROM RM 202 TO RM 224 PER SHD RE	PEDIGO PRODUCTS, INC. P-1572-4 FOUR HOOK	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
11		059223	2	PUMP, COMPRESSION LEG & FOOT COMPRESSION; COMPRESSION SLEEVE CONTAINS 3 BLADDERS TO DELIVER CIRCUMFERENTIAL COMPRESSION ALL-IN-ONE (WXDXH) IN: 7.70 x 7.30 x 6.80 LBS: 5.00 MM: 195.58 x 185.42 x 172.72 KG: 2.27 LOCATION: COUNTERTOP CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 WATS: 50 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 50VA;13' POWER CORD WITH HOSPITAL GRADE, REGION-SPECIFIC CORDAGE/PLUG OUTLET MUST BE GROUNDED RECHARGEABLE LITHIUM ION BATTERY 10.8V, 2200MAH	KENDALL HEALTHCARE, COVIDIEN 29525 700 SERIES	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: ACUTE EQUIPMENT STORAGE
ROOM NUMBER: 224

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		1023026 DUV01*	1	DISINFECTOR, ROOM, MOBILE TELESCOPING ROOM STERILIZATION UNIT; 55W UV-C LAMPS; (WXDXH) IN: 16.00 x 13.50 x 46.75 LBS: 44.00 MM: 406.40 x 342.90 x 1187.45 KG: 19.96 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 110 HERTZ: 60	FIRST RESPONDER TECHNOLOGIES STERIZAP UV-C ST E500UVST	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: ACUTE CARE
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 225

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		052851 HKP01	1	CART, HOUSEKEEPING SS;2 SHELF;FOLDING PLATFORM W/LOCKING DOOR;ZIP BAG;TRASH LID;TOOL BRACKET W/4 HANDLES & 3 HOOKS;TRAY TOP (WDXH) IN: 23.00 x 36.00 x 42.50 LBS: 32.00 MM: 584.20 x 914.40 x 1079.50 KG: 14.52 LOCATION: FLOOR	ROYCE ROLLS RINGER CO. F36 FOLDING F36 FOLDING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
4		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: OFFICE
ROOM NUMBER: 706

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 708

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1014559 RCK05	1	RACK, DRYING SOILED MAT TRANSFER CAR; 500 LB CAPACITY (WDXH) IN: 28.00 x 42.00 x 40.00 LBS: 50.00 MM: 711.20 x 1066.80 x 1016.00 KG: 22.68 LOCATION: FLOOR	CACTUS MAT MFR. CO. 6477 PORT-A-MAT	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: TRASH
ROOM NUMBER: 710

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015414 TRK02	1	TRUCK, TILT, WASTE/LINEN 16 CUFT MANUAL CUBE TRUCK; INCLUDES LID (WXDXH) IN: 30.40 x 44.00 x 35.50 LBS: 60.00 MM: 772.16 x 1117.60 x 901.70 KG: 27.22 LOCATION: FLOOR NOTES: BLACK IN COLOR; INCLUDES LID	TOTER INC. MMC16-10836 BLACK	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: STAFF TOILET
ROOM NUMBER: 712

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: DISHROOM
ROOM NUMBER: 713

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		094275	3	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		057206 GLV01	3	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: DIETARY
ROOM NAME: HOSPITAL DINING
ROOM NUMBER: 715

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		083455 COF01	1	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WXDXH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED
2		1006012 COF04	1	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WXDXH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1014420	1	WASTE RECEPTACLE 13GAL;OPEN TOP;BLUE (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 2026699 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED LOUNGE
ROOM NUMBER: 300

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015398	2	WASTE RECEPTACLE STEEL; INDOOR USE; INCLUDES ONE 24 GALLON SQUARE CONTAINER AND TWO 8 GALLON HALF-ROUND CONTAINERS; PREMIUM GLOSS POWDER (WDXDH) IN: 15.75 x 31.50 x 30.00 LBS: 81.50 MM: 400.05 x 800.10 x 762.00 KG: 36.97 LOCATION: FLOOR	TRASHCAN WAREHOUSE RC-KDSQ-3C-BT THREE-STREAM COMBO	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: PUBLIC TOILET
ROOM NUMBER: 301

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: PUBLIC TOILET
ROOM NUMBER: 302

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED ALCOVE 03
ROOM NUMBER: 303

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1001212 STR01	1	STRETCHER TRANSPORT STRETCHER 26" WIDE DECK, TO INCLUDE: TN TRANSPORT P8005 (WDXH) IN: 36.00 x 83.00 x 34.25 LBS: 260.00 MM: 914.40 x 2108.20 x 869.95 KG: 117.94 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	HILL-ROM P8005 TRANSPORT	OF/VI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: STAFF TOILET
ROOM NUMBER: 304

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: PATIENT TOILET
ROOM NUMBER: 305

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED ALCOVE 02
ROOM NUMBER: 306

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015397 CRL01	1	CART, LINEN 5" QUIET, NON-MARKING TENITE CASTERS AMONG THE BEST ON THE MARKET; REINFORCED INJECTION-MOLDED SHELVES WITH ½" LIP TO (WDXH) IN: 40.00 x 20.00 x 63.00 LBS: 50.00 MM: 1016.00 x 508.00 x 1600.20 KG: 22.68 LOCATION: FLOOR	DIRECT SUPPLY, INC LC36-4-42 IPU PREMIUM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		106032 ECG01	1	ELECTROCARDIOGRAPH WIRELESS;15" COLOR TOUCH SCREEN;UP TO 18 LEADS;DIMENSIONS REFLECT AN OPEN DISPLAY (38" HT WHEN CLOSED) TO INCLUDE: (WDXH) IN: 20.10 x 25.90 x 53.00 LBS: 200.00 MM: 510.54 x 657.86 x 1346.20 KG: 90.72 LOCATION: FLOOR DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: WIRELESS CONFIGURATION OTHER CONNECTION AVAILABLE: LAN 10/100 BASE-TX IEEE 802.3 ETHERNET VIA ON-BOARD RJ45 VOLTS: 120 HERTZ: 60 WATS: 75 ELECT NOTE: LITHIUM ION BATTERY (2 MODULES, HOT SWAPPABLE)	PHILIPS MEDICAL SYSTEMS 860315/860318 PAGEWRITER TC70	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1015745 PHY06	1	MONITOR, PHYSIOLOGICAL 9" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP, DUAL PRESS AND TEMP; INCLUDES (WDXH) IN: 21.00 x 21.00 x 54.30 LBS: 42.30 MM: 533.40 x 533.40 x 1379.22 KG: 19.19 LOCATION: FLOOR NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866060_NAM INTELLIVUE MX400	OF/VI	NEW	GROUP: 1 FIXED

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED VESTIBULE
ROOM NUMBER: 307

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1002727	2	CHAIR, TRANSPORT 800LB CAP;LIFTING ARMREST;NON-FOLDING;ANTI-TIP DESIGN TO INCLUDE: (WXDXH) IN: 31.00 x 39.00 x 41.00 LBS: 79.00 MM: 787.40 x 990.60 x 1041.40 KG: 35.84 LOCATION: FLOOR NOTES: MUST INCLUDE O2 HOLDER	STAXI CORPORATION LIMITED RA010-L-C RANGER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1006581	1	WHEELCHAIR, ADULT, BARIATRIC BARIATRIC WHEELCHAIR; WEIGHT CAPACITY: 700 LBS; PUSH-TO-LOCK (WXDXH) IN: 34.00 x 48.00 x 32.00 LBS: 68.00 MM: 863.60 x 1219.20 x 812.80 KG: 30.85 LOCATION: FLOOR NOTES: MUST INCLUDE O2 HOLDER	ALCO SALES & SERVICE COMPANY AL-79000 CLASSIC 700	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1015399	1	DISPENSER, PPE STATION KIT INCLUDES: - RS004-0212 RESPIRATORY HYGIENE STATION (WXDXH) IN: 22.25 x 17.00 x 58.06 LBS: 18.10 MM: 565.15 x 431.80 x 1474.73 KG: 8.22 LOCATION: FLOOR	BOWMAN MANUFACTURING COMPANY BD114-0012 COMPLIANCE KIT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: GREETER / REGISTRATION
ROOM NUMBER: 309

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TRIAGE
ROOM NUMBER: 310

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1000325 CHR07	1	CHAIR, RECLINER 3 POSITION RECLINER CHAIR; TRENDELENBURG; 350# WEIGHT CAPACITY; 21" SEAT WIDTH; FOUR SWIVEL, THREE LOCKING CASTERS (WDXDH) IN: 51.00 x 75.00 x 50.00 LBS: 145.00 MM: 1295.40 x 1905.00 x 1270.00 KG: 65.78 LOCATION: FLOOR NOTES: 2025.05.12 PER SHD REVIEW 2 DOC, CHANGED TO EXISTING CHANGE CODE: C CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	GRAHAM-FIELD HEALTH PRODUCTS FR566G LUMEX DELUXE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY;WALLBOARD;OPHTH/OTO;SPHYG;TEMP TO INCLUDE: (WDXDH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
3		097638 SCL02	1	SCALE, STAND-ON 1000LB CAP;TOUCH LESS SONAR HEIGHT ROD;USB PORT;BMI;HL7 (WDXDH) IN: 17.00 x 18.50 x 89.50 LBS: 22.00 MM: 431.80 x 469.90 x 2273.30 KG: 9.98 LOCATION: FLOOR DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: RS232 SERIAL & USB-B PORT VOLTS: 120 HERTZ: 60 ELECT NOTE: 12 AA BATTERIES OR AC ADAPTER	DETECTO SCALE COMPANY 809161198805 ICON DIGITAL	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		006656 MAY01	1	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TRIAGE
ROOM NUMBER: 310

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
9		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TRIAGE
ROOM NUMBER: 310

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		102576 EYE01	1	CHART, EYE, WALL MOUNTED TESTING VISUAL ACUITY EYE TEST CHART-10 FOOT; MATT (WXDXH) IN: 11.00 x 1.00 x 24.00 LBS: 1.00 MM: 279.40 x 25.40 x 609.60 KG: 0.46 LOCATION: WALL	WESTERN OPHTHALMICS CORP OS-1935 10FT - SNELLEN	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		087746 SCL03	1	SCALE, INFANT DIGITAL;HEAD/FOOT;WIRELESS EMR;LARGE TRAY; MEASURING ROD (WXDXH) IN: 24.40 x 15.60 x 8.50 LBS: 9.50 MM: 619.76 x 396.24 x 215.90 KG: 4.31 LOCATION: COUNTERTOP NOTES: WILL RESIDE ON SUPPLY CART - BETTER TO CHOOSE BATTERY POWERED SCALE DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 ELECT NOTE: 6 AA BATTERIES;9-12V WIRELESS USB POWER ADAPTOR	SECA CORPORATION 374 W/233 374	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		1006176 CRS03	1	CART, SUPPLY 3 3"-, 2 6"-, AND 1 9"-HIGH DRAWER CONFIGURATION;ACCOMMODATES A VARIETY OF INTERIOR AND EXTERIOR ACCESSORIES;LIGHTWEIGHT (WXDXH) IN: 29.25 x 21.25 x 42.50 LBS: 145.00 MM: 742.95 x 539.75 x 1079.50 KG: 65.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE SP30W6 PACE 30	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ON-CALL
ROOM NUMBER: 311

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: STAFF / MD WORKROOM
ROOM NUMBER: 312

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015781 MTC02	1	MONITOR, CENTRAL STATION PIIC IX SURVEILLANCE CENTRAL MONITORING; 23" TOUCH; CPU TO INCLUDE: (WXDXH) IN: 16.00 x 8.20 x 16.00 LBS: 30.00 MM: 406.40 x 208.28 x 406.40 KG: 13.61 LOCATION: COUNTERTOP NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 ELECT NOTE: CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. EMERGENCY POWER	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		1015755 PHY08	1	MONITOR, PHYSIOLOGIC, TRANSPORT INTELLIVUE MULTI-MEASUREMENT TRANSPORT ACQUISITION MODULE; 5 HOUR BATTERY; CHARGES VIA DOCKING STATION, 120VAC TO INCLUDE: (WXDXH) IN: 9.80 x 3.80 x 4.40 LBS: 3.10 MM: 248.92 x 96.52 x 111.76 KG: 1.41 LOCATION: COUNTERTOP NOTES: 2025.09.08 SHD CONFIRMED KEEP ON MEQ LIST, UPDATED LOCATION FROM SPACE TO COUNTER MOUNTED CHANGE CODE: C CHANGE NOTES: LOCATION CHANGE DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.3 ELECT NOTE: - REMOVABLE, RECHARGEABLE 5 HOUR BATTERY - 120 V AC-POWERED INTELLIVUE DOCK PROVIDES CHARGING AND NETWORK CONNECTIVITY - CAN BE ATTACHED TO A HOST MONITOR WHICH PROVIDES THE POWER TO THE X3	PHILIPS HEALTHCARE - MONITORING SYSTEMS 867030 INTELLIVUE X3	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: STAFF / MD WORKROOM
ROOM NUMBER: 312

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1015752	1	MONITOR ACCESSORY, MODULE MULTI-MEASUREMENT MODULE TO INCLUDE: (WXDXH) IN: 2.40 x 7.40 x 3.80 LBS: 1.60 MM: 60.96 x 187.96 x 96.52 KG: 0.73 LOCATION: SPACE NOTES: 2025.09.08 SHD CONFIRMED KEEP ON LIST DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: NETWORK CONNECTION REQUIRED LAN SIGNALS: IEEE 802.3 10BASE-T AND 100BASE-TX COMPLIANT SERIAL SIGNALS: RS-422 COMPLIANT REFERENCE VENDOR'S TECHNICAL DATA FOR DETAILS	PHILIPS MEDICAL SYSTEMS 867036 INTELLIVUE MMX	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1003252 CBW03	1	CABINET, WARMING, DUAL, FREESTANDING DUAL FREESTANDING FLUID AND BLANKET WARMING CABINET WITH WINDOW GLASS DOORS. FEATURES 18 INCH DEPTH PERFECT FOR BUILT-IN OR (WXDXH) IN: 30.90 x 21.70 x 67.25 LBS: 358.00 MM: 784.86 x 551.18 x 1708.15 KG: 162.39 LOCATION: FLOOR SPEC COMMENTS: 50" DEEP WITH DOOR OPEN NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO VOLTS: 120 HERTZ: 60 AMPS: 14.2 WATS: 1.7KW PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: SINGLE PHASE.;8FT CORD	PEDIGO PRODUCTS, INC. P-2148	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		091485 ICE01	1	ICE MAKER W/WATER DISPENSER 12LB STORAGE;260LB PROD/24 HRS;SS EXTERIOR;EXTERNAL AIR FILTER;AGION ANTIMICROBIAL;R-134A REFRIGERANT;H2 NUGGET STYLE ICE (WXDXH) IN: 17.00 x 25.00 x 35.00 LBS: 183.00 MM: 431.80 x 635.00 x 889.00 KG: 83.01 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 7 ELECT NOTE: SINGLE PHASE;15A MAX FUSE SEE VENDOR TECHNICAL INFORMATION	SCOTSMAN ICE SYSTEMS HID312A-1 MERIDIAN AIR COOL BTU: 3200	OF/CI	NEW	GROUP: 1 FIXED COLD WATER: YES DRAIN REQUIRED: YES PLUMB NOTE: SEE VENDOR TECHNICAL INFORMATION

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: STAFF / MD WORKROOM
ROOM NUMBER: 312

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		1021297	1	FILTER VALVE-IN-HEAD WATER FILTER SYSTEM. INSTALLS ON THE 3/4-INCH MAIN WATER LINE AT THE POINT OF ENTRY. FILTER HOUSING DESIGNED (WXDXH) IN: 5.00 x 5.00 x 13.25 LBS: 0.50 MM: 127.00 x 127.00 x 336.55 KG: 0.23 LOCATION: WALL CHANGE CODE: A CHANGE NOTES: ADDED PER SENECA, HGA HAS ALREADY ACCOUNTED FOR FI	CULLIGAN WATER SYSTEMS HF-360	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
							COLD WATER: YES PLUMB NOTE: SEE VENDOR SPECS	
7		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
					MECH NOTE: REFRIGERANT: R600A			
8		054875 IST01	1	LAB ANALYZER, CHEMISTRY, MANUAL HANDHELD; W/DOWNLOADER/RECHARGER NETWORK; UTILIZES TEST CARTRIDGE OR GLUCOSE STRIP; MEASURES: BLOOD (WXDXH) IN: 3.04 x 2.85 x 9.24 LBS: 2.00 MM: 77.22 x 72.39 x 234.70 KG: 0.91 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 0.9 WATS: 120 ELECT NOTE: REQUIRES: 2-9V LITHIUM OR NICKEL CADMIUM BATTERIES RECHARGER: 9V-120V AC ADAPTER	ABBOTT LABORATORIES 06F16-10/06F23-46 I-STAT 1	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
9		105759 GLU01	1	LAB ANALYZER, GLUCOSE TOUCH SCREEN; QR BARCODE SCANNER; 1,000 TEST STORAGE; 4,000 USER STORAGE; 500 TEST STORAGE; NUMERIC OR PASS/FAIL (WXDXH) IN: 3.10 x 5.80 x 1.18 LBS: 0.49 MM: 78.74 x 147.32 x 29.98 KG: 0.23 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES	NOVA BIOMEDICAL 44993 STATSTRIP GLU 2	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: STAFF / MD WORKROOM
ROOM NUMBER: 312

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS
				DATA NOTE: DATA REQUIRED: ETHERNET NETWORK DROP OR 802.11 A/B/G/N WIRELESS ACCESS. SEE CUTSHEET FOR FURTHER DETAILS. VOLTS: 120 HERTZ: 60 AMPS: .6 ELECT NOTE: STATSTRIP: 3.7V LI POLYMER RECHARGEABLE BATTERY ELECTRICAL INFO IS FOR DOCKING STATION			
10		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW
							GROUP: 3 MINOR MOVABLE
11		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW
							GROUP: 3 MINOR MOVABLE
12		1014420	1	WASTE RECEPTACLE 13GAL;OPEN TOP;BLUE (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 2026699 SLIM JIM	OF/OI	NEW
							GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: DRUG ROOM
ROOM NUMBER: 313

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		104140 MED01	1	DISPENSER, MEDICATION, AUTOMATED 6 DRAWER; LCD MONITOR; KEYBOARD; BARCODE SCANNER; BIOMETRIC ACCESS SYSTEM; SECURE HALF AND FULL HEIGHT CUBIE POCKETS AND (WXDXH) IN: 23.00 x 27.00 x 55.00 LBS: 166.00 MM: 584.20 x 685.80 x 1397.00 KG: 75.30 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 3 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 1A NOM,3 AMPS MAX LOAD CURRENT EMERGENCY POWER REQUIRED	BD CAREFUSION PYXIS MEDSTATION ES 323 6 DRAWER MAIN BTU: 409 MECH NOTE: 409 BTU/HR	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
2		089012 MED02	1	DISPENSER, MEDICATION, AUTOMATED 7-DRAWER AUXILIARY;18-GAUGE SHEET METAL;REMOVABLE, LOCKING REAR PANEL (WXDXH) IN: 23.00 x 27.00 x 47.00 LBS: 133.00 MM: 584.20 x 685.80 x 1193.80 KG: 60.33 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: CONNECTS TO MAIN MED DISPENSER VIA CABLE ELECT NOTE: POWERED BY MAIN VIA CABLE CONNECTION	BD CAREFUSION PYXIS 7-DRAWER AUXILIARY MEDSTATION ES 7-DRAWER AUXILIARY	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
3		105662 MED04	2	DISPENSER, MEDICATION, AUTOMATED DOUBLE COLUMN AUXILIARY;8 DOORS (WXDXH) IN: 52.00 x 28.00 x 80.00 LBS: 470.50 MM: 1320.80 x 711.20 x 2032.00 KG: 213.42 LOCATION: FLOOR NOTES: 1 IS AN ADD OVER CURRENT CONDITION ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 WATS: 65 ELECT NOTE: 120V;60HZ;65W;EMERGENCY POWER;DATA REQUIRED CONNECTS TO MAIN UNIT VIA CABLE	BD CAREFUSION PYXIS DOUBLE COLUMN AUX MEDSTATION ES BTU: 222 MECH NOTE: 222 BTU/HR	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: DRUG ROOM
ROOM NUMBER: 313

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		065859 SAFE01	1	SAFE, NARCOTICS SINGLE COLUMN INTEGRATED MAIN W/PRINTER SHELF (WXDXH) IN: 49.00 x 28.00 x 79.50 LBS: 369.00 MM: 1244.60 x 711.20 x 2019.30 KG: 167.38 LOCATION: FLOOR NOTES: ADDITION OVER CURRENT CONDITION ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: DATA PORT VOLTS: 120 HERTZ: 60 AMPS: 1.71 ELECT NOTE: EMERGENCY POWER REQUIRED	BD CAREFUSION PYXIS CIISAFE SINGLE/ACC	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
5		1020123 MBOX01	1	BOX, RETURN, DRUGS LOCKABLE; WALL MOUNTED; BEIGE METAL ANTI-THEFT BAFFLE; BOTTOM IS (WXDXH) IN: 12.00 x 5.20 x 6.25 LBS: 9.00 MM: 304.80 x 132.08 x 158.75 KG: 4.09 LOCATION: WALL NOTES: 2024-09 SHD CONFIRMED WALL MOUNTED CHANGE CODE: C CHANGE NOTES: CHANGE FROM OF/OI COUNTER MNT TO OF/CI WALL MNT	HEALTH CARE LOGISTICS 3783 LOCKABLE RETURN DRUG BOX	OF/CI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: DRUG ROOM
ROOM NUMBER: 313

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		105664 REF07	1	REFRIGERATOR, UPRIGHT, PHARMACY 13.3CUFT; 115V; SS; VCC COMPRESSOR-BASED; FORCED-AIR HYDROCARBON 1300-4000RPM SPEED RANGE; AUTO DEFROST; LED DIGITAL (WDXDH) IN: 24.80 x 31.20 x 70.50 LBS: 338.00 MM: 629.92 x 792.48 x 1790.70 KG: 153.32 LOCATION: FLOOR SPEC COMMENTS: DEPTH WITH DOOR OPEN AT 90 DEGREES: 52.3" NOTES: TO INCLUDE: SEISMIC ANCHORING KIT ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CHANGED FROM OF/OI TO OF/VI DATA NOTE: ACCESS PORT: TOP ACCESS PORT- FOR EXTERNAL MONITORING PROBE(S) 0.75 IN DIAMETER; LEFT SIDEWALL PORT - WITH INTERIOR AND EXTERIOR PLUGS 1.75 IN DIAMETER VOLTS: 115 HERTZ: 60 AMPS: 2.3 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A MINIMUM BRANCH CIRCUIT W/EARTH GROUND;8-10' POWER CORD W/NEMA 5-15 HOSP GRADE PLUG;CONTROLLER USES DISPOSABLE 9V BATTERIES	HELMER LABS, INC. HORIZON SERIES HPR113-GX	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
7		066552 PAN01	3	LOUVER PANEL, WALL MOUNTED STORAGE PANEL; 16 GAUGE STEEL W/ GRAY ENAMEL FINISH LOUVERED FOR BINS; 250LB LOAD CAPACITY ONCE FASTENED SECURELY TO	QUANTUM STORAGE SYSTEMS QLP-4819 ULTRA HANGING SYSTEMS	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
8	No image available	506580	1	ALLOWANCE, BIN (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: DRUG ROOM
ROOM NUMBER: 313

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		067123 CRU04	1	CART, UTILITY PLASTIC; 2 SHELF; LOCKABLE DOORS; SLIDING DRAWER; BLACK; 200L (WXDXH) IN: 18.63 x 33.63 x 37.75 LBS: 45.00 MM: 473.21 x 854.21 x 958.85 KG: 20.42 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 3457 -	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1005565 BIN01	1	BIN, SHREDDING SHREDDER COLLECTION BIN (WXDXH) IN: 19.00 x 16.00 x 27.00 LBS: 35.00 MM: 482.60 x 406.40 x 685.80 KG: 15.88 LOCATION: FLOOR SPEC COMMENTS: LEASED ITEM	RAPID SHRED SHORT CONSOLE	OF/OI	LEASED	GROUP: 3 MINOR MOVABLE
11		1020550 DSPP04B	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL CONTROLLED SUBSTANCE WASTE MANAGEMENT SYSTEM. COMPLIANCE WITH REGULATORY GUIDELINES. REDUCES OPPORTUNITIES FOR DRUG (WXDXH) IN: 15.50 x 9.50 x 12.50 LBS: 17.00 MM: 393.70 x 241.30 x 317.50 KG: 7.72 LOCATION: SPACE SPEC COMMENTS: WALL/COUNTER BRACKET INSTALLED SEPARATELY BY CONTRACTOR CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY 0085-013-000 SMART SINK	OF/OI	NEW	GROUP: 1 FIXED
12		1020549 DSPP04A	1	BRACKET CACTUS SMART SINK MOUNTING BRACKET. TO INCLUDE: (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: WALL SPEC COMMENTS: BRACKET INSTALLED BY CONTRACTOR, SMART SINK SENT TO SITE FOR OWNER INSTALL CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY CACTUS SMART SINK MOUNTING BRACKET 0085-000-003	OF/CI	NEW	GROUP: 1 FIXED
13		099447 STL03	1	STOOL, FOOT, WITH HANDRAIL TWO STEP;STEPS ARE 8" AND 16"HIGH;CHROME CORRUGATED RUBBER SURFACE;GREY FLOOR TIPS (WXDXH) IN: 16.00 x 12.00 x 47.00 LBS: 35.00 MM: 406.40 x 304.80 x 1193.80 KG: 15.88 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. P-18-A-K P-18-A-K	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: DRUG ROOM
ROOM NUMBER: 313

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
14		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
15		087350	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL 8 GAL;SLIDE LID;10/CS;WHITE (WDXH) IN: 15.50 x 11.00 x 17.75 LBS: 4.00 MM: 393.70 x 279.40 x 450.85 KG: 1.82 LOCATION: FLOOR	KENDALL HEALTHCARE, COVIDIEN 8851 PHARMA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		1015402	1	SHARPS, DISPOSAL, FLOOR BIN 8 GAL SHARPS WITH SLIDING LID (WDXH) IN: 10.50 x 15.50 x 17.25 LBS: 1.00 MM: 266.70 x 393.70 x 438.15 KG: 0.46 LOCATION: FLOOR <i>NOTES: 2024-0924 CHANGED FROM NEW TO INFO PER SHD</i>	COVIDIEN 8980S 8 GAL	OF/OI	INFO	GROUP: 3 MINOR MOVABLE
17		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: SECURE STORAGE
ROOM NUMBER: 314

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015640 CAB07	2	CABINET, STORAGE, BIN DOORS HAVE 3-POINT LOCKING SYSTEM – KEYS INCLUDED. (48) 5 1/2 X 4 X 3", (48) 7 1/2 X 4 X 3", (WXDXH) IN: 36.00 x 24.00 x 78.00 LBS: 333.00 MM: 914.40 x 609.60 x 1981.20 KG: 151.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	ULINE H-8347BLU 138 BLUE BIN	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023169 LHT01	1	LIGHT, EXAM LED;SINGLE LIGHT HEAD;ARM 1 DROP TUBE LENGTH: 750MM;ARM 1 LENGTH: 800MM (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ANCHOR DETAIL REQ: YES ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL SIRCA	OF/CI	NEW	GROUP: 1 FIXED
2		106126 STR09*	1	STRETCHER 700LB USER WT CAP;30" WIDE PATIENT SURFACE;BIG WHEEL STEERING;ELECTRIC LIFT BASE;INTEGRATED SCALE (WXDXH) IN: 38.00 x 85.00 x 37.25 LBS: 333.00 MM: 965.20 x 2159.00 x 946.15 KG: 151.05 LOCATION: FLOOR NOTES: 2025.09.18 STRETCHER MAKE/MODEL TBC CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60 ELECT NOTE: 120V, 50/60HZ	STRYKER CORP/MEDICAL DIV 1115-000-000XE PRIME X ELEC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
4		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WXDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
<i>NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025.</i> <i>WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY</i> CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE								
5		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS;INCLUDES MOUNTING HARDWARE AND (WXDXH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. <i>NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025.</i> <i>WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI</i> CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY;WALLBOARD;OPHTH/OTO;SPHYG;TEMP TO INCLUDE: (WXDXH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		094482 ISR01	1	INFUSION SYSTEM, RAPID RAPID INFUSER; 750ML/M; IV POLE; RESERVOIR HOLDER; LCD (WXDXH) IN: 14.00 x 14.00 x 46.00 LBS: 28.00 MM: 355.60 x 355.60 x 1168.40 KG: 12.71 LOCATION: FLOOR SPEC COMMENTS: IV POLE MOUNTABLE OR FREESTANDING. IV POLE DIAMETER RANGE OF POLE MOUNT: 1" - 1-1/4". DIMENSION WITH POLE WDXH 14X8X46 VOLTS: 115 HERTZ: 60 AMPS: 12.5 WATS: 1440 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: INTERNAL RECHARGEABLE BATTERY 20A AC RECEPTACLE, 3- PRONG	BELMONT INSTRUMENT CORPORATION 903-00039 RI-2	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
8		1014543 CPR01	1	COMPRESSION DEVICE CHEST COMPRESSION SYSTEM TO INCLUDE: 1EA LUCAS 3.0 CHEST COMPRESSION SYSTEM WITH, (WXDXH) IN: 21.00 x 22.00 x 10.00 LBS: 18.00 MM: 533.40 x 558.80 x 254.00 KG: 8.17 LOCATION: WALL VOLTS: 115 HERTZ: 60 AMPS: 2.3 WATS: 265 ELECT NOTE: FOR BATTERY CHARGING: 115VAC/60HZ/2.3A;CLASS II OUTPUT 24VDC;4.2A	STRYKER - PHYSIO-CONTROL, INC 99576-000043 LUCAS 3.0	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
9		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WXDXH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		006656 MAY01	2	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WXDXH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		1012767 CRS01	1	CABINET, SUPPLY STORAGE MOBILE STORAGE CABINET; STEEL CONSTRUCTION WITH POWDER COAT FINISH; FOUR ADJUSTABLE SHELVES; 400LB SHELF CAP; 750LB TOTAL (WDXH) IN: 48.00 x 24.00 x 84.00 LBS: 242.00 MM: 1219.20 x 609.60 x 2133.60 KG: 109.77 LOCATION: FLOOR NOTES: SEISMIC KIT CUSTOM UPON ORDER ANCHOR DETAIL REQ: YES	ULINE H-9092A CLEAR VIEW ASSEMBLED	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
12		072962 CRC03	1	CART, RESUSCITATION, CARDIAC PEDS;30";9 MULTICOLOR DRW;(8)3"DRAWERS;(1)6"DRAWER CARDIAC BOARD;STEEL BREAKAWAY LOCK;5"CASTERS (WDXH) IN: 46.09 x 25.00 x 45.50 LBS: 128.00 MM: 1170.69 x 635.00 x 1155.70 KG: 58.06 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE;	ARMSTRONG MEDICAL INDUSTRIES PREMIER BROSELOW PBL-PC-9	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
13		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
14		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICK302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		021273 KCB01	1	KICKBUCKET SS;REMOVABLE 12 QT. SS PAIL;TIP-RESISTANT;ROUND (WDXH) IN: 13.75 x 13.75 x 13.25 LBS: 8.00 MM: 349.25 x 349.25 x 336.55 KG: 3.63 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. P-1020-SS P-1020-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
16		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
17		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
18		1014992	2	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
19		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WXDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
20		1005731 CLK01	1	CLOCK, ELAPSED TIME SURFACE MOUNT 6 DIGIT; RED; 4' DIGIT HEIGHT; 4" SLOPE (WXDXH) IN: 23.50 x 2.40 x 8.00 LBS: 4.60 MM: 596.90 x 60.96 x 203.20 KG: 2.09 LOCATION: WALL NOTES: 2025.06.10 CHANGED FROM POE TO 120V PER ELECTRICAL FUTURE ITEM; ROUGH IN ONLY AT THIS TIME CHANGE CODE: C CHANGE NOTES: CHANGED FROM POE TO 120V VOLTS: 120 HERTZ: 60 ELECT NOTE: 10' CORD WITH PLUG	PRIMEX XRA7A203E LEVO SERIES	OF/CI	FUTURE	GROUP: 1 FIXED
21		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
22		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL NOTES: 1 SHARPS ON WALL; 1 SHARPS ON SUPPLY CART CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
23		1003160 DSPP03	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL LOCKING WALL BRACKET; WHITE/BLEU 3 GALLON PHARMASAFETY CONTAINER LEAK RESISTANT; TAMPER EVIDENT; SECURE HINGED LID; ABSORBENT PA (WXDXH) IN: 13.75 x 6.50 x 16.50 LBS: 5.00 MM: 349.25 x 165.10 x 419.10 KG: 2.27 LOCATION: WALL SPEC COMMENTS: DIMENSIONS OF BRACKET: 13.75" X 6.5" X 11.75" (W X D X H)	COVIDIEN 8528C/8836SA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 05, TRAUMA
ROOM NUMBER: 317


LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
24		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
25		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
26		1023059	1	I.V. INFUSION PUMP ACCESSORY 1000ML (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.05.12 ADDED METPAK 5275 PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	RIESTER 5275	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 04, BEHAVIORAL HEALTH
ROOM NUMBER: 318

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1	No image available	1014542 LHT04	1	LIGHT, EXAM, MOBILE EXAMINATION LIGHT (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: FLOOR VOLTS: 120 HERTZ: 60	SUNNER TECHNOLOGIES TS20	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WDXDH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR PHASE: EARLY-PO/LEASED NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON ELECTRIC, STANDARD MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXDH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 04, BEHAVIORAL HEALTH
ROOM NUMBER: 318

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED
5		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY; WALLBOARD; OPPTH/OTO; SPHYG; TEMP TO INCLUDE: (WDXDH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 04, BEHAVIORAL HEALTH
ROOM NUMBER: 318

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		006656 MAY01	1	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXDH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
10		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXDH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXDH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 04, BEHAVIORAL HEALTH
ROOM NUMBER: 318

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
13		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
14		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
16		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL NOTES: 1 SHARPS ON WALL; 1 SHARPS ON SUPPLY CART CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 04, BEHAVIORAL HEALTH
ROOM NUMBER: 318

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 03, ISOLATION
ROOM NUMBER: 319

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023169 LHT01	1	LIGHT, EXAM LED;SINGLE LIGHT HEAD;ARM 1 DROP TUBE LENGTH: 750MM;ARM 1 LENGTH: 800MM (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ANCHOR DETAIL REQ: YES ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL SIRCA	OF/CI	NEW	GROUP: 1 FIXED
2		1023695 STR08*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. ELECTRIC FUNCTIONALITY. 700 LB WEIGHTT CAPACITY (WXDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR PHASE: EARLY-PO/LEASED NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, ELECTRIC, GEL MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 03, ISOLATION
ROOM NUMBER: 319

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED
5		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY; WALLBOARD; OPHTH/OTO; SPHYG; TEMP TO INCLUDE: (WDXDH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 03, ISOLATION
ROOM NUMBER: 319

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		006656 MAY01	1	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 03, ISOLATION
ROOM NUMBER: 319

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
13		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
15		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL NOTES: 1 SHARPS ON WALL; 1 SHARPS ON SUPPLY CART CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
16		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 03, ISOLATION
ROOM NUMBER: 319

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		053947 CRI01	1	CART, ISOLATION 4 DRAWER;YELLOW;BASIC CART W/DRAWER; CONFIGURATION OF 1 EA.3" (WXDXH) IN: 33.50 x 25.00 x 46.25 LBS: 165.00 MM: 850.90 x 635.00 x 1174.75 KG: 74.85 LOCATION: FLOOR	ARMSTRONG MEDICAL INDUSTRIES AKL-Y-4 TWO TONE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 02
ROOM NUMBER: 320

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023169 LHT01	1	LIGHT, EXAM LED;SINGLE LIGHT HEAD;ARM 1 DROP TUBE LENGTH: 750MM;ARM 1 LENGTH: 800MM (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ANCHOR DETAIL REQ: YES ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL SIRCA	OF/CI	NEW	GROUP: 1 FIXED
2		1023695 STR08*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. ELECTRIC FUNCTIONALITY. 700 LB WEIGHTT CAPACITY (WDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, ELECTRIC, GEL MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 02
ROOM NUMBER: 320

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED
5		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY; WALLBOARD; OPPTH/OTO; SPHYG; TEMP TO INCLUDE: (WDXDH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 02
ROOM NUMBER: 320

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		006656 MAY01	1	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXDH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
10		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXDH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXDH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 02
ROOM NUMBER: 320

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
13		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
14		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
16		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 02
ROOM NUMBER: 320

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 01
ROOM NUMBER: 321

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023169 LHT01	1	LIGHT, EXAM LED;SINGLE LIGHT HEAD;ARM 1 DROP TUBE LENGTH: 750MM;ARM 1 LENGTH: 800MM (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ANCHOR DETAIL REQ: YES ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL SIRCA	OF/CI	NEW	GROUP: 1 FIXED
2		1023695 STR08*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. ELECTRIC FUNCTIONALITY. 700 LB WEIGHTT CAPACITY (WXDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, ELECTRIC, GEL MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1015743 PHY04	1	MONITOR, PHYSIOLOGICAL 12" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WXDXH) IN: 12.80 x 7.20 x 11.30 LBS: 12.30 MM: 325.12 x 182.88 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866062_NAM INTELLIVUE MX450	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 01
ROOM NUMBER: 321

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXDH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED
5		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPERATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		069814 DIAG01	1	DIAGNOSTIC SET GREEN TECHNOLOGY; WALLBOARD; OPPTH/OTO; SPHYG; TEMP TO INCLUDE: (WDXDH) IN: 34.00 x 6.00 x 12.00 LBS: 25.00 MM: 863.60 x 152.40 x 304.80 KG: 11.34 LOCATION: WALL VOLTS: 120 HERTZ: 60 AMPS: 18	WELCH ALLYN, INC. 77791-1MPX GS 777 SERIES	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 01
ROOM NUMBER: 321

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		006656 MAY01	1	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXDH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
10		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXDH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXDH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 01
ROOM NUMBER: 321


LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
13		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
14		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
16		087035 DSP01	2	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL NOTES: 1 SHARPS ON WALL; 1 SHARPS ON SUPPLY CART CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: TREATMENT ROOM 01
ROOM NUMBER: 321

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		084252		1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		089875		1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED ALCOVE 01
ROOM NUMBER: 325
ROOM COMMENT: ALCOVE LOCATED OUTSIDE OF TRAUMA ROOM

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		056002 CRC04	1	CART, RESUSCITATION, CARDIAC CART, RESUSCITATION, CARDIAC; 6 DRAWERS; 4-3", 1-6", 1-12", PVC BUMPER; LOCKING BREAKAWAY BAR, TWIN (WXDXH) IN: 43.00 x 25.00 x 47.00 LBS: 165.00 MM: 1092.20 x 635.00 x 1193.80 KG: 74.85 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE;	ARMSTRONG MEDICAL INDUSTRIES AR-6 PREMIER W/ER PK	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		082533 DEF01	1	DEFIBRILLATOR R SERIES ALS DEFIBRILLATOR; AED W/MANUAL OVERRIDE; GUIDELINES 2010 COMPATIBLE; CODE READINESS TESTING SYSTEM; RECTILINEAR (WXDXH) IN: 10.50 x 12.50 x 8.20 LBS: 15.20 MM: 266.70 x 317.50 x 208.28 KG: 6.90 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: OPTIONAL WIFI: 802.11B USB PORT 2.0 COMPATIBLE VOLTS: 120 HERTZ: 60 ELECT NOTE: SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER EMERGENCY POWER	ZOLL MEDICAL CORPORATION 30720005201310012 R SERIES ALS	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
3		094808 ASP01	1	PUMP, SUCTION/ASPIRATOR, GENERAL, PORTABLE PORTABLE; RECHARGEABLE BATTERY; LED DISPLAY; HEPA FILTER; BEMIS DISPOSABLE CANISTER SYSTEM; HIGH FLOW; HIGH VACUUM; (WXDXH) IN: 13.00 x 6.30 x 12.40 LBS: 8.90 MM: 330.20 x 160.02 x 314.96 KG: 4.04 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 ELECT NOTE: ELECTRICAL SHOWN FOR BATTERY CHARGER	LAERDAL MEDICAL CORP. LSU 4000 78002001	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
4		098605 RCK01	1	RACK, APRON, WALL MOUNTED 8 PEGS; 4 APRONS; ELECTRO-MAGNETIC POWDER COATING; STAINLESS STEEL (WXDXH) IN: 19.00 x 5.00 x 3.00 LBS: 7.20 MM: 482.60 x 127.00 x 76.20 KG: 3.27 LOCATION: WALL	PHILLIPS SAFETY PRODUCTS RAR-P4	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: ED ALCOVE 01
ROOM NUMBER: 325
ROOM COMMENT: ALCOVE LOCATED OUTSIDE OF TRAUMA ROOM

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1023154	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL;RECTANGULAR SHIELD TO BE PLACED ON PATIENT DURING PROCEDURES; CHOOSE SIZE IN (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE <i>NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED</i> CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BLOCKERS-SB	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		1023150	1	APRON, LEAD FREE XENOLITE 600-B -0.35MM PROTECTIVE MATERIAL; ONE-PIECE; OPENING IN BACK; STRETCH HOOK-AND-LOOP CLOSURE; FOAM SHOULDER (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE <i>NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED</i> CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS - EURO WRAPAROUND ECEURO-SB-B-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: EMERGENCY
ROOM NAME: AMBULANCE CANOPY
ROOM NUMBER: ZZZ

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		094151 CAB06	1	CABINET, STORAGE SS STORAGE CABINET;5 ADJUSTABLE SHELVES;LOCKABLE; DOLLY ADDS 5" TO HEIGHT AND 25 LB IN WEIGHT. TO INCLUDE: (WXDXH) IN: 36.00 x 24.00 x 73.00 LBS: 260.00 MM: 914.40 x 609.60 x 1854.20 KG: 117.94 LOCATION: FLOOR NOTES: SEISMIC: CUSTOM ORDER THROUGH COMPANY AT TIME OF PURCHASE	ULINE H-5588	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1023695 STR08*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. ELECTRIC FUNCTIONALITY. 700 LB WEIGHTT CAPACITY (WXDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRETCHER MAKE/MODEL TBC - THIS ONE LIKELY STRYKER ZOOM DRIVE AS I NEEDS TO BE POWERED CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: MANAGER OFFICE
ROOM NUMBER: 506

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: VIEWING / TECH WORK
ROOM NUMBER: 507

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		092852	1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WXDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: PATIENT TOILET
ROOM NUMBER: 508

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: ULTRASOUND
ROOM NUMBER: 509

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015461 IUS02	1	IMAGING, ULTRASOUND SYSTEM TO INCLUDE: 1 APLIO I700 PRISM ULTRASOUND SYSTEM (WXDXH) IN: 24.80 x 42.10 x 69.20 LBS: 253.50 MM: 629.92 x 1069.34 x 1757.68 KG: 114.99 LOCATION: FLOOR SPEC COMMENTS: ;SEE QUOTE FOR ADDITIONAL INFORMATION PHASE: EARLY-PO/LEASED DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: ETHERNET 10BASE-T/ 100BASE-TX/ GIGABIT; INTERNAL WIRELESS LAN (FIPS 140-2 COMPLIANT); DUAL BAND 2.4GHZ/5 GHZ VOLTS: 115 HERTZ: 60 ELECT NOTE: POWER CONSUMPTION 1440 VA PO REQ #: 944A-00001 CLIENT PO #: SNCA417	CANON MEDICAL SYSTEMS USA, INC. APLIO-I700-V6.5.100 I700 PRISM	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		1020574 STR06*	1	STRETCHER OB/GYN;4" THICK MATTRESS;ADJ FOOTRESTS;26" WIDE PT SURFACE;RETRACTABLE FIFTH-WHEEL STEERING; (WXDXH) IN: 31.50 x 82.00 x 35.50 LBS: 250.00 MM: 800.10 x 2082.80 x 901.70 KG: 113.40 LOCATION: FLOOR NOTES: 2025.09.11 UPDATED TO STRYKER GYNNIE PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV GYNNIE 1061000000	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		105868 CHR06	1	CHAIR, ULTRASOUND ULTRASOUND CHAIR. FEATURES: BACKWARDS TILT, LOCKABLE SEAT ADJUSTABLE, SEAT HEIGHT ADJUSTABLE 23"-33", SEAT DEPTH (WXDXH) IN: 28.74 x 28.74 x 50.00 LBS: 30.86 MM: 730.00 x 730.00 x 1270.00 KG: 14.00 LOCATION: FLOOR	HAG INC 8106 CAPISCO	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: ULTRASOUND
ROOM NUMBER: 509

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		072156 CAB03	1	CABINET, STORAGE, TRANSDUCER/PROBE 120V 3 PROBE POLYPROPYLENE CABINET; ACFHEPA-188 (WXDXH) IN: 26.00 x 10.00 x 88.00 LBS: 115.10 MM: 660.40 x 254.00 x 2235.20 KG: 52.21 LOCATION: FLOOR NOTES: 2025.01.20 MOUNTING DETAILS AND HARDWARE PROVIDED BY VENDOR. SEISMIC DETAILS AND INSTALL TO BE PROVIDED BY PROJECT ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CHANGED FROM OFOI TO OFCI VOLTS: 120 HERTZ: 60 AMPS: 0.7 WATS: 84	CS MEDICAL AC-TE-03 AC-TE-03	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
5		089695 WRM02	1	WARMER, GEL ULTRASOUND GEL WARMER; 3 BOTTLE; ADJUSTABLE TEMP; LCD TEMP (WXDXH) IN: 8.00 x 5.00 x 12.00 LBS: 3.00 MM: 203.20 x 127.00 x 304.80 KG: 1.37 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 0.26 WATS: 31 PLUG TYPE: Type B (NEMA 5-15)	PARKER LABORATORIES, INC. 83-03 (LCD TEMP) THERMASONIC	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
6		101226 DIS01	1	DISINFECTOR ULTRASOUND TRANSDUCER DISINFECTOR;7 MIN CYCLE TIME;ELECTRONIC CYCLE LOGGING;W/WALL-MOUNTED PRINTER;(PRINTER IS 5X9X7") (WXDXH) IN: 14.20 x 12.50 x 21.00 LBS: 56.90 MM: 360.68 x 317.50 x 533.40 KG: 25.81 LOCATION: WALL SPEC COMMENTS: CONTRACTOR TO INSTALL WALL MOUNT DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: TCP/IP PROTOCOL; ETHERNET CONNECTOR RJ45; USB PORT: TYPE A VOLTS: 120 HERTZ: 60 AMPS: 6 WATS: 720 PLUG TYPE: Type B (NEMA 5-15)	GE MEDICAL SYSTEMS E8350MA TROPHON 2	OF/CI	NEW	GROUP: 1 FIXED
7		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: ULTRASOUND
ROOM NUMBER: 509

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
9		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
10		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
12		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: ULTRASOUND
ROOM NUMBER: 509

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		1001188 CBW01	1	CABINET, WARMING 2.5 CUBIC FT; 6 BLANKET CAPACITY; COUNTERTOP MODEL; STACKABLE; NO (WXDXH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP SPEC COMMENTS: CLEARANCES: TOP/LEFT/RIGHT: 1", BACK 3" NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO ANCHOR DETAIL REQ: NO CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 AMPS: 5 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 0.6KW;15A;125V PLUG;HOSPITAL GRADE PLUG	ENTHERMICS MEDICAL SYSTEMS TITAN SERIES EC250 BTU: 2050	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
14		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR NOTES: 2025.09.11 ADDED TO LIST PER SHD CHANGE CODE: A CHANGE NOTES: ADDED TO LIST PER SHD	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: IMG DRESS
ROOM NUMBER: 511

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM


FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY
ROOM NUMBER: 512

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1022578 IRD02*	1	IMAGING, RADIOGRAPHIC SYSTEM AUTO POSITIONING SYSTEM;X-RAY SYSTEM W/OTC+WALLSTAND+TABLE +SYSTEM CABINET TO INCLUDE: (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: REFERENCE VENDOR'S TECHNICAL DATA DEDICATED CIRCUIT: YES ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00003 CLIENT PO #: SNCA418	CANON CANADA, INC. OMNERA500A.100 OMNERA 500A	OF/VI	NEW	GROUP: 1 FIXED
2		098605 RCK01	1	RACK, APRON, WALL MOUNTED 8 PEGS;4 APRONS;ELECTRO-MAGNETIC POWDER COATING;STAINLESS STEEL (WDXDH) IN: 19.00 x 5.00 x 3.00 LBS: 7.20 MM: 482.60 x 127.00 x 76.20 KG: 3.27 LOCATION: WALL	PHILLIPS SAFETY PRODUCTS RAR-P4	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		1001743	4	APRON, LEAD COMFORTLITE WAIST ADJUSTABLE APRON; 0.35MM, 0.5MM LEAD EQUIV; VELCRO CLOSURES; ELASTIC BACK SECTIONS (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE	AADCO MEDICAL, INC. LP-CX--- COMFORTLITE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
4		1023151	1	APRON, LEAD FREE XENOLITE 600-B -0.35MM PROTECTIVE MATERIAL; ONE-PIECE; OPENING IN BACK; STRETCH HOOK-AND-LOOP CLOSURE; FOAM SHOULDER (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS CONFIRMED CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS - EURO WRAPAROUND ECEURO-SB-C-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY
ROOM NUMBER: 512

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1023150	1	APRON, LEAD FREE XENOLITE 600-B -0.35MM PROTECTIVE MATERIAL; ONE-PIECE; OPENING IN BACK; STRETCH HOOK-AND-LOOP CLOSURE; FOAM SHOULDER (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS - EURO WRAPAROUND	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		1023154	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL;RECTANGULAR SHIELD TO BE PLACED ON PATIENT DURING PROCEDURES; CHOOSE SIZE IN (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BLOCKERS-SB	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
8		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY
ROOM NUMBER: 512

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		006717 RES01	1	RESTRAINT, IMMOBILIZER, CHILD SMALL;FOR X-RAY USE;INFANT TO 3 1/2 YEARS;ADJUSTABLE SEAT PLATFORM TO FACILITATE OBLIQUE AND LATERAL VIEWS;SPRING BALANCED (WXDXH) IN: 20.00 x 22.00 x 35.00 LBS: 48.00 MM: 508.00 x 558.80 x 889.00 KG: 21.78 LOCATION: FLOOR	MXR/SOURCEONE 292205/MODERN WAY PIGG-O-STAT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1003609	1	ACCESSORY GENERAL SPONGE KIT;LIGHT GREEN;POLYURETHANE FOAM TO INCLUDE: (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: ASSORTED MEASUREMENTS	TECHNO-AIDE STEALTH-COTE SPONGER KIT 127595	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		1020084 BRD01A	1	ACCESSORY, WALL MOUNT WALL MOUNT FOR PATIENT TRANSFER BOARD (WXDXH) IN: 5.50 x 1.75 x 3.00 LBS: 0.02 MM: 139.70 x 44.45 x 76.20 KG: 0.01 LOCATION: WALL SPEC COMMENTS: ; NOTES: 2024-10-31 - ADDED FOR BOARD, PATIENT TRANSFER ID # 1020086 CHANGE CODE: A CHANGE NOTES: ADDED WALL MOUNT FOR TRASFER BOARD	MJM INTERNATIONAL 618-HDPE	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
13		1020086 BRD01B	1	BOARD, PATIENT TRANSFER 500LB WEIGHT CAP; PVC TRANSFER BOARD; EASY TO CLEAN (WXDXH) IN: 35.50 x 70.00 x 1.25 LBS: 26.00 MM: 901.70 x 1778.00 x 31.75 KG: 11.80 LOCATION: SPACE SPEC COMMENTS: PURCHASE WALL MOUNT 1 EA. PVC618HDP WALL BRACKET SEPARATELY FOR CONTRACTOR INSTALL CHANGE CODE: C CHANGE NOTES: CHANGED TO OF/OI, CHANGED TO "SPACE"	MEDLINE INDUSTRIES, INC. PVC618HDP PVC618HDP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY
ROOM NUMBER: 512

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
14		1004559	2	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
16		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WXDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
17		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WXDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY
ROOM NUMBER: 512

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
19		094275		1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: GENERAL RADIOLOGY CONTROL
ROOM NUMBER: 513

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015460 ICT01	1	IMAGING, CT SYSTEM TO INCLUDE: 1 PRIME SP 80 DESCRIPTION V10 (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: FLOOR SPEC COMMENTS: ;OVERALL: WIDTH: 7'-0 5/8" DEPTH: 17'-6 3/4" HEIGHT: 6'-5 1/2" PATIENT COUCH WIDTH: 2'-2" DEPTH: 9'-5 3/4" HEIGHT: 3'-4 3/4" GANTRY WIDTH: 8'-11 1/16" DEPTH: 2'-10 13/16" HEIGHT: 6'-2" TO 6'-10" ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO VOLTS: 480 HERTZ: 60 AMPS: 100 ELECT NOTE: REFERENCE VENDORS TECHNICAL DATA PO REQ #: 944A-00004 CLIENT PO #: SNCA419	CANON MEDICAL SYSTEMS USA, INC. AQUILION PRIME PRIME SP10.000	OF/VI	NEW	GROUP: 1 FIXED
					SELLER: CANON MEDICAL SYSTEMS USA, INC.			

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
2		106403 INJ01	1	INJECTOR DUAL INJECTION SYRINGE HEAD; LED DISPLAY TO INCLUDE: (WXDXH) IN: 12.10 x 9.00 x 25.80 LBS: 30.40 MM: 307.34 x 228.60 x 655.32 KG: 13.79 LOCATION: CEILING SPEC COMMENTS: ;REFERENCE VENDOR'S TECHNICAL DATA FOR ARM LENGTHS AND INSTALLATION INFORMATION BASE UNIT: 11.0X9.0X11.3 WDXH IN INCHES, 13.6LBS. WORKSTATION: 15.1X7.9X10.8 WDXH IN INCHES, 15.8LBS ANCHOR DETAIL REQ: YES DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: OPTIONAL VIRTUAL-CARE REMOTE SUPPORT VOLTS: 120 HERTZ: 60 AMPS: 2.5 WATS: 300	MEDRAD/BAYER HEALTHCARE STELLANT FLEX	OF/VI	NEW	GROUP: 1 FIXED
3		056002 CRC04	1	CART, RESUSCITATION, CARDIAC CART, RESUSCITATION, CARDIAC; 6 DRAWERS;4-3", 1-6",1-12",PVC BUMPER;LOCKING BREAKAWAY BAR, TWIN (WXDXH) IN: 43.00 x 25.00 x 47.00 LBS: 165.00 MM: 1092.20 x 635.00 x 1193.80 KG: 74.85 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE; NOTES: PLACE ON NORTH WALL; SHARED BETWEEN CT AND XRAY ROOMS	ARMSTRONG MEDICAL INDUSTRIES AR-6 PREMIER W/ER PK	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		082533 DEF01	1	DEFIBRILLATOR R SERIES ALS DEFIBRILLATOR;AED W/MANUAL OVERRIDE;GUIDELINES 2010 COMPATIBLE;CODE READINESS TESTING SYSTEM;RECTILINEAR (WXDXH) IN: 10.50 x 12.50 x 8.20 LBS: 15.20 MM: 266.70 x 317.50 x 208.28 KG: 6.90 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: OPTIONAL WIFI: 802.11B USB PORT 2.0 COMPATIBLE VOLTS: 120 HERTZ: 60 ELECT NOTE: SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER EMERGENCY POWER	ZOLL MEDICAL CORPORATION 30720005201310012 R SERIES ALS	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		094808 ASP01	1	PUMP, SUCTION/ASPIRATOR, GENERAL, PORTABLE PORTABLE; RECHARGEABLE BATTERY; LED DISPLAY; HEPA FILTER; BEMIS DISPOSABLE CANISTER SYSTEM; HIGH FLOW; HIGH VACUUM; (WDXH) IN: 13.00 x 6.30 x 12.40 LBS: 8.90 MM: 330.20 x 160.02 x 314.96 KG: 4.04 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 ELECT NOTE: ELECTRICAL SHOWN FOR BATTERY CHARGER	LAERDAL MEDICAL CORP. LSU 4000 78002001	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
6		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
7		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1004558	1	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		1014992	1	REGULATOR, SUCTION, CONTINUOUS & INTERMITTENT ANALOG CONTINUOUS/INTERMITTENT VACUUM REGULATOR WITH OHMEDA DIAMOND ADAPTER AND TUBING NIPPLE PATIENT CONNECTION. PUSH-TO-SE (WDXH) IN: 2.80 x 4.80 x 6.50 LBS: 0.00 MM: 71.12 x 121.92 x 165.10 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION 8701-1251-901 PUSH-TO-SET	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
11		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
13		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
14		066923 WRM01	1	WARMER, CONTRAST MEDIA .65 CF;BI-METALLIC THERMOSTAT;2 FIXED SHELVES; EXTERIOR STEEL CONSTRUCTION; POWDER COAT FINISH (WDXH) IN: 10.20 x 18.50 x 16.30 LBS: 16.00 MM: 259.08 x 469.90 x 414.02 KG: 7.26 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60	CONE INSTRUMENTS 216105	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
15		1020084 BRD01A	1	ACCESSORY, WALL MOUNT WALL MOUNT FOR PATIENT TRANSFER BOARD (WXDXH) IN: 5.50 x 1.75 x 3.00 LBS: 0.02 MM: 139.70 x 44.45 x 76.20 KG: 0.01 LOCATION: WALL SPEC COMMENTS: ; NOTES: 2024-10-31 - ADDED FOR BOARD, PATIENT TRANSFER ID # 1020086 CHANGE CODE: A CHANGE NOTES: ADDED WALL MOUNT FOR TRASFER BOARD	MJM INTERNATIONAL 618-HDPE	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
16		1020086 BRD01B	1	BOARD, PATIENT TRANSFER 500LB WEIGHT CAP; PVC TRANSFER BOARD; EASY TO CLEAN (WXDXH) IN: 35.50 x 70.00 x 1.25 LBS: 26.00 MM: 901.70 x 1778.00 x 31.75 KG: 11.80 LOCATION: SPACE SPEC COMMENTS: PURCHASE WALL MOUNT 1 EA. PVC618HDP WALL BRACKET SEPARATELY FOR CONTRACTOR INSTALL CHANGE CODE: C CHANGE NOTES: CHANGED TO OF/OI, CHANGED TO "SPACE"	MEDLINE INDUSTRIES, INC. PVC618HDP PVC618HDP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
17		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT SCAN
ROOM NUMBER: 514

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
19		1003160 DSPP03	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL LOCKING WALL BRACKET;WHITE/BUE 3 GALLON PHARMASAFETY CONTAINER LEAK RESISTANT;TAMPER EVIDENT;SECURE HINGED LID;ABSORBENT PA (WDXH) IN: 13.75 x 6.50 x 16.50 LBS: 5.00 MM: 349.25 x 165.10 x 419.10 KG: 2.27 LOCATION: WALL SPEC COMMENTS: DIMENSIONS OF BRACKET: 13.75" X 6.5" X 11.75" (W X D X H)	COVIDIEN 8528C/8836SA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT CONTROL
ROOM NUMBER: 515

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1	No image available	069542 ICR01	1	IMAGING, READER, CR TABLE TOP;IP CASSETTES;UP TO 57 CASSETTES PER HOUR (WXDXH) IN: 22.00 x 21.00 x 15.00 LBS: 86.00 MM: 558.80 x 533.40 x 381.00 KG: 39.01 LOCATION: COUNTERTOP ELECT NOTE: 120V;60HZ;1.9 AMPS	FUJIFILM MEDICAL SYSTEMS USA, INC. FCR PRIMA T CR1R-392	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1001188 CBW01	1	CABINET, WARMING 2.5 CUBIC FT; 6 BLANKET CAPACITY; COUNTERTOP MODEL; STACKABLE; NO (WXDXH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP SPEC COMMENTS: CLEARANCES: TOP/LEFT/RIGHT: 1", BACK 3" NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 5 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 0.6KW;15A;125V PLUG;HOSPITAL GRADE PLUG	ENTHERMICS MEDICAL SYSTEMS TITAN SERIES EC250 BTU: 2050	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
3		098605 RCK01	1	RACK, APRON, WALL MOUNTED 8 PEGS;4 APRONS;ELECTRO-MAGNETIC POWDER COATING;STAINLESS STEEL (WXDXH) IN: 19.00 x 5.00 x 3.00 LBS: 7.20 MM: 482.60 x 127.00 x 76.20 KG: 3.27 LOCATION: WALL	PHILLIPS SAFETY PRODUCTS RAR-P4	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
4		1001743	4	APRON, LEAD COMFORTLITE WAIST ADJUSTABLE APRON; 0.35MM, 0.5MM LEAD EQUIV; VELCRO CLOSURES; ELASTIC BACK SECTIONS (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE	AADCO MEDICAL, INC. LP-CX--- COMFORTLITE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: CT CONTROL
ROOM NUMBER: 515

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1023140	2	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; FRONT AND BACK PROTECTION; FRONT BUCKLE CLOSURE; SIZE: MALE MEDIUM (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFIRMED CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-REGULAR VEST ECVEST-SB-B-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		1023145	2	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; PARTIAL OVERLAP DESIGN; FRONT HOOK-AND-LOOP CLOSURE W/ADJUSTABLE 2" BUCKLE CLOSUR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFIRMED CHANGE CODE: C CHANGE NOTES: 2025.05.12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-KILT ECKILT-SB-B-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: IMAGING
ROOM NAME: ALCOVE, MOBILE XRAY
ROOM NUMBER: 516

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1016294 IMX01	1	IMAGING, X-RAY UNIT, MOBILE, BATTERY MOBILE DIGITAL X-RAY; 360 DEGREE ROTATION OF THE PORTABLE X AND Y AXIS; INCLUDES: (WXDXH) IN: 21.30 x 99.00 x 87.80 LBS: 1146.00 MM: 541.02 x 2514.60 x 2230.12 KG: 519.82 LOCATION: FLOOR SPEC COMMENTS: ; VOLTS: 120 HERTZ: 60 PO REQ #: 944A-00005 CLIENT PO #: SNCA420	CANON, USA, INC. SOLTUS-500	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: MICROBIOLOGY
ROOM NUMBER: 500

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1005344 BIO01	1	CABINET, SAFETY, BIOLOGICAL 3' BIOLOGICAL SAFETY CABINET W/ 10" SASH OPENING, UV LIGHT, SERVICE FIXTURE AND VACU-PASS PORTAL; ELECTRONICALLY (WXDXH) IN: 42.30 x 31.20 x 61.70 LBS: 510.00 MM: 1074.42 x 792.48 x 1567.18 KG: 231.34 LOCATION: FLOOR SPEC COMMENTS: INCLUDE INSIDE DELIVERY AND SET-UP REQUIRES 3 RD PARTY HOOD CERTIFICATION CONFIRM VENTED OR NON-VENTED; <i>NOTES: 2024.08.12 VENDOR CONFIRMED SEISMIC REQUIRED. ADD OPTIONAL SEISMIC KIT (VENDOR SET IN PLACE, CONTRACTOR INSTALL SEISMIC). SENT INFO TO KPFF.</i> ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 12 WATS: 1380 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 10FT POWER CORD W/PLUG	LABCONCO CORPORATION 302311001_3 PURIFIER LOGIC+ (3 FT)	OF/VI	NEW	GROUP: 1 FIXED
2		068742 CEN04	1	LAB CENTRIFUGE, COUNTER TOP 8 SPEED SETTINGS;20-1060 XG FORCES;6 CYCLE TIMES (WXDXH) IN: 6.60 x 8.60 x 5.30 LBS: 5.50 MM: 167.64 x 218.44 x 134.62 KG: 2.50 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60	IRIS DIAGNOSTICS STATSPIN CYTOFUGE 2 CF02	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: MICROBIOLOGY
ROOM NUMBER: 500

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1001788 QUI01	1	LAB ANALYZER, CHEMISTRY, IMMUNO COUNTERTOP IMMUNOASSAY ANALYZER FOR USE WITH CASSETTE-BASED IMMUNOFLUORESCENT IN VITRO DIAGNOSTIC ASSAYS MANUFACTURED BY (WXDXH) IN: 4.75 x 8.75 x 4.75 LBS: 2.00 MM: 120.65 x 222.25 x 120.65 KG: 0.91 LOCATION: COUNTERTOP NOTES: 2025.08.11 REQUIRES CERNER INTERFACE 2025.0522 - UPDATED FROM EXISTING QUIDEL #20310 INCLUDES PROMO BUNDLE, TO NEW QUIDEL #20299 NO PROMO BUNDLE PER SENECA SS CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISTING TO NEW, CHANGED MODEL DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ45 LAN INTERFACE CONNECTOR VOLTS: 120 HERTZ: 60 AMPS: 2.93 WATS: 300 ELECT NOTE: ELECTRICAL REFLECTS POWER ADAPTER. SYSTEM COMES WITH PRE-INSTALLED RECHARGEABLE BATTERY.	QUIDEL 20299 SOFIA 2	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		071746 INC03	1	LAB INCUBATOR, COUNTER TOP 18L; 65CF; MECH (WXDXH) IN: 10.24 x 18.50 x 16.34 LBS: 15.90 MM: 260.10 x 469.90 x 415.04 KG: 7.22 LOCATION: COUNTERTOP NOTES: 2025.0526 CHANGED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: EXISTING TO NEW STATUS CHANGE VOLTS: 120 HERTZ: 60 AMPS: .45 PLUG TYPE: Type B (NEMA 5-15)	THERMO FISHER SCIENTIFIC 50125590 IMC18	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		1022426 LAB01	1	ANALYZER, LAB, MICROBIOLOGY STANDARDIZED, AUTOMATED PANEL INTERPRETATION. DATA MANAGEMENT SYSTEM. PROVIDES MIC FOR BLOOD AND URINE LEVELS WITH (WXDXH) IN: 19.00 x 23.00 x 10.00 LBS: 40.00 MM: 482.60 x 584.20 x 254.00 KG: 18.15 LOCATION: COUNTERTOP SPEC COMMENTS: DIMENSIONS ARE FOR ANALYZER ONLY: COMPUTER AND MONITOR: 17.3"W X 15.5"D KEYBOARD: 19.4"W X 8.3"D REPORT PRINTER: 18"W X 16"D	BECKMAN COULTER CORP. AUTOSCAN -4 B1018-502	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: MICROBIOLOGY
ROOM NUMBER: 500

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				NOTES: 2025-0811 REQUIRES NETWORK CONNECTION AND CERNER INTERFACE 2025-0616 UPDATED ID PER SENECA SS 6/6/2023: LOOK AT BUYING NEW DUE TO SERVICE CONTRACT INCLUDES PRO PC WORKSTATION CHANGE CODE: C CHANGE NOTES: UPDATED SPEC DATA HARDWIRED: YES WIRELESS: NO VOLTS: 115 HERTZ: 60 AMPS: 2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: INCLUDES PRO PC WORKSTATION AND PRINTER. ADDITIONAL ELECTRICAL OUTLETS REQUIRED FOR EACH POWER CORD LENGTH: 6.5 IN. CABLE TO COMPUTER LENGTH: 8.5 IN. DIGITAL INTERFACE: RS-232C (9600 BAUD)	BTU: 630			
6		1007249 LAB02	1	LAB ANALYZER, BACTERIA MICROSCAN TURBIDITY METER (WXDXH) IN: 6.00 x 4.00 x 4.00 LBS: 0.00 MM: 152.40 x 101.60 x 101.60 KG: 0.00 LOCATION: COUNTERTOP NOTES: 2025.06.20 UPDATED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: UPDATED FROM EXISTING TO NEW VOLTS: 120 HERTZ: 60	SIEMENS 5400-3126D AUTOSCAN TURBIDITY METER	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
7		1000226 LBM03	1	LAB MICROSCOPE LABORATORY MICROSCOPE; QUINTUPLE NOSE PIECE; 5 LENSES (WXDXH) IN: 12.00 x 15.00 x 18.00 LBS: 16.00 MM: 304.80 x 381.00 x 457.20 KG: 7.26 LOCATION: COUNTERTOP NOTES: 2025-0714 WILL LIVE IN CABINET VOLTS: 115 HERTZ: 60	OLYMPUS AMERICA INC. BH2	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
8		1023674 QIA02*	1	LAB ANALYZER, MOLECULAR, AUTOMATED 1 MODULE; ON-SITE MOLECULAR (PCR) POLYMERASE CHAIN REACTION) TESTING; HIGH THROUGHPUT, RANDOM ACCESS W/ABILITY TO RUN UP TO (WXDXH) IN: 15.00 x 16.00 x 14.00 LBS: 32.00 MM: 381.00 x 406.40 x 355.60 KG: 14.52 LOCATION: COUNTERTOP SPEC COMMENTS: CONTROL STATION (W X D X H): 8.52" X 12.77" X 3.82"; 18 LBS MODULES ONLY: 4.13"; 14 LBS EA.	BIOMERIEUX, INC. FAST-ASY-0001 SPOTFIRE	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: MICROBIOLOGY
ROOM NUMBER: 500

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				NOTES: 2025-0918 UPDATED MODEL TO SPOTFIRE WITH 1 MODULE PER SHD 2025-0811 REQUIRES CERNER INTERFACE 2025-0617 CHANGE FROM EXISTING TO X1 NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISTING TO NEW, CHANGED MODEL DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: ETHERNET OR WIFI CONENCTIVITY VOLTS: 120 HERTZ: 60				
9		063128 MIX01	1	LAB MIXER VARIABLE SPEED ANALOG MIXER; 100-3200 RPM; FOR FLASKS, BEAKERS, TUBES AND MICROWELL PLATES; (WDXH) IN: 5.50 x 8.00 x 4.80 LBS: 11.70 MM: 139.70 x 203.20 x 121.92 KG: 5.31 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 WATS: 150	THERMO FISHER SCIENTIFIC VORTEX MIXERS 02-215-365	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
10		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		053801 DSP02	1	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART;WALL ENCLOSURE W/CONTAINER INSERT (WDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL NOTES: 6/6/2023: PLACE NEAR MICROSCOPES	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: MICROBIOLOGY
ROOM NUMBER: 500

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
14		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: PATIENT TOILET (BLOOD DRAW)
ROOM NUMBER: 501

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: BLOOD DRAW
ROOM NUMBER: 502

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		069766 CRP01	1	CART, PHLEBOTOMY PHLEBOTOMY;4 DRAWERS;12 CLEAR TILT BINS; LOCKING SHARPS BRACKET AND WASTE DISPOSAL BIN INCLUDED;4 DIVIDABLE LOCKABLE (WDXH) IN: 27.00 x 24.50 x 37.50 LBS: 105.00 MM: 685.80 x 622.30 x 952.50 KG: 47.63 LOCATION: FLOOR	WATERLOO INDUSTRIES, INC. MSWA-3469-WHT MSWA-3469-WHT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		081174 CHR01	1	CHAIR, EXAM/TREATMENT PHLEBOTOMY/ENT CHAIR; 375LBS CAP; ADJUSTABLE SEAT HEIGHT (WDXH) IN: 31.50 x 67.00 x 51.00 LBS: 216.00 MM: 800.10 x 1701.80 x 1295.40 KG: 97.98 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 ELECT NOTE: 120V;60HZ	UMF MEDICAL 8000 SERIES 8678	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
3		1005432 CHR02	1	CHAIR, BLOOD DRAW ULTRA COMFORT HEAVY DUTY CHAIR DESIGNED WITH WIDER SEAT FOR UP 750LBS PATIENTS OR CHILD ON A PARENT'S LAP; PADDED ARM BOARD (WDXH) IN: 41.00 x 35.00 x 41.75 LBS: 78.00 MM: 1041.40 x 889.00 x 1060.45 KG: 35.39 LOCATION: FLOOR NOTES: 2025-0811 MATCH UPHOLSTERY COLOR OF CHR01 (LIGHT BLUE / GREY)	UMF MEDICAL 8690	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		088476 MAY02	1	STAND, MAYO SS; 5-LEG CAST ALUMINUM BASE; 29.5"-49" HEIGHT ADJ; 2" (WDXH) IN: 19.13 x 12.65 x 49.00 LBS: 14.00 MM: 485.91 x 321.31 x 1244.60 KG: 6.36 LOCATION: FLOOR SPEC COMMENTS: ;	BLICKMAN, INC. 0661501000 1501	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
5		053801 DSP02	2	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART;WALL ENCLOSURE W/CONTAINER INSERT (WDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: BLOOD DRAW
ROOM NUMBER: 502

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR NOTES: RED	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		057206 GLV01	2	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		055214 DCA01	1	LAB ANALYZER, CHEMISTRY, AUTOMATED MEASURES BLOOD/URINE HGBA1C & MICROALBUMIN; CALIBRATED W/BAR CODE CARD STORES UP TO 16 TEST RESULTS/16 CONTROL RESULTS (WXDXH) IN: 11.50 x 10.50 x 9.00 LBS: 8.55 MM: 292.10 x 266.70 x 228.60 KG: 3.88 LOCATION: COUNTERTOP NOTES: 2025-0811 REQUIRES NETWORK CONNECTION AND CERNER INTERFACE DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 4	SIEMENS MEDICAL SYSTEMS/CORPORATE 23-312-998 DCA VANTAGE	OF/VI	EXISTING	GROUP: 3 MINOR MOVABLE
1		107077 ABX01	1	ANALYZER, LAB, HEMATOLOGY COMPACT 5-PART DIFFERENTIAL ANALYZER; SINGLE SAMPLE ANALYSIS IN OPEN MODE; FULLY INTEGRATED IPU; LCD COLOR TOUCHSCREEN; 25 (WXDXH) IN: 13.78 x 17.72 x 18.12 LBS: 77.17 MM: 350.02 x 450.09 x 460.25 KG: 35.01 LOCATION: COUNTERTOP NOTES: 2025-0811 REQUIRES NETWORK CONNECTION AND CERNER INTERFACE 2025-0618 UPDATED FROM 2 TO 1 XN-550 PER SENECA INCLUDES WORKSTATION MONITOR AND PRINTER CHANGE CODE: R CHANGE NOTES: REDUCED QTY FROM 2 TO 1 DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 ELECT NOTE: 235VA	SYSMEX CORPORATION XN SERIES 350	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1003596 CLN01	1	LAB ANALYZER, URINE ANALYZER;AUTOMATIC CHECKS;STRIP TEST;CONNECTIVITY READY; TOUCHSCREEN;950 PATIENT TEST MEMORY;URINE, KIDNEY DISEASE AND (WXDXH) IN: 6.70 x 10.70 x 6.20 LBS: 3.65 MM: 170.18 x 271.78 x 157.48 KG: 1.66 LOCATION: COUNTERTOP SPEC COMMENTS: W/ OPTIONS: CLINITEK STATUS+ ANALYZER STARTER PACK NOTES: 2025-0811 REQUIRES CERNER INTERFACE 2025.05.21 CHANGED FROM EXISTING TO NEW PER LAB SS CHANGE CODE: C CHANGE NOTES: EXISTING TO NEW STATUS CHANGE DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RS-232 SERIAL PORT VOLTS: 120 HERTZ: 60 ELECT NOTE: 6 AA BATTERIES, NON-RECHARGEABLE	SIEMENS CLINITEK STATUS PLUS 1780 / STARTUA	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
1		1022886 FZP01	1	FREEZER, PLASMA PLASMA FREEZER. 4.9 CUBIC FEET/133 LITERS. -15Â°C TO -30Â°C. 3 ADJUSTABLE WIRE SHELVES. REMOTE ALARM TERMINALS. (WXDXH) IN: 24.00 x 26.00 x 33.40 LBS: 220.00 MM: 609.60 x 660.40 x 848.36 KG: 99.80 LOCATION: FLOOR NOTES: 2025.07.08 UPDATED FROM EXISTING TO NEW PER SENECA SS. MODEL PROVIDED BY MCKESSON 2024 - JEWETT DTPM PROBE TEMP MONITORS INCLUDED IN SPEC CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISTING TO NEW, MODEL UPDATED VOLTS: 115 HERTZ: 60 AMPS: 10 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 10' POWER CORD	THERMO FISHER SCIENTIFIC JPL430A 430A	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
1		098908 SYS01	1	LAB ANALYZER, COAGULATION, AUTOMATED COAGULATION ANALYZER FOR CLOTTING, CHROMOGENIC, AND IMMUNOLOGIC DETECTION METHODS. FEATURES (WXDXH) IN: 22.50 x 19.50 x 19.50 LBS: 94.60 MM: 571.50 x 495.30 x 495.30 KG: 42.91 LOCATION: COUNTERTOP SPEC COMMENTS: PROVIDE SPACE FOR 5L EXTERNAL WASTE AND WATER CONTAINERS.	SIEMENS MEDICAL SYSTEMS/CORPORATE 10712039 SYSMEX CA-660	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				NOTES: 2025-0811 REQUIRES CERNER INTERFACE 2025.05.22 CHANGED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISITNG TO NEW DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: HOST CONNECTION: BIDIRECTIONAL RS-232C SERIAL PORT INTERFACE: ASTM, CA-100 PROTOCOL VOLTS: 120 HERTZ: 60 WATS: 300 DEDICATED CIRCUIT: YES				
2		1014128 EXL01	1	LAB ANALYZER, CHEMISTRY, AUTOMATED INTEGRATED CHEMISTRY AND IMMUNOASSAY ANALYZER. CHEMISTRY AND IMMUNOASSAY TECHNOLOGIES SHARING NO MAJOR COMPONENTS, TO HELP (WXDXH) IN: 80.07 x 36.77 x 62.99 LBS: 1675.50 MM: 2033.78 x 933.96 x 1599.95 KG: 760.00 LOCATION: FLOOR NOTES: 2025-0811 REQUIRES NETWORK CONNECTION AND CERNER INTERFACE 2025.0522 - UPDATED FROM EXISTING EXL 200 TO NEW ATELICA CI 1900 6/6/2023: OEM RELOCATION, RECERT. DEDICATED CIRCUIT NEEDED, DRAIN NEEDED. FUTURE 220V CHANGE CODE: C CHANGE NOTES: UPDATED FROM EXISTING EXL 200 TO NEW ATELICA DATA HARDWIRED: YES WIRELESS: NO VOLTS: 240 HERTZ: 60 WATS: 1530 PLUG TYPE: Locking-NEMA L6-30 ELECT NOTE: 30A	SIEMENS CI 1900 ATELLICA CI 1900	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
					BTU: 5210/hr		COLD WATER: YES TREATED WATER: YES DRAIN REQUIRED: YES PLUMB NOTE: 2.0-4.5 L/MIN AT TEMP OF 10-30° C. <33 L/H; SPECIAL REAGENT WATER. BACTERIAL (50 CFU/ML), TOTAL ORGANIC CARBON (500 PPB)	
2		097096 ABX02*	1	LAB ANALYZER, HEMATOLOGY, AUTOMATED AUTOMATED SAMPLE LOADER; INTEGRATED RACK SAMPLING; AUTOMATIC RERUN AND REFLEX (WXDXH) IN: 17.70 x 26.00 x 17.70 LBS: 125.00 MM: 449.58 x 660.40 x 449.58 KG: 56.70 LOCATION: COUNTERTOP SPEC COMMENTS: MONITOR: 10.5X9.5X8.25 WXDXH IN INCHES RECOMMENDED TABLE SIZE: 76" LONG X 30" DEEP X 33" HIGH REFERENCE VENDOR'S TECHNICAL DATA FOR FULL DETAILS	SYSMEX CORPORATION XN-550 XN-550	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

COLD WATER: YES
TREATED WATER: YES
DRAIN REQUIRED: YES
PLUMB NOTE: 2.0-4.5 L/MIN AT TEMP OF 10-33 L/H; SPECIAL REAGENT WATER. BACTERIAL (50 CFU/ML), TOTAL ORGANIC CARBON (500 PPB)

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				NOTES: 2025-0811 REQUIRES CERNER INTERFACE 2025-0618 ADDED 1 XN-550 PER SENECA INCLUDES WORKSTATION MONITOR CHANGE CODE: A CHANGE NOTES: ADDED ANALYZER TO LIST DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: TWO ETHERNET (RJ45 JACK) CONNECTIONS TO LAN (INCLUDING PATCH CABLES) VOLTS: 115 HERTZ: 60 AMPS: 20 WATS: 800 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 1PH; DEDICATED 15 - 20 AMP CIRCUIT BREAKER REQUIRED. TWO 15-20 AMP RECEPTACLES REQUIRED.				
2		096137 CEN05	1	LAB CENTRIFUGE, COUNTER TOP URINALYSIS PACKAGE 4: 1-CENTRIFUGE, 1-BUILT-IN, ANGLE ROTOR (WXDXH) IN: 9.00 x 10.30 x 13.90 LBS: 22.00 MM: 228.60 x 261.62 x 353.06 KG: 9.98 LOCATION: COUNTERTOP NOTES: 2025.0522 - UPDATED FROM EXISTING HORIZON 642VES TO NEW HETTICH EBA 200 PER SENECA SS VOLTS: 115 HERTZ: 60 AMPS: 1.8	HETTICH 200URINALYSIS4 EBA 200	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1023325 DATA01*	1	DATA LOGGER, TEMPERATURE PROBE WIFI DATA LOGGER. CONFIGURABLE ALARM LEVELS. ALARMS AND DATA VIEWABLE ON BROWSER OR MOBILE APP. INCLUDES: WALL BRACKET (WXDXH) IN: 2.75 x 1.41 x 3.22 LBS: 0.00 MM: 69.85 x 35.82 x 81.79 KG: 0.00 LOCATION: SPACE NOTES: 2025-0811 UPDATED TO VFC 311 PER SHD 2025.07.08 ADDED PER SENECA SS CHANGE CODE: A CHANGE NOTES: ADDED TEMP MONITOR PER SENECA VOLTS: 120 HERTZ: 60	CONTROL CORPORATION OF AMERICA VFC 311		NEW	GROUP: 2 MAJOR MOVABLE
2		1007220 CEN02	1	LAB CENTRIFUGE, COUNTER TOP 6 TUBES CAPACITY; 5500 RPM; 20 SECOND ACCELL/DECEL; 1-30 (WXDXH) IN: 11.00 x 13.50 x 8.00 LBS: 12.00 MM: 279.40 x 342.90 x 203.20 KG: 5.45 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 WATS: 220	DRUCKER COMPANY 00-076-009-004 DASH APEX 6	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1012828 ALE01	1	ANALYZER, LAB, CHEMISTRY/DRUG, POINT-OF-CARE POINT-OF-CARE CHEMISTRY/DRUG LAB ANALYZER. FEATURES 16 IMMUNOASSAYS AVAILABLE WITH RESULTS IN ABOUT 15 MINUTES USING WHOLE (WDXDH) IN: 6.25 x 8.50 x 2.75 LBS: 2.00 MM: 158.75 x 215.90 x 69.85 KG: 0.91 LOCATION: COUNTERTOP NOTES: 2025-0811 REQUIRES CERNER INTERFACE 2025.06.02 UPDATED FROM EXISITNG TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: UPDATED FROM EXISITNG TO NEW DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RS-232 COMPUTER INTERFACE PORT VOLTS: 120 HERTZ: 60 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 6V DC AT 1 AMP – SUPPLIED VIA 4 AA BATTERIES OR AC/DC CONVERTER	QUIDEL TRIAGE METERPRO	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1022618 THD01*	1	WORK TABLE, MOBILE HEAVY DUTY LAB WORKBENCH. MOBILE. ADJUSTABLE. STAINLESS STEEL TOP. 72" W X 30" D (WDXDH) IN: 72.00 x 30.00 x 37.00 LBS: 0.00 MM: 1828.80 x 762.00 x 939.80 KG: 0.00 LOCATION: FLOOR NOTES: 2025-0618 ADDED WORKBENCH FOR ABX01A AND ABX01B, PRINTER TO LIVE ON LOWER SHELF CHANGE CODE: A CHANGE NOTES: ADDED TABLE TO LIST	WORKPLACE MODULAR SYSTEMS 1000 SERIES 30X72	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		098427 LBM02	1	LAB MICROSCOPE ERGO CERAMIC STG PKG;NIKON ECLIPSE CI-L MICROSCOPE FRAME;VARIABLE LED ILLUMINATION SYSTEM & CORD;ERGONOMIC TILTING AND (WDXDH) IN: 8.78 x 20.00 x 16.00 LBS: 30.00 MM: 223.02 x 508.00 x 406.40 KG: 13.61 LOCATION: COUNTERTOP NOTES: 2025-0616 UPDATED FROM NEW TO EXISTING PER SENECA SS CHANGE CODE: C CHANGE NOTES: NEW TO EXISTING STATUS CHANGE VOLTS: 120 HERTZ: 60 AMPS: .37 WATS: 6 ELECT NOTE: 3 CONDUCTOR GROUNDING TYPE SVT, NO. 18 AWG, 3M LONG MAX. RATED AT 125VAC	NIKON INSTRUMENTS INC. ECLIPSE CI-LED MBA950LP	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1023684 DATA02*	1	BRACKET, ALARM MOUNT, WALL WALL MOUNTING BRACKET IS COMPATIBLE WITH ALL EL-WIFI SENSOR UNITS. THE BRACKET FEATURES THREE MOUNTING HOLES TO ENSURE SECUR (WXDXH) IN: 3.00 x 2.00 x 4.00 LBS: 0.00 MM: 76.20 x 50.80 x 101.60 KG: 0.00 LOCATION: WALL NOTES: 2025.09.18 ADDED WALL MOUNT BRACKET FOR DATA LOGGER SEPARATELY FOR OF/CI. MOUNT ON WALL ABOVE COUNTER CHANGE CODE: A CHANGE NOTES: ADDED WALL MOUNT BRACKET FOR DATA LOGGER	CONTROL CORPORATION OF AMERICA EL-WIFI WALL MOUNT BRACKET	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
3		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1022131 WTS02*	1	WATER SYSTEM 70L TANK; 30 LITERS PER HOUR; 60-100 PSI; 5-35 DEGREES C; 295 LBS WET WEIGHT (WXDXH) IN: 15.00 x 23.00 x 44.00 LBS: 105.00 MM: 381.00 x 584.20 x 1117.60 KG: 47.63 LOCATION: FLOOR NOTES: 2025.06.02 UPDATED FROM EXISTING MW-15 TO NEW MW-30 PER SENECA FOR NEW ATELICA CHANGE CODE: C CHANGE NOTES: UPDATED SPEC, CHANGED FROM EXISITNG TO NEW VOLTS: 110 HERTZ: 60 WATS: 25	MED WATER MW 30 SERIES	OF/VI	NEW	GROUP: 1 FIXED
4		106606 GLO01	1	LAB ANALYZER, HEMATOLOGY, AUTOMATED AUTOMATED CELL COUNTER FOR CSF; CALCULATES LOW LEVEL TNC AND RBC RESULTS IN CELLS/UL; GENERATES RESULTS WITH ONLY 30 UL OF (WXDXH) IN: 6.00 x 8.00 x 10.00 LBS: 40.00 MM: 152.40 x 203.20 x 254.00 KG: 18.15 LOCATION: COUNTERTOP SPEC COMMENTS: INCLUDES: INSTRUMENT: 8X6X10 COMPUTER: 8X20X16 VACUUM STATION: 4.3X4.8X4.3 NOTES: 2025-0616 CONFIRMED EXISITNG PER SENECA SS CHANGE CODE: A CHANGE NOTES: ADDED TO EQUIPMENT LIST	ADVANCED INSTRUMENTS, INC. GLOCYTE GLOCYTE	OF/VI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: (2) USB 2.0 PORTS VOLTS: 120 HERTZ: 60 AMPS: 1 ELECT NOTE: UNIVERSAL AC/DC POWER SUPPLY 100-240VAC INSTRUMENT: 12 VDC/2A VACUUM STATION: 5 VDC/0.5A				
4		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WXDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR NOTES: RED	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		087033 REF03	1	REFRIGERATOR, BLOOD BANK 115V;4.9 CF;UNDERCOUNTER;SINGLE SOLID DOOR;2 SS DRAWERS;CHART RECORDER, 1°C TO +8°C (WXDXH) IN: 24.00 x 28.90 x 34.00 LBS: 212.00 MM: 609.60 x 734.06 x 863.60 KG: 96.17 LOCATION: FLOOR NOTES: 2025.06.02 UPDATED FROM EXISTING TO NEW CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISITNG TO NEW VOLTS: 115 HERTZ: 60 AMPS: 8 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: SINGLE PHASE;	THERMO FISHER SCIENTIFIC/JEWETT 30404J2A0ZZD0H0A JBB404A MECH NOTE: 727 BTU/HR AVG HEAT REJECTION	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1022428 LMG01	2	LIGHT, TASK, TABLE MOUNT, W/MAGNIFIER SPRING COUNTERBALANCE CONTROL LOCK KNOBS (WDXH) IN: 4.00 x 2.50 x 36.00 LBS: 0.00 MM: 101.60 x 63.50 x 914.40 KG: 0.00 LOCATION: COUNTERTOP NOTES: 2025-0616 UPDATED ID PER SENECA SS CHANGE CODE: C CHANGE NOTES: MODEL CHANGE PER SHD VOLTS: 120 HERTZ: 60 WATS: 60 ELECT NOTE: 60W BULB	FISHER SCIENTIFIC MAGNIFIER 2X 15078220	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		1007255 CELL01	1	LAB ANALYZER, HEMATOLOGY, AUTOMATED 240V;BLOOD CELL DIFFERENTIAL ANALYZER; CAPTURES DIGITAL IMAGES OF CELLS FROM BLOOD SMEARS; CAN BE USED AS EITHER (WDXH) IN: 11.00 x 15.40 x 14.60 LBS: 24.00 MM: 279.40 x 391.16 x 370.84 KG: 10.89 LOCATION: COUNTERTOP NOTES: 2025.03.14 CHANGED TO EXISTING. THIS HAS BEEN PURCHASED THROUGH SENECA AUXILIARY DOLLARS AND NOT THROUGH PROJECT BUDGET AS PER JUDY. CHANGE CODE: C CHANGE NOTES: NEW TO EXISTING STATUS CHANGE DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: LIS SUPORT; ETHERNET 10/100 MBPS VOLTS: 240 HERTZ: 60 AMPS: 1.2	CELLAVISION, INC. DC-1	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
5		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1023325 DATA01*	1	DATA LOGGER, TEMPERATURE PROBE WIFI DATA LOGGER. CONFIGURABLE ALARM LEVELS. ALARMS AND DATA VIEWABLE ON BROWSER OR MOBILE APP. INCLUDES: WALL BRACKET (WXDXH) IN: 2.75 x 1.41 x 3.22 LBS: 0.00 MM: 69.85 x 35.82 x 81.79 KG: 0.00 LOCATION: SPACE NOTES: 2025-0811 UPDATED TO VFC 311 PER SHD 2025.07.08 ADDED PER SENECA SS CHANGE CODE: A CHANGE NOTES: ADDED TEMPERATURE MONITOR PER SENECA VOLTS: 120 HERTZ: 60	CONTROL CORPORATION OF AMERICA VFC 311		NEW	GROUP: 2 MAJOR MOVABLE
5		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WXDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		103350 RFZ03	1	REFRIGERATOR/FREEZER, UPRIGHT 115V;BIOMEDICAL;VACCINE STORAGE;14.7CF REF; 6.2CF FREEZER;MICROPROCESSOR TEMP CONTROL;LED DISPLAY;ALARMS;4 CASTERS;2 LEVELIN (WXDXH) IN: 35.40 x 28.10 x 75.20 LBS: 357.00 MM: 899.16 x 713.74 x 1910.08 KG: 161.94 LOCATION: FLOOR SPEC COMMENTS: DEPTH WITH DOORS OPEN: 28.10" + 17.7" = 45.8" NOTES: 2025.07.14 CHANGED FROM EXISTING TO NEW CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISITNG TO NEW DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: REMOTE ALARM CONTACTS VOLTS: 115 HERTZ: 60 AMPS: 4.30 WATS: 510 PLUG TYPE: Type B (NEMA 5-15)	FOLLETT PRODUCTS, LLC (MIDDLEBY) MPR-715F-PA/PNAMPR715FPA REFFZR21-GS MECH NOTE: REFRIGERANT: R-513A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		071820 DIF01	1	LAB COUNTER, BLOOD CELL 12 CELL COUNT;HEMATOLOGY;LCD DISPLAY;PERCENT OR WBC VALUES;ALARMS EVERY 100 (WXDXH) IN: 6.00 x 9.00 x 2.25 LBS: 3.00 MM: 152.40 x 228.60 x 57.15 KG: 1.37 LOCATION: COUNTERTOP NOTES: 2025.0522 UPDATED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: EXISTING TO NEW STATUS CHANGE VOLTS: 120 HERTZ: 60	MODULUS DATA SYSTEMS, INC. DIFFCOUNT III 10-312	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		053801 DSP02	1	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART;WALL ENCLOSURE W/CONTAINER INSERT (WXDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL NOTES: 6/6/2023: PLACE NEAR MICROSCOPES	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		1023684 DATA02*	1	BRACKET, ALARM MOUNT, WALL WALL MOUNTING BRACKET IS COMPATIBLE WITH ALL EL-WIFI SENSOR UNITS. THE BRACKET FEATURES THREE MOUNTING HOLES TO ENSURE SECUR (WXDXH) IN: 3.00 x 2.00 x 4.00 LBS: 0.00 MM: 76.20 x 50.80 x 101.60 KG: 0.00 LOCATION: WALL NOTES: 2025.09.18 ADDED WALL MOUNT BRACKET FOR DATA LOGGER SEPARATELY FOR OF/CI. MOUNT ON WALL ABOVE COUNTER CHANGE CODE: A CHANGE NOTES: ADDED WALL MOUNT BRACKET FOR DATA LOGGER	CONTROL CORPORATION OF AMERICA EL-WIFI WALL MOUNT BRACKET	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
7		066907 REF05	1	REFRIGERATOR, LABORATORY 115V;78.8 CF;3 GLASS DOORS;1-8 DEGREES C12 SHELVES;7 DAY CHART RECORDER;AUTOMATIC DEFROST;AUDIBLE VISUAL ALARMS;DUAL WHEEL (WXDXH) IN: 85.00 x 37.20 x 79.20 LBS: 805.00 MM: 2159.00 x 944.88 x 2011.68 KG: 365.15 LOCATION: FLOOR SPEC COMMENTS: ; NOTES: 2024-09 DOES NOT REQUIRE BMS CONNECTION PER SHD ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 16 ELECT NOTE: 5-20P PLUG	REVCO, A THERMO FISHER COMPANY RGL7504A REVCO BTU: 1979	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		088030 BLK01	1	LAB HEATING BLOCK 1 BLOCK CAPACITY;130 DEG C TO 60 DEG C;HI/LOW THERMOSTATS;SEPARATE ON/OFF POWER SWITCH;PILOT LIGHT INDICATE HEAT (WDXH) IN: 7.00 x 8.00 x 4.00 LBS: 5.00 MM: 177.80 x 203.20 x 101.60 KG: 2.27 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: .83 WATS: 100 ELECT NOTE: 120 VOLTS; 50/60 HERTZ; 100 WATTS; 0.83 AMPS; LINE CORD AND PLUG	THERMO FISHER SCIENTIFIC 2050Q	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
7		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
7		098717 CEN06	1	LAB CENTRIFUGE, SERO-FUGE 2400RPM/500XG OR 3400RPM/1049XG;TIME DISPLAY;W/LID;ON-BOARD TACHOMETER;SUPPLIED ROTOR HAS NUMBERED SLOTS FOR SAMPLE (WDXH) IN: 17.00 x 13.00 x 12.00 LBS: 28.00 MM: 431.80 x 330.20 x 304.80 KG: 12.71 LOCATION: COUNTERTOP VOLTS: 115 HERTZ: 60 AMPS: 6.5 ELECT NOTE: CONNECT TO 3-WIRE GROUNDED AC RECEPTACLE	CLAY ADAMS BD#420352/MC#493688 SEROFUGE 2002 MECH NOTE: 175BTU/HR;12" CLEARANCE ALL SIDES	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
8		084252	1	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR NOTES: RED	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1006966 SED01	1	ANALYZER, ERYTHROCYTE SEDIMENTATION RATE (ESR), AUTOMATED ERYTHROCYTE SEDIMENTATION RATE (ESR) ANALYZER; FULLY AUTOMATED; REQUIRES 100 MICROLITERS OF SAMPLE; INCLUDES INTERNAL (WXDXH) IN: 7.10 x 10.40 x 9.50 LBS: 9.80 MM: 180.34 x 264.16 x 241.30 KG: 4.45 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ-45 ETHERNET FOR LIS CONNECTION VOLTS: 120 HERTZ: 60 WATS: 60 ELECT NOTE: 100-240VAC	ALCOR SCIENTIFIC, INC. MINIISED	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
8		1004090	1	WASTE RECEPTACLE 13GAL; OPEN TOP; GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
8		1022425 INC01	1	ANALYZER, BLOOD BANK CENTRIFUGE & INCUBATOR; INTEGRATED ID-MTS GEL CARDS; ALARM SYSTEM; TEMP. MONITOR; 10 CARD CAPACITY CENTRIFUGE; 20 CARD (WXDXH) IN: 22.67 x 8.67 x 12.80 LBS: 24.00 MM: 575.82 x 220.22 x 325.12 KG: 10.89 LOCATION: COUNTERTOP NOTES: 2025-0714 UPDATED BACK TO EXISTING PER UPDATED SENECA REVIEW 2025-0616 UPDATED FROM EXISTING TO NEW PER SENECA SS NEED WORKSPACE FOR PC CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISTING TO NEW VOLTS: 120 HERTZ: 60	ORTHO CLINICAL DIAGNOSTICS, INC. ORTHO WORKSTATION KIT 6904630AS	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
9		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WXDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		1023693 LBM01*	2	LAB MICROSCOPE BINOCULAR MICROSCOPE. SPECIFICALLY PRECONFIGURED FOR HEMATOLOGY/ ONCOLOGY LABS. 4X, 10X, 40X, 50X OIL AND 100X OIL (WXDXH) IN: 8.00 x 10.00 x 16.00 LBS: 14.00 MM: 203.20 x 254.00 x 406.40 KG: 6.36 LOCATION: COUNTERTOP NOTES: 2025-0714 INCREASED QTY FROM 1 TO 2 PER UPDATED SENECA REVIEW, UPDATED MODEL TO LUMEON #597 2025-0602 UPDATED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: INCREASED QTY, CHANGED FROM EXIST TO NEW	MCKESSON 597 LUMEON / 885235	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1007240 BAC01	1	ANALYZER, BLOOD CULTURE BLOOD CULTURE DIAGNOSTICS SYSTEM; HOLDS UP TO 40 (WXDXH) IN: 26.60 x 23.00 x 15.40 LBS: 70.10 MM: 675.64 x 584.20 x 391.16 KG: 31.80 LOCATION: COUNTERTOP NOTES: 2025.02.06 UPDATED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: UPDATED FROM EXISTING TO NEW VOLTS: 120 HERTZ: 60	BECTON DICKINSON 442296 BACTEC FX40	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
9		1013950 INC02	1	LAB ANALYZER ORTHO OPTIX WORKS WITH THE ORTHO WORKSTATION, THE ONLY COMPACT 2-IN-1 BLOOD TESTING SCREENING SYSTEM; 120VAC (WXDXH) IN: 8.46 x 17.13 x 9.45 LBS: 11.90 MM: 214.89 x 435.11 x 240.03 KG: 5.40 LOCATION: COUNTERTOP NOTES: WORKS WITH THE INC01 ORTHO WORKSTATION 2025.07.14 UPDATED BACK TO EXISTING PER UPDATED SENECA REVIEW 2025.06.02 UPDATED FROM EXISTING TO NEW PER SENECA SS DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: NETWORK CONNECTION IS NEEDED FOR LIS AND E- CONNECTIVITY VOLTS: 120 HERTZ: 60 AMPS: 150 VA	ORTHO CLINICAL DIAGNOSTICS, INC. OPTIX	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		053801 DSP02	1	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART;WALL ENCLOSURE W/CONTAINER INSERT (WXDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL NOTES: 6/6/2023: PLACE NEAR MICROSCOPES	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
10		1022429 RCKR01	1	LAB ROCKER, TUBE SINGLE ON/OFF CONTROL SWITCH 15 TUBE MODEL (WXDXH) IN: 14.00 x 5.00 x 5.50 LBS: 0.00 MM: 355.60 x 127.00 x 139.70 KG: 0.00 LOCATION: COUNTERTOP NOTES: 2025-0616 UPDATED FROM EXISTING TO NEW PER SENECS SS CHANGE CODE: C CHANGE NOTES: CHANGED FORM EXISTING TO NEW VOLTS: 120 HERTZ: 60	MCKESSON BLD TUBE 15PLACE 862482	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		093825 LPS01	1	LAB PLASMA SYSTEM THAWS FROZEN BAGS OF PLASMA;UP TO 4 BAGS;600 ML BAG SIZE;14 MIN THAW TIME;37 DEG. CELSIUS TEMP PRESET POINT;DRAIN PORT ON (WXDXH) IN: 9.00 x 21.00 x 14.00 LBS: 0.00 MM: 228.60 x 533.40 x 355.60 KG: 0.00 LOCATION: COUNTERTOP VOLTS: 115 HERTZ: 60 WATS: 770	CYTOTHERM PLASMA THAWER CT-4S	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		100924 IMM01	1	LAB PROCESSOR, BLOOD ACCOMODATES 32 STRIPS; COLOR IMAGING ADD SYSTEM FLUID OR REMOVE WASTE ANYTIME DURING OPERATION;AUTOMATED (WXDXH) IN: 45.00 x 23.00 x 26.00 LBS: 104.00 MM: 1143.00 x 584.20 x 660.40 KG: 47.18 LOCATION: COUNTERTOP <i>NOTES: THIS WILL BE A FUTURE PIECE OF EQUIPMENT PER DAVE MONDAY, AND WILL REPLACE SOME OF THE OTHER EQUIPMENT TO MAKE SPACE FOR IT WHEN ADDED IN THE FUTURE</i> DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: PC TO INSTRUMENT: SERIAL RS232 INSTRUMENT INTERNAL: SERIAL RS232 MONITOR TO PC: USB VOLTS: 120 HERTZ: 60 AMPS: 2 WATS: 250 ELECT NOTE: POWER SUPPLY: 5X11X3	IMMUCOR, INC. - LUMENA ECHO	OF/VI	FUTURE	GROUP: 2 MAJOR MOVABLE
12		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		062537 VWR01	1	LAB VIEWER, TEST TUBE 5 WATT HALOGEN BULB;ON/OFF SWITCH;FLEXIBLE NECK; INCLUDES (WXDXH) IN: 5.00 x 8.00 x 12.00 LBS: 0.00 MM: 127.00 x 203.20 x 304.80 KG: 0.00 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 WATS: 18	THERMO FISHER SCIENTIFIC 22-363-560 AGGLUTINATION	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
13		105022 CEN03	1	LAB CENTRIFUGE, COUNTER TOP PROGRAM UP TO 10 CUSTOM TIME;RPM AND G-FORCE SETTINGS;SET AND LOCK 3-PREFERRED CYCLES;LIGHTED LID INDICATE CYCLE (WXDXH) IN: 13.00 x 15.00 x 9.00 LBS: 30.00 MM: 330.20 x 381.00 x 228.60 KG: 13.61 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60	DRUCKER COMPANY DASH FLEX 12	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		014328 BLK02	1	LAB HEATING BLOCK 2 BLOCK CAPACITY; BIMETALLIC CONTROLLED HEATER TEMPERATURE UNIFORMITY +/- .5 DEGREE C (WDXH) IN: 8.00 x 9.00 x 4.00 LBS: 5.00 MM: 203.20 x 228.60 x 101.60 KG: 2.27 LOCATION: COUNTERTOP NOTES: TYPICAL, NOT ACTUAL VOLTS: 120 HERTZ: 60 AMPS: 1.25 WATS: 150 ELECT NOTE: LINE CORD AND PLUG	THERMO FISHER SCIENTIFIC 2052Q MULTI-BLOK	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
14		041208 CEN01	1	LAB CENTRIFUGE, COUNTER TOP MAXIMUM SPEED: 13,200 RPM; 48 X 0.5ML CAPACITY; MAXIMUM RCF: 16,250XG; MAXIMUM SAMPLE VOLUME: 36ML (WDXH) IN: 12.00 x 13.30 x 19.80 LBS: 36.00 MM: 304.80 x 337.82 x 502.92 KG: 16.33 LOCATION: COUNTERTOP NOTES: LOCATED WITH THE OTHER CENTRIFUGE UNITS, ACROSS FROM THE AUTO CHEM ANALYZER VOLTS: 120 HERTZ: 60 AMPS: 6.3	INTERNATIONAL EQUIPMENT CO. (IEC) 2498 MICROMAX	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
14		053801 DSP02	1	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART; WALL ENCLOSURE W/CONTAINER INSERT (WDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL NOTES: 6/6/2023: PLACE NEAR MICROSCOPES	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	2	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
16		1004090	1	WASTE RECEPTACLE 13GAL; OPEN TOP; GRAY (WDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY
ROOM NUMBER: 503

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY (PHLEBOTOMY DESK)
ROOM NUMBER: 503D

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1019034 RFZ01	1	REFRIGERATOR/FREEZER, UNDERCOUNTER 3.2CF CAP;REVERSIBLE DOOR SWING (RH DEFAULT);ADJ SHELF;2 DOOR SHELVES;AVAILABLE IN WHITE (W) OR BLACK/STAINLESS STEEL (WXDXH) IN: 19.00 x 19.75 x 32.75 LBS: 51.00 MM: 482.60 x 501.65 x 831.85 KG: 23.14 LOCATION: FLOOR NOTES: ER 07-24-24: REPLACED DISCONTINUED ID 098003 WITH 1019034. ANCHOR DETAIL REQ: NO CHANGE CODE: C CHANGE NOTES: UPDATED MODEL, UPDATED DIMENSIONS VOLTS: 115 HERTZ: 60 AMPS: 1.1 ELECT NOTE: 5' POWER CORD	SUMMIT APPLIANCE DIV. CP34[COLOR]ADA CP34	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY (PROCESSING)
ROOM NUMBER: 503P

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		105022 CEN03	1	LAB CENTRIFUGE, COUNTER TOP PROGRAM UP TO 10 CUSTOM TIME;RPM AND G-FORCE SETTINGS;SET AND LOCK 3-PREFERRED CYCLES;LIGHTED LID INDICATE CYCLE (WDXH) IN: 13.00 x 15.00 x 9.00 LBS: 30.00 MM: 330.20 x 381.00 x 228.60 KG: 13.61 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60	DRUCKER COMPANY DASH FLEX 12	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		1007239	1	STAMP, TIME AND DATE REGISTERS TIME, DATE, NUMBERS, AND CUSTOMIZED MESSAGES (WDXH) IN: 6.20 x 7.10 x 7.10 LBS: 4.00 MM: 157.48 x 180.34 x 180.34 KG: 1.82 LOCATION: COUNTERTOP DATA NOTE: USB 2.0 VOLTS: 120 HERTZ: 60 AMPS: .6	LATHEM TIME RECORDER COMPANY LT5000	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
3		053801 DSP02	1	DISPOSAL CONTAINER, SHARPS W/ENCLOSURE 12 QUART;WALL ENCLOSURE W/CONTAINER INSERT (WDXH) IN: 15.00 x 7.00 x 18.00 LBS: 6.00 MM: 381.00 x 177.80 x 457.20 KG: 2.73 LOCATION: WALL NOTES: 6/6/2023: PLACE NEAR MICROSCOPES	KENDALL HEALTHCARE, COVIDIEN 85301H/8536SA	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
4		1014551	1	WASTE CONTAINER, BIOHAZARDOUS LIGHTWEIGHT ABS BENCHTOP WASTE DISPOSAL WITH FLIP-TOP LID (WDXH) IN: 5.00 x 3.88 x 7.63 LBS: 2.00 MM: 127.00 x 98.56 x 193.81 KG: 0.91 LOCATION: COUNTERTOP	MARKETLAB, INC. ML16505 1.4 Q BENCHTOP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
6		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: LABORATORY (PROCESSING)
ROOM NUMBER: 503P

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WDXDH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
8	No image available	101177 ESR01	1	LAB ANALYZER, ELECTROLYTE HAND-HELD TESTING DEVICE UTILIZING HOVA LACTATE TEST STRIPS (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.02.06 UPDATED FROM EXISTING TO NEW PER SENECA SS CHANGE CODE: C CHANGE NOTES: CHANGED FROM EXISTING TO NEW	NOVA BIOMEDICAL - LACTATE PLUS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: OFFICE / BREAKROOM
ROOM NUMBER: 504

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		008082 REF02		1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		090202 COF03*		1	COFFEE BREWER SINGLE-CUP BREWER; BREW SIZES: 6/8/10 OZ; 48OZ WATER RESERVOIR; AUTO ON/OFF; REMOVABLE DRIP TRAY; COMMERCIAL CERTIFIED (WXDXH) IN: 9.50 x 12.25 x 13.25 LBS: 14.00 MM: 241.30 x 311.15 x 336.55 KG: 6.36 LOCATION: COUNTERTOP NOTES: 2025-0714 UPDATED TO NON PLUMBED KUREIG PER UPDATED SENECA REVIEW CHANGE CODE: C CHANGE NOTES: CHANGED MODEL VOLTS: 120 HERTZ: 60 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. COMM BREWING K145 853653/23145NS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		055770 OVM01		1	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WXDXH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
4		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: OFFICE / BREAKROOM
ROOM NUMBER: 504

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1005146		1	WASTE RECEPTACLE, RECYCLE 7GAL RECTANGULAR RECYCLING WASTEBASKET; PLASTIC; BLUE; PACK OF 12 (WDXH) IN: 10.50 x 14.50 x 15.00 LBS: 1.88 MM: 266.70 x 368.30 x 381.00 KG: 0.86 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG295673BLUE BRUTE RECYCLING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: LAB
ROOM NAME: SUPPLY STORAGE
ROOM NUMBER: 505

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	3	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: STORAGE
ROOM NUMBER: Z400

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015148 FRK01	1	TRUCK, FORKLIFT FULLY POWERED STRADDLE STACKER LIFT TRUCK; 65" LIFT' 2650 LB CAPACITY; 120V AC (WDXH) IN: 59.50 x 62.00 x 65.00 LBS: 1400.00 MM: 1511.30 x 1574.80 x 1651.00 KG: 635.03 LOCATION: FLOOR SPEC COMMENTS: ; VOLTS: 120 HERTZ: 60 ELECT NOTE: UNIT INCLUDES (2) 12V / 85AH MAINTENANCE-FREE BATTERIES, BATTERY DISCHARGE GAUGE AND AN INTEGRAL 15 AMP 120V CHARGER.	GLOBAL INDUSTRIAL WB988955	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		053947 CRI01	2	CART, ISOLATION 4 DRAWER;YELLOW;BASIC CART W/DRAWER; CONFIGURATION OF 1 EA.3" (WDXH) IN: 33.50 x 25.00 x 46.25 LBS: 165.00 MM: 850.90 x 635.00 x 1174.75 KG: 74.85 LOCATION: FLOOR	ARMSTRONG MEDICAL INDUSTRIES AKL-Y-4 TWO TONE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
3		092852	1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1014831	2	SHELVING, STEEL 5 SHELF; STEEL DECK; 3000 LBS CAPACITY; ADJUSTABLE SHELVING (WDXH) IN: 72.00 x 24.00 x 72.00 LBS: 338.00 MM: 1828.80 x 609.60 x 1828.80 KG: 153.32 LOCATION: FLOOR	GLOBAL INDUSTRIAL 236618 EXTRA HEAVY DUTY	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
5		078823 SHW02	4	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: STORAGE
ROOM NUMBER: Z400

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		037382	1	RACK, CRUTCH CRUTCH/CANE RACK;30 PAIR OF CRUTCHES;10 CANES MOUNTING HARDWARE NOT SUPPLIED (WDXH) IN: 60.00 x 16.00 x 17.00 LBS: 24.00 MM: 1524.00 x 406.40 x 431.80 KG: 10.89 LOCATION: WALL	BAILEY MFG. COMPANY 790 790	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
7		056756	2	LOUVER PANEL, WALL MOUNTED 1000LB CAPACITY;GRAY;16 GAUGE STEEL;POWDER COAT FINISH;ATTACHES DIRECTLY TO WALL;MOUNTING SCREWS (WDXH) IN: 36.00 x 1.00 x 61.00 LBS: 40.00 MM: 914.40 x 25.40 x 1549.40 KG: 18.15 LOCATION: WALL	AKRO-MILS 30161 -	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
8		000156	2	LOUVER PANEL, WALL MOUNTED STORAGE;MODULAR TYPE;LOUVERED FOR BINS;WALL MOUNT FOR VERTICAL AND HORIZONTAL SUSPENSION;HEAVY DUTY STEEL W/HORIZONTAL RIB (WDXH) IN: 35.75 x 0.31 x 19.00 LBS: 11.25 MM: 908.05 x 7.88 x 482.60 KG: 5.11 LOCATION: WALL NOTES: PACK OF 4 EACH	AKRO-MILS 30636 30636	CF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: MAINTENANCE
ROOM NUMBER: Z401

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015154	1	BENCH, TECH WORKSTATION, 6 FT. COMPOSITE WOOD TOP (MODEL H-1137-WOOD); ACTUAL ADJUSTABLE HEIGHT (INCLUDING TOP): 28 1/8 - 38 1/4"; STEEL HOUSING: 69 X 25 (WXDXH) IN: 30.00 x 72.00 x 38.00 LBS: 329.00 MM: 762.00 x 1828.80 x 965.20 KG: 149.24 LOCATION: FLOOR NOTES: MUST INCLUDE COMPOSITE WOOD TOP	ULINE H-9636-WOOD	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015155	1	BENCH, TECH WORKSTATION, 6 FT. INCLUDES TABLE, FULL DEPTH BOTTOM SHELF, DRAWER AND UPRIGHTS.TABLE HEIGHT ADJUSTS FROM 30-37". 500 LB. CAPACITY. (WXDXH) IN: 30.00 x 72.00 x 37.00 LBS: 231.00 MM: 762.00 x 1828.80 x 939.80 KG: 104.78 LOCATION: FLOOR	GLOBAL INDUSTRIAL H-6342-WOOD	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1015158	3	ELECTRICAL OUTLET STRIP 15 AMP/120V. 15 FOOT CORD. DIMENSIONS: 19.8" X 1.9" X 1.7" (WXDXH) IN: 2.00 x 19.00 x 2.00 LBS: 2.00 MM: 50.80 x 482.60 x 50.80 KG: 0.91 LOCATION: SPACE VOLTS: 120 HERTZ: 60	GLOBAL INDUSTRIAL WB500883	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
4		1014539	1	LIGHT, MAGNIFIER 36" REACH;5" DIAMETER LENS;ROUND MAGNIFIER LIGHT;5-DIOPTER (2.25X);600 WHITE LEDS;SPRING BALANCED ARM;HEAVY DUTY MOUNTING (WXDXH) IN: 0.00 x 36.00 x 0.00 LBS: 6.70 MM: 0.00 x 914.40 x 0.00 KG: 3.04 LOCATION: SPACE VOLTS: 120 HERTZ: 60 AMPS: 0.3 ELECT NOTE: 5 FT POWER CORD	AVEN TOOLS 26501-LED/26505-CL-HD PROVUE SUPERSLIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		1015157	1	WORK PLATFORM, MOBILE 72"W X 36"D MOBILE PRODUCTION WORKBENCH - POWER APRON - SHOP TOP SQUARE EDGE BLUE; PRE-WIRED POWER APRON INCLUDES ONE USB (WXDXH) IN: 72.00 x 36.00 x 36.00 LBS: 0.00 MM: 1828.80 x 914.40 x 914.40 KG: 0.00 LOCATION: FLOOR VOLTS: 115 HERTZ: 60	GLOBAL INDUSTRIAL WB254000WBBL	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: MAINTENANCE
ROOM NUMBER: Z401

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		1022463 CFS01*	1	CABINET, SAFETY, FLAMMABLE STORAGE 45 GAL CAP;2 DOORS SELF CLOSE;2 SHELF;YELLOW;CONFORMS TO OSHA/NFPA STANDARDS;WARNING LABEL;3 POINT LATCH SYSTEM (WXDXH) IN: 43.00 x 18.00 x 65.00 LBS: 313.00 MM: 1092.20 x 457.20 x 1651.00 KG: 141.98 LOCATION: FLOOR SPEC COMMENTS: UNIT CAN BE CONNECTED TO BUILDING EXHAUST IF REQUIRED, BUT NOT RECOMMENDED BY THE VENDOR NOTES: PER SHD: THIS EQUIPMENT WILL BE INSTALLED IN THE SUPPORT SERVICES BUILDING. WILL BE VENT TO THE OUTSIDE AND WILL BE OF/CI WHEN THE SPECIFICATIONS FOR THE SUPPORT SERVICES BUILDING ARE COMPLETED. HGA TO INCORPORATE OF/CI INFO TO SUPPORT SERVICES BLDG BID DOCS CHANGE CODE: C CHANGE NOTES: CHANGED TO VENTED AND OF/CI	JUSTRITE MANUFACTURING COMPANY 894520	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
					VENT REQUIRED: YES MECH NOTE: DUAL 2" NPT VENTS MAY BE CONNECTED TO EXHAUST SYSTEM. MAXIMUM DUCT LENGTH OF 25"			
7		1015151	3	PEGBOARD, WALL MOUNTED CONTENTS INCLUDE (2) 18 IN. X 36 IN. LOCBOARDS; 28 HOOKS: (1) MULTI-PRONG TOOL HOLDER; (1) MULTI-RING TOOL HOLDER; (2) 4	GLOBAL INDUSTRIAL LB18-BKIT WGB2143599	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
				(WXDXH) IN: 36.00 x 1.00 x 72.00 LBS: 22.00 MM: 914.40 x 25.40 x 1828.80 KG: 9.98 LOCATION: WALL				
8		1004790	1	BIN, ACCESSORY, PANEL, LOUVERED MOBILE; DOUBLE SIDED 100LB WEIGHT CAPACITY; TO INCLUDE:	AKRO-MILS 30676	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
				(WXDXH) IN: 36.19 x 25.00 x 66.38 LBS: 110.00 MM: 919.23 x 635.00 x 1686.06 KG: 49.90 LOCATION: FLOOR				

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: MAINTENANCE
ROOM NUMBER: Z401

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		506580	1	ALLOWANCE, BIN (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1015152	1	BIN, SUPPLY BIN DIVIDERS PROVIDE SAFE AND ORGANIZED COMPONENT STORAGE. SOLD IN PACKAGE OF 24, PRICE PER PACKAGE. (WDXDH) IN: 8.00 x 0.00 x 3.00 LBS: 3.00 MM: 203.20 x 0.00 x 76.20 KG: 1.37 LOCATION: SPACE NOTES: PACKAGE OF 24	GLOBAL INDUSTRIAL 40150 WB752372	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		1015153	2	BIN, SUPPLY PLASTIC STACKABLE BINS - 11 X 5 1/2 X 5", BLUE, CASE OF 12 (WDXDH) IN: 5.50 x 11.00 x 5.00 LBS: 0.61 MM: 139.70 x 279.40 x 127.00 KG: 0.28 LOCATION: SPACE NOTES: CASE OF 12 EACH	ULINE S-12415BLU	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: STAFF TOILET
ROOM NUMBER: Z402

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: PURCHASING MAINTENANCE OFFICE
ROOM NUMBER: Z403

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		094527	2	CABINET, STORAGE METAL STORAGE CABINET;BLACK;KEYLESS LOCK;72" OVERALL HEIGHT;ASSEMBLED;4 (WXDXH) IN: 36.00 x 18.00 x 72.00 LBS: 149.00 MM: 914.40 x 457.20 x 1828.80 KG: 67.59 LOCATION: FLOOR	GRAINGER, INC. 22ND35 EA4E361872-09	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: HOUSEKEEPING OFFICE
ROOM NUMBER: Z404

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: MAINTENANCE BUILDING
ROOM NAME: MAINTENANCE OFFICE
ROOM NUMBER: Z405

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		1023669 KEY01*	1	CABINET, KEY 100 SINGLE KEY CAP; KEY LOCK; INCLUDES 2 KEYS, KEY TAGS, NUMBERED KEY HOOKS, MOUNTING HARDWARE, INDEX SHEET (WXDXH) IN: 14.75 x 3.00 x 21.75 LBS: 11.65 MM: 374.65 x 76.20 x 552.45 KG: 5.29 LOCATION: WALL NOTES: 2025.09.08 ADDED PER SHD FINAL REVIEW DOC. SHD TO CONFIRM IF PREFERENCE IS FOR OF/OI OR OF/CI CHANGE CODE: A CHANGE NOTES: ADDED PER SHD	BARSKA CB12964	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: RT WORK / STORAGE
ROOM NUMBER: 604

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		065434 SHW03	3	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WXDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1003148 VNT02	1	VENTILATOR ICU;TRANSPORT;ADULT/PEDS;8.4"COLOR TOUCHSCREEN 41 MONITORING PARAMETERS;5.5 HOURS BACKUP BATTERY, TROLLEY WITH (WXDXH) IN: 24.80 x 24.80 x 64.30 LBS: 40.80 MM: 629.92 x 629.92 x 1633.22 KG: 18.51 LOCATION: FLOOR DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: COMMUNICATION INTERFACE: COM1 PORT FOR CONNECTION TO A REMOTE MONITOR NURSE CALL INTERFACE RJ-45 ETHERNET CONNECTOR USB CONNECTOR REFERENCE VENDOR'S TECHNICAL DATA VOLTS: 120 HERTZ: 60 WATS: 150 ELECT NOTE: POWER CONSUMPTION:50-150W; LITHIUM-ION BATTERY SUPPLIED BY HAMILTON MEDICAL ONLY	HAMILTON MEDICAL, INC. T1 T1	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
3		1000454 VNT01	2	VENTILATOR V60 TRSPT DOM, CART, ARM, O2 MAN, NIC UNAS;PEDIATRIC- ADULT;BIPAP/NON-INVASIVE;CPAP;S/T;PCV;TOUCH-SCREEN;6 HOUR LI-ION (WXDXH) IN: 16.00 x 17.00 x 46.00 LBS: 80.00 MM: 406.40 x 431.80 x 1168.40 KG: 36.29 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 WATS: 300 ELECT NOTE: LITHIUM ION BATTERY (6-HR OPERATION)	PHILIPS MEDICAL SYSTEMS V60 RESPIRONICS V60	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: RT WORK / STORAGE
ROOM NUMBER: 604

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		076164 ABG01	1	LAB ANALYZER, BLOOD GAS 60 SEC RESULTS;W/CART;TOUCH SCREEN/BARCODE READER CLOT DETECTION & MANAGEMENT;CAN USE WHOLE BLOOD (WDXH) IN: 11.50 x 16.00 x 21.50 LBS: 36.50 MM: 292.10 x 406.40 x 546.10 KG: 16.56 LOCATION: FLOOR DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RS232 PORT; 10BASE T ETHERNET VOLTS: 120 HERTZ: 60	SIEMENS MEDICAL SOLUTIONS USA, INC. 10492730/10698368 RP500	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
6		1000341	1	STOOL STOOL GAS CYLINDER; 5-CASTER; (WDXH) IN: 12.00 x 12.00 x 24.00 LBS: 0.00 MM: 304.80 x 304.80 x 609.60 KG: 0.00 LOCATION: FLOOR	STORAGE SYSTEMS UNLIMITED - P-526-GS	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
7		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: RT WORK / STORAGE
ROOM NUMBER: 604

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
9		1006395 HUM02	1	HUMIDIFIER HEAT/HUMIDIFICATION;HIGH FLOW;WITH STAND;DELIVERY OF INSPIRED GAS 1-40LPM;BUILT-IN BLENDER O2 SENSOR; BATTERY BACKUP;USES (WDXH) IN: 25.00 x 25.00 x 72.00 LBS: 40.00 MM: 635.00 x 635.00 x 1828.80 KG: 18.15 LOCATION: FLOOR DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: ELECTRONIC MEDICAL RECORDS (EMR): RS-232 BI-DIRECTIONAL COMMUNICATION NURSE CALL: 3-PIN 3.5 MM STEREO AUDIO JACK VOLTS: 120 HERTZ: 60 WATS: 200 ELECT NOTE: 200VA IN WARM-UP; 80VA STEADY-STATE BATTERY BACKUP POWER: QTY 4, 4.8V NICKEL-METAL HYDRIDE AA BATTERIES (NOT USER REPLACEABLE)	VAPOTHERM PF-UNIT-US PRECISION FLOW	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE MEDICAL GAS: YES PLUMB NOTE: MEDICAL AIR AND O2 AT BETWEEN 4 AND 85 PSI INLET; DISS FITTINGS
10		094333 HUM01	1	HUMIDIFIER HEATS/HUMIDIFIES PATIENT RESP GASES; USER CAN SET OXYGEN PERCENTAGE, TEMP, & FLOW RATE OF DELIVERED GAS; APPROX 400ML (WDXH) IN: 25.00 x 25.00 x 84.00 LBS: 50.00 MM: 635.00 x 635.00 x 2133.60 KG: 22.68 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 ELECT NOTE: 200VA DURING WARM-UP;80VA STEADY STATE BACK UP BATTERY 4.8V NICKEL-METAL HYDRIDE	VAPOTHERM PF-PLUS-US PRECISION FLOW+	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: THERAPY
ROOM NUMBER: 801

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023661 PBR02*	1	PARALLEL BARS 600# CAP.; FOLDABLE; (4) 4" LOCKING CASTERS; ANTI-SLIP PLATFORM; TRANSPORT MODE DIMENSIONS: (WXDXH) IN: 45.50 x 60.00 x 41.00 LBS: 0.00 MM: 1155.70 x 1524.00 x 1041.40 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.16 SHD CONFIRMED ARMEDICA PORTABLE PARALLEL BARS ANCHOR DETAIL REQ: NO	ARMEDICA MANUFACTURING CORPORATION PTPB1000	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		103132 STP01	1	EXERCISE UNIT, STAIRCASE TRAINING STAIRS WITH BUS STEP;(3) 4" RISERS ON ONE SIDE AND (2)6" RISERS (WXDXH) IN: 31.00 x 55.00 x 48.00 LBS: 61.00 MM: 787.40 x 1397.00 x 1219.20 KG: 27.67 LOCATION: FLOOR	PERFORMANCE HEALTHCARE 561363 081535582	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1007219 EXR03	1	EXERCISE UNIT, CROSS TRAINER, ELLIPTICAL RECUMBENT LINEAR CROSS TRAINER; SWIVEL SEAT; ADJUSTABLE RESISTANCE; 18 SET PROGRAMS; 9" LCD DISPLAY; 450 LB USER WEIGHT (WXDXH) IN: 30.00 x 60.00 x 40.00 LBS: 210.00 MM: 762.00 x 1524.00 x 1016.00 KG: 95.26 LOCATION: FLOOR NOTES: 2025.05.12 UPDATED FROM NEW TO EXISTING MODEL PHYSIO STEP LXT PER SHD SPREADSHEET ANCHOR DETAIL REQ: NO CHANGE CODE: C CHANGE NOTES: UPDATED TO EXISITNG MODEL VOLTS: 110 HERTZ: 60	HEALTHCARE INTERNATIONAL PHYSIOSTEP LXT LXT-700	OF/VI	EXISTING	GROUP: 3 MINOR MOVABLE
4		065262 EXR02	1	EXERCISE UNIT, ERGOMETER, ARM UPPER/LOWER BODY EXERCISER;MAGNETIC RESISTANCE DIGITAL DISPLAY;ADJUSTABLE BASE (WXDXH) IN: 17.00 x 9.00 x 17.00 LBS: 18.00 MM: 431.80 x 228.60 x 431.80 KG: 8.17 LOCATION: COUNTERTOP	PERFORMANCE HEALTHCARE 554035 MAGNETIC RESIST	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: THERAPY
ROOM NUMBER: 801

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		087084 TBR01	1	TABLE, WORK, ADJUSTABLE HAND CRANK PATIENT TABLE WITH TILT TOP;HEIGHT ADJ RANGE 27"-39";1-1/8" ALL LAMINATE TOP;ALUMINUM AND POWDER-COATED STAND (WDXH) IN: 36.00 x 32.00 x 39.00 LBS: 0.00 MM: 914.40 x 812.80 x 990.60 KG: 0.00 LOCATION: FLOOR	CLINTON INDUSTRIES 76-36C	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		092792 MAT01	1	TABLE, MAT, MOTORIZED WT CAP 900LBS;20-30" HT RANGE;HAND CONTROL SWITCH;2" FOAM PADDING;PERMABLOK COATING (WDXH) IN: 48.00 x 84.00 x 30.00 LBS: 270.00 MM: 1219.20 x 2133.60 x 762.00 KG: 122.47 LOCATION: FLOOR ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 2.4 ELECT NOTE: HOSPITAL GRADE PLUG	PERFORMANCE HEALTHCARE A370562 PERFORMA POWER	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
7		092980 FRM01	1	FRAME, STANDING ELECTRIC SIT-TO-STAND STANDING FRAME; FOUR POINT SYSTEM STABILIZATION AT FEET, KNEES, BUTTOCKS AND (WDXH) IN: 42.00 x 38.00 x 53.00 LBS: 295.00 MM: 1066.80 x 965.20 x 1346.20 KG: 133.81 LOCATION: FLOOR VOLTS: 115 HERTZ: 60 AMPS: 3.5	PERFORMANCE HEALTHCARE 961522B METRON ELITE	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
8		102949 PUL01	1	EXERCISE UNIT, PULLEY, WALL 2-42" ALUMINUM WALL TRACKS W/CONNECTIONS;3 ANCHORS INSTRUCTION POSTER;7 PROGRESSIVE RESISTANCE BANDS (WDXH) IN: 6.00 x 2.00 x 84.00 LBS: 15.00 MM: 152.40 x 50.80 x 2133.60 KG: 6.81 LOCATION: WALL	PERFORMANCE HEALTHCARE THERA-BAND CLX WALL 087106779	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
9		078097 RCK07	3	EXERCISE UNIT, WEIGHTS ACCESSORY, RACK, STORAGE STRIP;WALL MOUNTED HOLDS 11 DUMBBELLS;GRAY PRE-DRILLED MOUNTING (WDXH) IN: 3.00 x 2.00 x 55.00 LBS: 0.00 MM: 76.20 x 50.80 x 1397.00 KG: 0.00 LOCATION: WALL ANCHOR DETAIL REQ: YES	CLINTON INDUSTRIES 5113K -	OF/CI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: THERAPY
ROOM NUMBER: 801

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		094356	1	EXERCISE UNIT, DUMBBELLS SET OF 10 NEOPRENE COATED CAST IRON DUMBBELLS; 1 EA, 1-10 LBS; WEIGHT IS TH (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 55.00 MM: 0.00 x 0.00 x 0.00 KG: 24.95 LOCATION: SPACE	PERFORMANCE HEALTHCARE 081513878 INDIVIDUAL NEOPRENE DUMBBELLS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		092574	1	EXERCISE UNIT, WEIGHT BAR SET OF 5;1,2,3,4,5LB WEIGHTS;GRAY BAR W/CLR LABEL (WDXH) IN: 30.00 x 1.50 x 1.50 LBS: 15.00 MM: 762.00 x 38.10 x 38.10 KG: 6.81 LOCATION: SPACE	PERFORMANCE HEALTHCARE 540014 081035013	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		1005075	1	EXERCISE UNIT, WEIGHT, WRIST AND ANKLE 18 PIECES;ANKLE/WRIST;2 EA. 1LB, 1.5LB, 2LB. (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 73.00 MM: 0.00 x 0.00 x 0.00 KG: 33.12 LOCATION: SPACE	PERFORMANCE HEALTH 5170Y SET Y	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
13		106653 STL02	2	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: THERAPY
ROOM NUMBER: 801

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
16		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: PT/OT/RT
ROOM NAME: THERAPY STORAGE
ROOM NUMBER: 802

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		071749 RCK03	1	RACK, CRUTCH FOR CRUTCHES,CANES,WALKERS & ROLLS;ALL STEEL GREY POWDER COAT (WXDXH) IN: 33.00 x 12.00 x 4.00 LBS: 13.00 MM: 838.20 x 304.80 x 101.60 KG: 5.90 LOCATION: WALL	IDEAL RR31	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: LOBBY
ROOM NUMBER: 100

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015398		2	WASTE RECEPTACLE STEEL; INDOOR USE; INCLUDES ONE 24 GALLON SQUARE CONTAINER AND TWO 8 GALLON HALF-ROUND CONTAINERS; PREMIUM GLOSS POWDER (WXDXH) IN: 15.75 x 31.50 x 30.00 LBS: 81.50 MM: 400.05 x 800.10 x 762.00 KG: 36.97 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC	TRASHCAN WAREHOUSE RC-KDSQ-3C-BT THREE-STREAM COMBO	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015399		1	DISPENSER, PPE STATION KIT INCLUDES: - RS004-0212 RESPIRATORY HYGIENE STATION (WXDXH) IN: 22.25 x 17.00 x 58.06 LBS: 18.10 MM: 565.15 x 431.80 x 1474.73 KG: 8.22 LOCATION: FLOOR	BOWMAN MANUFACTURING COMPANY BD114-0012 COMPLIANCE KIT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: CHAPEL
ROOM NUMBER: 102

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: PUBLIC TOILET (QTY 2)
ROOM NUMBER: 104
ROOM COMMENT: RM 104, 105 NEAR MALL

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: DECON
ROOM NUMBER: 400

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		088075 STR04	1	STRETCHER, SHOWER PVC;WATER PROOF FOAM PAD;DROP SIDE (WXDXH) IN: 28.00 x 79.00 x 34.00 LBS: 130.00 MM: 711.20 x 2006.60 x 863.60 KG: 58.97 LOCATION: FLOOR	LUMEX, INC. GRF-8005	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3	No image available	065637	1	ORGANIZER, PPE HOLDS 3 SINGLE GLOVE BOXES;1 LG & 2 SM POCKETS;2 LG POCKETS ON BACK OF CADDY W/SIDE ACCESS;TWO HOOKS FOR DOOR HANGING;CAN (WXDXH) IN: 33.00 x 5.00 x 19.00 LBS: 0.00 MM: 838.20 x 127.00 x 482.60 KG: 0.00 LOCATION: SPACE	ROYAL ARMS MEDICAL INC. DOOR HANGING Q9	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: LACTATION
ROOM NUMBER: 401

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: LOCKERS, MALE / FEMALE (QTY 2)
ROOM NUMBER: 402
ROOM COMMENT: RM 402, 404

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		089875	2	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: LOCKERS, TOILET (QTY 2)
ROOM NUMBER: 403
ROOM COMMENT: RM 403, 405

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: STAFF BREAK
ROOM NUMBER: 406

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		091485 ICE01	1	ICE MAKER W/WATER DISPENSER 12LB STORAGE;260LB PROD/24 HRS;SS EXTERIOR;EXTERNAL AIR FILTER;AGION ANTIMICROBIAL;R-134A REFRIGERANT;H2 NUGGET STYLE ICE (WXDXH) IN: 17.00 x 25.00 x 35.00 LBS: 183.00 MM: 431.80 x 635.00 x 889.00 KG: 83.01 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 7 ELECT NOTE: SINGLE PHASE;15A MAX FUSE SEE VENDOR TECHNICAL INFORMATION	SCOTSMAN ICE SYSTEMS HID312A-1 MERIDIAN AIR COOL	OF/CI	NEW	GROUP: 1 FIXED
2		1021297	1	FILTER VALVE-IN-HEAD WATER FILTER SYSTEM. INSTALLS ON THE 3/4-INCH MAIN WATER LINE AT THE POINT OF ENTRY. FILTER HOUSING DESIGNED (WXDXH) IN: 5.00 x 5.00 x 13.25 LBS: 0.50 MM: 127.00 x 127.00 x 336.55 KG: 0.23 LOCATION: WALL CHANGE CODE: A CHANGE NOTES: ADDED PER SENECA, HGA HAS ALREADY ACCOUNTED FOR FI	CULLIGAN WATER SYSTEMS HF-360	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		1002050 RFZ02	2	REFRIGERATOR/FREEZER, UPRIGHT 18.3CF;TOP MOUNT FREEZER; WHITE; NO ICE (WXDXH) IN: 30.00 x 30.38 x 66.38 LBS: 164.00 MM: 762.00 x 771.66 x 1686.06 KG: 74.39 LOCATION: FLOOR SPEC COMMENTS: CLEARANCES: 3/8" SIDES AND TOP, 1" BACK VOLTS: 120 HERTZ: 60 AMPS: 15 WATS: 1800 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A CIRCUIT;3 PRONG GROUNDING PLUG	FRIGIDAIRE FFHT1835VW	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: STAFF BREAK
ROOM NUMBER: 406

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		055770 OVM01	1	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WXDXH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP NOTES: BUILT IN BELOW CABINET VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		083455 COF01	1	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WXDXH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED
6		1006012 COF04	1	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WXDXH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: STAFF BREAK
ROOM NUMBER: 406

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		092852		1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: EMERGENCY DRY FOOD STORAGE
ROOM NUMBER: 409

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		065434 SHW03	3	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WXDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1020132 DRCK01*	5	RACK, DUNNAGE ALUMINUM DUNNAGE RACK; 2,000 LBS CAPACITY; ALL-WELDED BULK STORAGE; RUST AND CORROSION RESISTANT; NSF CERTIFIED. (WXDXH) IN: 48.00 x 24.00 x 12.00 LBS: 12.00 MM: 1219.20 x 609.60 x 304.80 KG: 5.45 LOCATION: FLOOR NOTES: 2024.08.30 REMOVED PALLET SHELVING, CHANGE TO DUNNAGE RACKS MIN 9" OFF THE FLOOR AND 24" DEEP, 4FT INCREMENTS AND FILL THE WALL AS PER HGA/SENECA. ANCHORING WILL NOT BE REQUIRED WITH THIS SET UP. ANCHOR DETAIL REQ: NO CHANGE CODE: C CHANGE NOTES: CHANGE TO DUNNAGE RACKS	ULINE H-7497	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: GENERAL STORAGE
ROOM NUMBER: 410

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		065434 SHW03	6	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: SOILED LINEN
ROOM NUMBER: 411

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100858 TRK01	6	TRUCK/BIN, LINEN 500LB; 16 CUFT CAPACITY; MANUAL UNLOADING; BLUE (WDXH) IN: 30.38 x 44.00 x 35.50 LBS: 60.50 MM: 771.66 x 1117.60 x 901.70 KG: 27.45 LOCATION: FLOOR NOTES: INCLUDES LID; COLOR BLUE FOR ALL	TOTER INC. MMC16-90701 CUBE HEAVY DUTY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: CLEAN LINEN STORAGE
ROOM NUMBER: 412

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		065434 SHW03	3	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WXDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1015397 CRL01	2	CART, LINEN 5" QUIET, NON-MARKING TENTE CASTERS AMONG THE BEST ON THE MARKET; REINFORCED INJECTION-MOLDED SHELVES WITH ½" LIP TO (WXDXH) IN: 40.00 x 20.00 x 63.00 LBS: 50.00 MM: 1016.00 x 508.00 x 1600.20 KG: 22.68 LOCATION: FLOOR	DIRECT SUPPLY, INC LC36-4-42 IPU PREMIUM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 413

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		052851 HKP01	1	CART, HOUSEKEEPING SS;2 SHELF;FOLDING PLATFORM W/LOCKING DOOR;ZIP BAG;TRASH LID;TOOL BRACKET W/4 HANDLES & 3 HOOKS;TRAY TOP (WDXDH) IN: 23.00 x 36.00 x 42.50 LBS: 32.00 MM: 584.20 x 914.40 x 1079.50 KG: 14.52 LOCATION: FLOOR	ROYCE ROLLS RINGER CO. F36 FOLDING F36 FOLDING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXDH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXDH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
4		100388 SCB02	1	FLOOR MACHINE, SCRUBBER, POLISHER WALK BEHIND;SCRUBBING PATH OF 14.5";BATTERY OPERATED (WDXDH) IN: 18.70 x 48.00 x 42.70 LBS: 176.00 MM: 474.98 x 1219.20 x 1084.58 KG: 79.84 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 ELECT NOTE: 1-84AH AGM BATTERY (12V) ELECTRICAL SHOWN FOR CHARGING	NILFISK ADVANCE 183603 SC351	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		093627 VAC04	1	VACUUM, WET/DRY 19GAL TANK;FRONT MOUNT SQUEEGEE;ERGO TIP-&-POUR DESIGN;9GAL DUST BAG (WDXDH) IN: 24.00 x 26.00 x 41.00 LBS: 61.00 MM: 609.60 x 660.40 x 1041.40 KG: 27.67 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 WATS: 1100 ELECT NOTE: 50' POWER CORD	NILFISK ADVANCE 56383408 VL500-75	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 413

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		1003893 VAC01	1	CLEANER, CARPET / FLOOR, VACUUM, PORTABLE UPRIGHT VACUUM W/ 18-INCH CLEANING PATH;AUTOMATIC BRUSH-HEIGHT ADJUSTMENT; 40 FOOT POWER CORD (WDXH) IN: 18.00 x 11.00 x 46.00 LBS: 19.00 MM: 457.20 x 279.40 x 1168.40 KG: 8.62 LOCATION: FLOOR NOTES: MODEL NOT EXACT. PLACEHOLDER FOR EXISTING VACUUM. VOLTS: 110 HERTZ: 60 AMPS: 1.19 ELECT NOTE: REFERENCE VENDOR TECHNICAL DATA	KARCHER SENSOR XP 18 10126130	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
MECH NOTE: REFERENCE VENDOR TECHNICAL DATA								

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: BIOHAZARD WASTE
ROOM NUMBER: 414

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100859 WST02	2	CONTAINER, BIOHAZARD RED; 64 GALLON; 2 WHEEL; HINGED LID; 2-10"WHEELS; 2-4"CASTERS; PLASTIC (WDXH) IN: 24.25 x 32.00 x 44.25 LBS: 45.30 MM: 615.95 x 812.80 x 1123.95 KG: 20.55 LOCATION: FLOOR	TOTER INC. RMC64-00RED RMC64	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SHARED SERVICES
ROOM NAME: STAFF TOILET
ROOM NUMBER: 601

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: PUBLIC TOILET (QTY 2)
ROOM NUMBER: 803
ROOM COMMENT: RM 803, 804 NEAR THERAPY

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: DINING ROOM
ROOM NUMBER: 806

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		089875	2	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		092852	1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		057206 GLV01	2	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: RECREATION ROOM
ROOM NUMBER: 809

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1002050 RFZ02	1	REFRIGERATOR/FREEZER, UPRIGHT 18.3CF;TOP MOUNT FREEZER; WHITE; NO ICE (WDXDH) IN: 30.00 x 30.38 x 66.38 LBS: 164.00 MM: 762.00 x 771.66 x 1686.06 KG: 74.39 LOCATION: FLOOR SPEC COMMENTS: CLEARANCES: 3/8" SIDES AND TOP, 1" BACK VOLTS: 120 HERTZ: 60 AMPS: 15 WATS: 1800 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A CIRCUIT;3 PRONG GROUNDING PLUG	FRIGIDAIRE FFHT1835VW	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		1013974 RNG01	1	RANGE, ELECTRIC 30" INDUCTION RANGE WITH AIR FRYING OPTION (WDXDH) IN: 29.88 x 29.25 x 48.00 LBS: 250.00 MM: 758.96 x 742.95 x 1219.20 KG: 113.40 LOCATION: FLOOR NOTES: 2024.09.24 PLUG, NOT HARDWIRED. PLUGGED INTO AN OUTLET WITH SHUT OFF SWITCH REQUIRED - HGA TO PROVIDE CORRECT OUTLET ANCHOR DETAIL REQ: TBD VOLTS: 240 HERTZ: 60 AMPS: 40 DEDICATED CIRCUIT: YES ELECT NOTE: POWER SUPPLY CONNECTION AT MIDDLE BOTTOM REAR	FRIGIDAIRE GCRI3058AF GALLERY 30	OF/CI	NEW	GROUP: 1 FIXED
3		1015636 OVM02	1	OVEN, DOMESTIC, MICROWAVE, BUILT-IN 30 INCH OVER THE RANGE MICROWAVE OVEN WITH 1.8 CU. FT. CAPACITY, 1000 COOKING WATTS, DUCTED VENTING, 300 CFM, 10 POWER (WDXDH) IN: 29.88 x 15.75 x 16.38 LBS: 50.00 MM: 758.83 x 400.05 x 415.93 KG: 22.68 LOCATION: SPACE NOTES: 2024.08.30 BUILT-IN ABOVE STOVE. MUST INCLUDE 3-SEC CHILD LOCK. ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CHANGED TO OFCI VOLTS: 120 HERTZ: 60 AMPS: 15 WATS: 1000	FRIGIDAIRE FMOW1852AS 1.8 CU. FT OVER-THE-RANGE	OF/CI	NEW	GROUP: 1 FIXED
4		083455 COF01	1	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WDXDH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: RECREATION ROOM
ROOM NUMBER: 809

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS
				SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)			COLD WATER: YES PLUMB NOTE: WATER PRESSURES: 40PSIG TO 125PSIG 1/4"FPT PORTS IN/OUT FILTER WATER CONNECTION-3/4" MALE GARDEN HOSE CONNECTION
5		1006012 COF04	1	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WXDXH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW GROUP: 3 MINOR MOVABLE
6		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW GROUP: 3 MINOR MOVABLE
7		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW GROUP: 3 MINOR MOVABLE
8		092852	1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WXDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: RECREATION ROOM
ROOM NUMBER: 809

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		089875		1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: FAMILY ROOM
ROOM NUMBER: 811

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: LOCKERS, FEMALE
ROOM NUMBER: 812

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		089875		1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: LOCKERS, TOILET (QTY 2)
ROOM NUMBER: 813
ROOM COMMENT: RM 813, 849

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: MED ROOM
ROOM NUMBER: 814

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		104140 MED01	1	DISPENSER, MEDICATION, AUTOMATED 6 DRAWER; LCD MONITOR; KEYBOARD; BARCODE SCANNER; BIOMETRIC ACCESS SYSTEM; SECURE HALF AND FULL HEIGHT CUBIE POCKETS AND (WXDXH) IN: 23.00 x 27.00 x 55.00 LBS: 166.00 MM: 584.20 x 685.80 x 1397.00 KG: 75.30 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 3 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 1A NOM,3 AMPS MAX LOAD CURRENT EMERGENCY POWER REQUIRED	BD CAREFUSION PYXIS MEDSTATION ES 323 6 DRAWER MAIN BTU: 409 MECH NOTE: 409 BTU/HR	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
2		089012 MED02	1	DISPENSER, MEDICATION, AUTOMATED 7-DRAWER AUXILIARY;18-GAUGE SHEET METAL;REMOVABLE, LOCKING REAR PANEL (WXDXH) IN: 23.00 x 27.00 x 47.00 LBS: 133.00 MM: 584.20 x 685.80 x 1193.80 KG: 60.33 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: CONNECTS TO MAIN MED DISPENSER VIA CABLE ELECT NOTE: POWERED BY MAIN VIA CABLE CONNECTION	BD CAREFUSION PYXIS 7-DRAWER AUXILIARY MEDSTATION ES 7-DRAWER AUXILIARY	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
3		1012767 CRS01	1	CABINET, SUPPLY STORAGE MOBILE STORAGE CABINET; STEEL CONSTRUCTION WITH POWDER COAT FINISH; FOUR ADJUSTABLE SHELVES; 400LB SHELF CAP; 750LB TOTAL (WXDXH) IN: 48.00 x 24.00 x 84.00 LBS: 242.00 MM: 1219.20 x 609.60 x 2133.60 KG: 109.77 LOCATION: FLOOR NOTES: SEISMIC KIT CUSTOM UPON ORDER ANCHOR DETAIL REQ: YES CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	ULINE H-9092A CLEAR VIEW ASSEMBLED	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		105843 REF04	1	REFRIGERATOR, UNDERCOUNTER, MEDICAL GRADE 5.3 CF;SOLID FIELD REVERSIBLE DOOR;TEMP RANGE 2-10 DEG C;7" I.C3 COLOR TOUCHSCREEN INTERFACE;AUTO DEFROST;R600A ISOBUTANE (WXDXH) IN: 25.00 x 29.00 x 34.00 LBS: 174.00 MM: 635.00 x 736.60 x 863.60 KG: 78.93 LOCATION: FLOOR	HELMER LABS, INC. ILR105-GX I SERIES ADA	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: MED ROOM
ROOM NUMBER: 814

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				NOTES: 2024-09 SHD CONFIRMED REMOT CONNECTIONS NOT REQUIRED ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CAHNGED FROM OF/OI TO OF/VI DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ45 ETHERNET VOLTS: 115 HERTZ: 60 AMPS: 1.4 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 8-10' POWER CORD;12V RECHARGEABLE LEAD ACID BATTERY FOR DISPLAY & ALARM BACKUP	BTU: 460 MECH NOTE: REAR ACCESS PORT FOR EXT MONITORING PROBES			
5		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
6		105414 CRTM01	1	CART, MEDICATION ACS CART;KEYLESS PUSH LOCK;NO HANDLE DK CREME; CASTERS (WXDXH) IN: 26.00 x 24.00 x 44.00 LBS: 123.00 MM: 660.40 x 609.60 x 1117.60 KG: 55.80 LOCATION: FLOOR	CAPSA SOLUTIONS AVALO ACS 10 HIGH	OF/VI	EXISTING	GROUP: 3 MINOR MOVABLE
7		1020550 DSPP04B	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL CONTROLLED SUBSTANCE WASTE MANAGEMENT SYSTEM. COMPLIANCE WITH REGULATORY GUIDELINES. REDUCES OPPORTUNITIES FOR DRUG (WXDXH) IN: 15.50 x 9.50 x 12.50 LBS: 17.00 MM: 393.70 x 241.30 x 317.50 KG: 7.72 LOCATION: SPACE SPEC COMMENTS: WALL/COUNTER BRACKET INSTALLED SEPARATELY BY CONTRACTOR CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY 0085-013-000 SMART SINK	OF/OI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: MED ROOM
ROOM NUMBER: 814

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1020549 DSPP04A	1	BRACKET CACTUS SMART SINK MOUNTING BRACKET. TO INCLUDE: (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: WALL SPEC COMMENTS: BRACKET INSTALLED BY CONTRACTOR, SMART SINK SENT TO SITE FOR OWNER INSTALL CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY CACTUS SMART SINK MOUNTING BRACKET 0085-000-003	OF/CI	NEW	GROUP: 1 FIXED
9		1020123 MBOX01	1	BOX, RETURN, DRUGS LOCKABLE; WALL MOUNTED; BEIGE METAL ANTI-THEFT BAFFLE; BOTTOM IS (WXDXH) IN: 12.00 x 5.20 x 6.25 LBS: 9.00 MM: 304.80 x 132.08 x 158.75 KG: 4.09 LOCATION: WALL NOTES: 2024-09 SHD CONFIRMED WALL MOUNTED CHANGE CODE: C CHANGE NOTES: CHANGE FROM OF/OI COUNTER MNT TO OF/CI WALL MNT	HEALTH CARE LOGISTICS 3783 LOCKABLE RETURN DRUG BOX	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
10		087350	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL 8 GAL;SLIDE LID;10/CS;WHITE (WXDXH) IN: 15.50 x 11.00 x 17.75 LBS: 4.00 MM: 393.70 x 279.40 x 450.85 KG: 1.82 LOCATION: FLOOR	KENDALL HEALTHCARE, COVIDIEN 8851 PHARMA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
12		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: MED ROOM
ROOM NUMBER: 814

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: BEAUTY PARLOR
ROOM NUMBER: 815

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		056183 CHR05	1	CHAIR, SALON HYDRAULIC HEIGHT ADJUST;HEAVY DUTY BASE;RECLINING LOGO FOOTREST;CALF BARS (WDXDH) IN: 30.00 x 40.00 x 35.00 LBS: 0.00 MM: 762.00 x 1016.00 x 889.00 KG: 0.00 LOCATION: FLOOR ANCHOR DETAIL REQ: NO	COLLINS MANUFACTURING 9020 CONTINENTAL	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1008257 DRY01	1	DRYER, HAIR FEATURES LARGE HOOD; INCLUDES VARIABLE TIME AND TEMPERATURE SETTINGS; 1000 WATTS OF DYING POWER (WDXDH) IN: 17.50 x 17.50 x 72.00 LBS: 45.00 MM: 444.50 x 444.50 x 1828.80 KG: 20.42 LOCATION: FLOOR NOTES: 2024-09-30 SHD CONFIRMED MOBILE FLOOR UNIT CHANGE CODE: C CHANGE NOTES: CHANGED FROM OF/CI TO OF/OI VOLTS: 110 HERTZ: 60	COLLINS MANUFACTURING 1433S TEMPTURA ULTRA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXDH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		099002 VAC03	1	VACUUM STATIONARY VACUUM;AUTO/MANUAL FUNCTIONS 4.8 LITER CAPACITY;HIGH EFFICIENCY FILTRATION (WDXDH) IN: 13.00 x 7.50 x 20.50 LBS: 15.60 MM: 330.20 x 190.50 x 520.70 KG: 7.08 LOCATION: FLOOR VOLTS: 120 HERTZ: 60	EYE VAC EYEVAC PRO	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXDH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: BEAUTY PARLOR
ROOM NUMBER: 815

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		081549 WSH02	1	WASHER/DRYER, CLOTHES, DOMESTIC HE TOP LOAD;3.8 CF WASHER;FRONT LOAD 5.5 CF DRYER 8 WASH CYCLES;STAINLESS STEEL WASHER/DRYER (WXDXH) IN: 27.00 x 31.50 x 77.00 LBS: 275.00 MM: 685.80 x 800.10 x 1955.80 KG: 124.74 LOCATION: FLOOR ANCHOR DETAIL REQ: NO VOLTS: 240 HERTZ: 60 AMPS: 30 ELECT NOTE: 5.1 CONNECTED LOAD HEATING ELEMENT @ 240V 4,500	FRIGIDAIRE FFLE3911QW WHITE	OF/CI	NEW	GROUP: 1 FIXED
					VENT REQUIRED: YES MECH NOTE: 56FT MAX EXHAUST DUCT LENGTH POWER SUPPLY AND WATER INLET CONNECTION LOCATION:MIDDLE REAR	COLD WATER: YES HOT WATER: YES DRAIN REQUIRED: YES PLUMB NOTE: PRESSURE BTWN 10-120 PSI WITH NOT MORE THAN 10PSI H/C WATER FAUCETS MUST BE INSTALLED WITHIN 42" OF WASHER'S WATER INLET AND MUST BE 3/4" W/ THREADING FOR LAUNDRY HOSE CONNECTION TO CONNECT TO 48" INLET HOSES DRAIN STANDPIPE REQUIRES DIAMETER OF 1 1/4" MIN & HEIGHT OF 33"MIN/96"MAX ABOVE FLOOR,CAPABLE OF ELIMINATING 17 GALLONS PER MINUTE ATTACHED 52" DRAIN HOSE CAN REACH 74"H STANDPIPE WATER INLET CONNECTION LOCATION:MIDDLE REAR		

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SEMI-PRIVATE PATIENT BEDROOM (QTY 12)
ROOM NUMBER: 816
ROOM COMMENT: RM 816, 819, 821, 824, 826, 828, 852, 854, 856, 859, 861, 864

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1020630 BED02	24	BED, RESIDENTIAL, LONG TERM CARE RESIDENTIAL STYLE LONG TERM CARE BED. HEAD RAIL WITH CONTROLS, FOOT RAIL WITH NO CONTROLS, FOUR FLOOR BRAKES. SLEEP DECK (WDXDH) IN: 42.00 x 91.50 x 30.00 LBS: 275.00 MM: 1066.80 x 2324.10 x 762.00 KG: 124.74 LOCATION: FLOOR NOTES: 2025.05.12 QTY MAY CHANGE DURING PROCUREMENT VOLTS: 120 HERTZ: 60	HILL-ROM LTC-320 LTC320	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		093154 TBO01	24	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25";LOW PROFILE;TUNING FORK BASE DESIGN;SGL THERMOFOIL TOP TO INCLUDE: (WDXDH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1006224 CRS02	12	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXDH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		1004559	24	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXDH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	12	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXDH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SEMI-PRIVATE PATIENT BEDROOM (QTY 12)
ROOM NUMBER: 816
ROOM COMMENT: RM 816, 819, 821, 824, 826, 828, 852, 854, 856, 859, 861, 864

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		087035 DSP01	12	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
7		004363	36	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR NOTES: 1 PER BED AND 1 AT SINK	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: RESIDENT TOILET (QTY 12)
ROOM NUMBER: 817
ROOM COMMENT: RM 817, 820, 822, 825, 827, 829, 853, 855, 857, 860, 862, 865

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	12	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SOILED HOLDING 01
ROOM NUMBER: 818

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100858 TRK01	1	TRUCK/BIN, LINEN 500LB; 16 CUFT CAPACITY; MANUAL UNLOADING; BLUE (WXDXH) IN: 30.38 x 44.00 x 35.50 LBS: 60.50 MM: 771.66 x 1117.60 x 901.70 KG: 27.45 LOCATION: FLOOR NOTES: INCLUDES LID	TOTER INC. MMC16-90701 CUBE HEAVY DUTY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015414 TRK02	1	TRUCK, TILT, WASTE/LINEN 16 CUFT MANUAL CUBE TRUCK; INCLUDES LID (WXDXH) IN: 30.40 x 44.00 x 35.50 LBS: 60.00 MM: 772.16 x 1117.60 x 901.70 KG: 27.22 LOCATION: FLOOR NOTES: BLACK IN COLOR; INCLUDES LID	TOTER INC. MMC16-10836 BLACK	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: GENERAL STORAGE 01
ROOM NUMBER: 823

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	4	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1013976	2	SEAT, TOILET, RAISED FREESTANDING TOILET SEAT RAISER, WITH BACK SUPPORT (WXDXH) IN: 22.00 x 22.00 x 22.00 LBS: 8.60 MM: 558.80 x 558.80 x 558.80 KG: 3.91 LOCATION: SPACE NOTES: 2025.09.18 MOVED FROM ALCOVE 830 TO STORAGE ROOM PER SHD CHAIR WILL FLOAT BETWEEN RESIDENT ROOMS CHANGE CODE: C CHANGE NOTES: MOVED ROOMS	ETAC 81702020 SWIFT FREESTANDING	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
3		1013976	4	SEAT, TOILET, RAISED FREESTANDING TOILET SEAT RAISER, WITH BACK SUPPORT (WXDXH) IN: 22.00 x 22.00 x 22.00 LBS: 8.60 MM: 558.80 x 558.80 x 558.80 KG: 3.91 LOCATION: SPACE NOTES: 2025.09.18 MOVED FROM ALCOVE 830 TO STORAGE ROOM PER SHD CHAIR WILL FLOAT BETWEEN RESIDENT ROOMS CHANGE CODE: C CHANGE NOTES: MOVED ROOMS	ETAC 81702020 SWIFT FREESTANDING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1009029	1	CHAIR, COMMODE 350 LBS CAPACITY; WHITE; FOLDING; STANDARD SEATING PACKAGE OF 4 G1-200X1 (WXDXH) IN: 21.00 x 14.50 x 22.00 LBS: 13.00 MM: 533.40 x 368.30 x 558.80 KG: 5.90 LOCATION: FLOOR NOTES: 2025.09.18 MOVED FROM ALCOVE 830 TO STORAGE ROOM PER SHD CHAIR WILL FLOAT BETWEEN RESIDENT ROOMS	MEDLINE INDUSTRIES, INC. G1-200X4 G1-200X1	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: EQUIPMENT ALCOVE
ROOM NUMBER: 830

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023032 LFT05*	1	LIFT, PATIENT, MOBILE 500# CAP; POWER BASE; AUDIBLE ALERT; SIX-POINT CRADLE W/SPRING CLIPS; INCLUDES BATTERY AND EXTERNAL CHARGER STAND (WDXDH) IN: 45.80 x 52.00 x 76.30 LBS: 156.50 MM: 1163.32 x 1320.80 x 1938.02 KG: 70.99 LOCATION: FLOOR SPEC COMMENTS: EXTERNAL WIDTH WHEN OPEN: 45.8" EXTERNAL WIDTH WHEN CLOSED: 31.6" NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW ELECT NOTE: RECHARGEABLE BATTERY	DRIVE MEDICAL FLP500 LEVANTAR	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		074071 LFT02	1	LIFTER, PATIENT MOBILE;400LB CAP;ACTIVE STANDING/TRANSFER AID; FLIP DOWN SWIVEL (WDXDH) IN: 34.65 x 36.22 x 41.50 LBS: 65.00 MM: 880.11 x 919.99 x 1054.10 KG: 29.49 LOCATION: FLOOR SPEC COMMENTS: SPACE REQUIREMENT: 87" X 79" CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC	ARJOHUNTLEIGH NTB2000 SARA STEDY	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
3		1001245 MVS02	1	MONITOR, VITAL SIGNS MOBILE STAND;10.1" LCD TOUCHSCREEN DISPLAY;EWS (EARLY WARNING SYSTEM); MASIMO RAINBOW SET OPTIONS (SPO2, RRA, SPHB);THERMAL (WDXDH) IN: 21.00 x 21.00 x 49.85 LBS: 39.00 MM: 533.40 x 533.40 x 1266.19 KG: 17.70 LOCATION: FLOOR NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: CONNECT DIRECTLY INTO EMR OR HIS VIA HL-7 VOLTS: 120 HERTZ: 60 WATS: 60 ELECT NOTE: RECHARGEABLE LITHIUM ION BATTERY 11.1V 7800MAH	PHILIPS MEDICAL SYSTEMS 863380/989803176601 EARLYVUE VS30 W/ROLLSTAND	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: EQUIPMENT ALCOVE
ROOM NUMBER: 830

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1010678	1	CHAIR, COMMODE, BARIATRIC 1000 LB CAP; DROP ARM; SEAT HEIGHT ADJ 17.5"-21.5"; ADJUSTABLE LEGS; POWDER COATED FINISH STEEL FRAME IS RUST-RESISTANT; (WXDXH) IN: 31.00 x 20.50 x 31.50 LBS: 30.00 MM: 787.40 x 520.70 x 800.10 KG: 13.61 LOCATION: FLOOR SPEC COMMENTS: DEPTH AND HEIGHT ARE APPROXIMATE. VENDOR UNABLE TO GIVE EXACT MEASUREMENTS. NOTES: CHAIR WILL FLOAT BETWEEN RESIDENT ROOMS	MEDLINE INDUSTRIES, INC. G1-400BDX1 G1-400BDX1	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		1023049	1	CHAIR, COMMODE, SHOWER HYBRID COMMODE AND SHOWER CHAIR; ALUMINUM FRAME; TILTING; ADJUSTABLE HEADREST AND FOOTREST; 20" SEAT WIDTH; FOUR LOCKING 5" (WXDXH) IN: 24.00 x 40.00 x 41.00 LBS: 0.00 MM: 609.60 x 1016.00 x 1041.40 KG: 0.00 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW	HEALTHLINE MEDICAL EZEE LIFE 2970009301	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: NURSE STATION (QTY 2)
ROOM NUMBER: 833
ROOM COMMENT: RM 833, 868

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		056002 CRC04	1	CART, RESUSCITATION, CARDIAC CART, RESUSCITATION, CARDIAC; 6 DRAWERS;4-3", 1-6",1-12",PVC BUMPER;LOCKING BREAKAWAY BAR, TWIN (WXDXH) IN: 43.00 x 25.00 x 47.00 LBS: 165.00 MM: 1092.20 x 635.00 x 1193.80 KG: 74.85 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE;	ARMSTRONG MEDICAL INDUSTRIES AR-6 PREMIER W/ER PK	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		082533 DEF01	1	DEFIBRILLATOR R SERIES ALS DEFIBRILLATOR;AED W/MANUAL OVERRIDE;GUIDELINES 2010 COMPATIBLE;CODE READINESS TESTING SYSTEM;RECTILINEAR (WXDXH) IN: 10.50 x 12.50 x 8.20 LBS: 15.20 MM: 266.70 x 317.50 x 208.28 KG: 6.90 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: OPTIONAL WIFI: 802.11B USB PORT 2.0 COMPATIBLE VOLTS: 120 HERTZ: 60 ELECT NOTE: SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER EMERGENCY POWER	ZOLL MEDICAL CORPORATION 30720005201310012 R SERIES ALS	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
3		094808 ASP01	1	PUMP, SUCTION/ASPIRATOR, GENERAL, PORTABLE PORTABLE; RECHARGEABLE BATTERY; LED DISPLAY; HEPA FILTER; BEMIS DISPOSABLE CANISTER SYSTEM; HIGH FLOW; HIGH VACUUM; (WXDXH) IN: 13.00 x 6.30 x 12.40 LBS: 8.90 MM: 330.20 x 160.02 x 314.96 KG: 4.04 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 ELECT NOTE: ELECTRICAL SHOWN FOR BATTERY CHARGER	LAERDAL MEDICAL CORP. LSU 4000 78002001	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: NURSE STATION (QTY 2)
ROOM NUMBER: 833
ROOM COMMENT: RM 833, 868

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023001 TMP02	2	THERMOMETER, ELECTRONIC, WALL MOUNTED TEMPORAL ARTERY PROFESSIONAL THERMOMETER; KEYLESS WALL MOUNT W/HIDDEN RELEASE LEVER; NO CABLE (WDXDH) IN: 3.23 x 0.00 x 6.20 LBS: 0.00 MM: 82.05 x 0.00 x 157.48 KG: 0.00 LOCATION: WALL NOTES: 2025.05.12 UPDATED MODEL FROM 5000 TO 5000S WITH NO CABLE KEYLESS WALL MOUNT PER SHD CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: 9-VOLT BATTERY	EXERGEN CORPORATION TAT-5000S 124275S/134305	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
5		105759 GLU01	1	LAB ANALYZER, GLUCOSE TOUCH SCREEN;QR BARCODE SCANNER;1,000 TEST STORAGE; 4,000 USER STORAGE; 500 TEST STORAGE;NUMERIC OR PASS/FAIL (WDXDH) IN: 3.10 x 5.80 x 1.18 LBS: 0.49 MM: 78.74 x 147.32 x 29.98 KG: 0.23 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: DATA REQUIRED: ETHERNET NETWORK DROP OR 802.11 A/B/G/N WIRELESS ACCESS. SEE CUTSHEET FOR FURTHER DETAILS. VOLTS: 120 HERTZ: 60 AMPS: .6 ELECT NOTE: STATSTRIP: 3.7V LI POLYMER RECHARGEABLE BATTERY ELECTRICAL INFO IS FOR DOCKING STATION	NOVA BIOMEDICAL 44993 STATSTRIP GLU 2	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
6		105759 GLU01	1	LAB ANALYZER, GLUCOSE TOUCH SCREEN;QR BARCODE SCANNER;1,000 TEST STORAGE; 4,000 USER STORAGE; 500 TEST STORAGE;NUMERIC OR PASS/FAIL (WDXDH) IN: 3.10 x 5.80 x 1.18 LBS: 0.49 MM: 78.74 x 147.32 x 29.98 KG: 0.23 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: DATA REQUIRED: ETHERNET NETWORK DROP OR 802.11 A/B/G/N WIRELESS ACCESS. SEE CUTSHEET FOR FURTHER DETAILS. VOLTS: 120 HERTZ: 60 AMPS: .6 ELECT NOTE: STATSTRIP: 3.7V LI POLYMER RECHARGEABLE BATTERY ELECTRICAL INFO IS FOR DOCKING STATION	NOVA BIOMEDICAL 44993 STATSTRIP GLU 2	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: ALCOVE, NOUR (QTY 2)
ROOM NUMBER: 833A
ROOM COMMENT: RM 833A, 868A; LOCATED BEHIND NURSE STATIONS

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		091485 ICE01	2	ICE MAKER W/WATER DISPENSER 12LB STORAGE;260LB PROD/24 HRS;SS EXTERIOR;EXTERNAL AIR FILTER;AGION ANTIMICROBIAL;R-134A REFRIGERANT;H2 NUGGET STYLE ICE (WDXDH) IN: 17.00 x 25.00 x 35.00 LBS: 183.00 MM: 431.80 x 635.00 x 889.00 KG: 83.01 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 7 ELECT NOTE: SINGLE PHASE;15A MAX FUSE SEE VENDOR TECHNICAL INFORMATION	SCOTSMAN ICE SYSTEMS HID312A-1 MERIDIAN AIR COOL	OF/CI	NEW	GROUP: 1 FIXED
2		1021297	2	FILTER VALVE-IN-HEAD WATER FILTER SYSTEM. INSTALLS ON THE 3/4-INCH MAIN WATER LINE AT THE POINT OF ENTRY. FILTER HOUSING DESIGNED (WDXDH) IN: 5.00 x 5.00 x 13.25 LBS: 0.50 MM: 127.00 x 127.00 x 336.55 KG: 0.23 LOCATION: WALL CHANGE CODE: A CHANGE NOTES: ADDED PER SENECA, HGA HAS ALREADY ACCOUNTED FOR FI	CULLIGAN WATER SYSTEMS HF-360	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		1002050 RFZ02	2	REFRIGERATOR/FREEZER, UPRIGHT 18.3CF;TOP MOUNT FREEZER; WHITE; NO ICE (WDXDH) IN: 30.00 x 30.38 x 66.38 LBS: 164.00 MM: 762.00 x 771.66 x 1686.06 KG: 74.39 LOCATION: FLOOR SPEC COMMENTS: CLEARANCES: 3/8" SIDES AND TOP, 1" BACK VOLTS: 120 HERTZ: 60 AMPS: 15 WATS: 1800 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A CIRCUIT;3 PRONG GROUNDING PLUG	FRIGIDAIRE FFHT1835VW	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: ALCOVE, NOUR (QTY 2)
ROOM NUMBER: 833A
ROOM COMMENT: RM 833A, 868A; LOCATED BEHIND NURSE STATIONS

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		055770 OVM01	2	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WXDXH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
5		083455 COF01	2	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WXDXH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED
6		1006012 COF04	2	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WXDXH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1004090	2	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: ALCOVE, NOUR (QTY 2)
ROOM NUMBER: 833A
ROOM COMMENT: RM 833A, 868A; LOCATED BEHIND NURSE STATIONS

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1014420	2	WASTE RECEPTACLE 13GAL;OPEN TOP;BLUE (WDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 2026699 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: CENTRAL BATHING (QTY 2)
ROOM NUMBER: 834
ROOM COMMENT: RM 834, 869

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		088075 STR04	1	STRETCHER, SHOWER PVC;WATER PROOF FOAM PAD;DROP SIDE (WDXDH) IN: 28.00 x 79.00 x 34.00 LBS: 130.00 MM: 711.20 x 2006.60 x 863.60 KG: 58.97 LOCATION: FLOOR NOTES: ONLY IN ROOM 819	LUMEX, INC. GRF-8005	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
2		1001188 CBW01	2	CABINET, WARMING 2.5 CUBIC FT; 6 BLANKET CAPACITY; COUNTERTOP MODEL; STACKABLE; NO (WDXDH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP SPEC COMMENTS: CLEARANCES: TOP/LEFT/RIGHT: 1", BACK 3" NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 5 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 0.6KW;15A;125V PLUG;HOSPITAL GRADE PLUG	ENTHERMICS MEDICAL SYSTEMS TITAN SERIES EC250 BTU: 2050	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
3		105040 CBW02	2	CABINET, WARMING, WIPES FOR USE WITH COMFORT PERSONAL CLEANSING BRAND;ACCOMMODATES 12 INDIVIDUAL PACKAGES;125 DEGREES F TARGET PRODUCT (WDXDH) IN: 14.69 x 15.25 x 16.88 LBS: 30.00 MM: 373.13 x 387.35 x 428.76 KG: 13.61 LOCATION: COUNTERTOP NOTES: SITS ON TOP OF CBW01 ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 7 DEDICATED CIRCUIT: YES ELECT NOTE: SINGLE PHASE;DEDICATED CIRCUIT RECOMMENDED	SAGE PRODUCTS INC. 7937 12-SLOT	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		1003921 LHM01	2	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXDH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: CENTRAL BATHING (QTY 2)
ROOM NUMBER: 834
ROOM COMMENT: RM 834, 869

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		057206 GLV01	2	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: TOILET, CENTRAL BATHING (QTY 2)
ROOM NUMBER: 835
ROOM COMMENT: RM 835, 870

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: DON OFFICE
ROOM NUMBER: 836

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1005565 BIN01	1	BIN, SHREDDING SHREDDER COLLECTION BIN (WXDXH) IN: 19.00 x 16.00 x 27.00 LBS: 35.00 MM: 482.60 x 406.40 x 685.80 KG: 15.88 LOCATION: FLOOR SPEC COMMENTS: LEASED ITEM CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	RAPID SHRED SHORT CONSOLE	OF/OI	LEASED	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: DON STORAGE
ROOM NUMBER: 837

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		084955 TCRT01	1	CART, TELEMEDICINE MOBILE TELE-COMMUNICATION SYSTEM;FULLY INTEGRATED 24" HD LCD DISPLAY;INTEGRATED CAMERA & MICROPHONE (WXDXH) IN: 23.00 x 27.00 x 71.00 LBS: 230.00 MM: 584.20 x 685.80 x 1803.40 KG: 104.33 LOCATION: FLOOR NOTES: 2025.09.18 KEEP ON LIST FOR BUDGET, WILL DETERMINE PREFERRED MAKE/MODEL DURING PROCUREMENT DATA HARDWIRED: YES WIRELESS: NO VOLTS: 125 HERTZ: 60 ELECT NOTE: 45 AH LITHIUM IRON BATTERY RECHARGED BY HOSPITAL GRADE PLUG ONLY	AVIZIA - CA750	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SPECIAL PURPOSE ROOM (PRIVATE & ADA)
ROOM NUMBER: 840

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1020630 BED02	1	BED, RESIDENTIAL, LONG TERM CARE RESIDENTIAL STYLE LONG TERM CARE BED. HEAD RAIL WITH CONTROLS, FOOT RAIL WITH NO CONTROLS, FOUR FLOOR BRAKES. SLEEP DECK (WXDXH) IN: 42.00 x 91.50 x 30.00 LBS: 275.00 MM: 1066.80 x 2324.10 x 762.00 KG: 124.74 LOCATION: FLOOR PHASE: EARLY-PO/LEASED NOTES: 2025.05.12 QTY MAY CHANGE DURING PROCUREMENT VOLTS: 120 HERTZ: 60 PO REQ #: 944A-00008 CLIENT PO #: SNCA1309	HILL-ROM LTC-320 LTC320	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WXDXH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
4		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SPECIAL PURPOSE ROOM (PRIVATE & ADA)
ROOM NUMBER: 840

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		093154 TB001	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25";LOW PROFILE;TUNING FORK BASE DESIGN;SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: TOILET, PRIVATE RESIDENT (QTY 2)
ROOM NUMBER: 841
ROOM COMMENT: RM 841, 843

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: QUIET ROOM (PRIVATE & ADA)
ROOM NUMBER: 842

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1020630 BED02	1	BED, RESIDENTIAL, LONG TERM CARE RESIDENTIAL STYLE LONG TERM CARE BED. HEAD RAIL WITH CONTROLS, FOOT RAIL WITH NO CONTROLS, FOUR FLOOR BRAKES. SLEEP DECK (WXDXH) IN: 42.00 x 91.50 x 30.00 LBS: 275.00 MM: 1066.80 x 2324.10 x 762.00 KG: 124.74 LOCATION: FLOOR PHASE: EARLY-PO/LEASED NOTES: 2025.05.12 QTY MAY CHANGE DURING PROCUREMENT VOLTS: 120 HERTZ: 60 PO REQ #: 944A-00007 CLIENT PO #: SNCA1256	HILL-ROM LTC-320 LTC320	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WXDXH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: QUIET ROOM (PRIVATE & ADA)
ROOM NUMBER: 842

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		093154 TB001	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25";LOW PROFILE;TUNING FORK BASE DESIGN;SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: GENERAL STORAGE 02
ROOM NUMBER: 844

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	4	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR NOTES: 2025.06.04 REDUCED FROM 5 TO 4 PER DH ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: REDUCED FROM 5 TO 4	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1006551 RCK06	1	RACK, CLOTHES, MOBILE SINGLE ROLLING CLOTHES RACK (WXDXH) IN: 50.00 x 22.50 x 65.00 LBS: 28.00 MM: 1270.00 x 571.50 x 1651.00 KG: 12.71 LOCATION: FLOOR	ULINE H-1546	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1006551 RCK06	1	RACK, CLOTHES, MOBILE SINGLE ROLLING CLOTHES RACK (WXDXH) IN: 50.00 x 22.50 x 65.00 LBS: 28.00 MM: 1270.00 x 571.50 x 1651.00 KG: 12.71 LOCATION: FLOOR	ULINE H-1546	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: CLEAN SUPPLY
ROOM NUMBER: 847

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		089770 PAN02	6	LOUVER PANEL, WALL MOUNTED 175LB LOAD CAPACITY;ADJ BIN STORAGE;16 GAUGE STEEL W/OYSTER WHITE / BEIGE BAKED-ENAMEL FINISH; (WDXH) IN: 36.00 x 1.00 x 19.00 LBS: 13.00 MM: 914.40 x 25.40 x 482.60 KG: 5.90 LOCATION: WALL ANCHOR DETAIL REQ: YES	QUANTUM STORAGE SYSTEMS QLP-3619HC	OF/CI	NEW	GROUP: 1 FIXED
2	No image available	506580	1	ALLOWANCE, BIN (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		065434 SHW03	3	SHELVING, WIRE 5 SHELF W/SOLID GALV BOTTOM SHELF;STATIONARY TO INCLUDE: (WDXH) IN: 60.00 x 24.00 x 74.50 LBS: 131.00 MM: 1524.00 x 609.60 x 1892.30 KG: 59.43 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. A2460NC/2460FG/74P SASE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		084882 CRK02	2	CART, CYLINDER, MEDICAL GAS 4 CAPACITY TILT BACK CART; E/D SIZE; W/HEAVY DUTY 5" CASTERS (WDXH) IN: 11.00 x 11.00 x 39.00 LBS: 21.00 MM: 279.40 x 279.40 x 990.60 KG: 9.53 LOCATION: FLOOR	W.T. FARLEY, INC. - CR-DC04E	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: LOCKERS, MALE
ROOM NUMBER: 848

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		089875		1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: GENERAL STORAGE 03
ROOM NUMBER: 858

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	4	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: SOILED HOLDING 02
ROOM NUMBER: 863

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100858 TRK01	1	TRUCK/BIN, LINEN 500LB; 16 CUFT CAPACITY; MANUAL UNLOADING; BLUE (WXDXH) IN: 30.38 x 44.00 x 35.50 LBS: 60.50 MM: 771.66 x 1117.60 x 901.70 KG: 27.45 LOCATION: FLOOR NOTES: INCLUDES LID	TOTER INC. MMC16-90701 CUBE HEAVY DUTY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015414 TRK02	1	TRUCK, TILT, WASTE/LINEN 16 CUFT MANUAL CUBE TRUCK; INCLUDES LID (WXDXH) IN: 30.40 x 44.00 x 35.50 LBS: 60.00 MM: 772.16 x 1117.60 x 901.70 KG: 27.22 LOCATION: FLOOR NOTES: BLACK IN COLOR; INCLUDES LID	TOTER INC. MMC16-10836 BLACK	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 871

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		052851 HKP01	1	CART, HOUSEKEEPING SS;2 SHELF;FOLDING PLATFORM W/LOCKING DOOR;ZIP BAG;TRASH LID;TOOL BRACKET W/4 HANDLES & 3 HOOKS;TRAY TOP (WDXH) IN: 23.00 x 36.00 x 42.50 LBS: 32.00 MM: 584.20 x 914.40 x 1079.50 KG: 14.52 LOCATION: FLOOR	ROYCE ROLLS RINGER CO. F36 FOLDING F36 FOLDING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: OFFICE
ROOM NUMBER: 872

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: EQUIPMENT ALCOVE
ROOM NUMBER: 874

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023032 LFT05*	1	LIFT, PATIENT, MOBILE 500# CAP; POWER BASE; AUDIBLE ALERT; SIX-POINT CRADLE W/SPRING CLIPS; INCLUDES BATTERY AND EXTERNAL CHARGER STAND (WXDXH) IN: 45.80 x 52.00 x 76.30 LBS: 156.50 MM: 1163.32 x 1320.80 x 1938.02 KG: 70.99 LOCATION: FLOOR SPEC COMMENTS: EXTERNAL WIDTH WHEN OPEN: 45.8" EXTERNAL WIDTH WHEN CLOSED: 31.6" NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: RECHARGEABLE BATTERY	DRIVE MEDICAL FLP500 LEVANTAR	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		074071 LFT02	1	LIFTER, PATIENT MOBILE;400LB CAP;ACTIVE STANDING/TRANSFER AID; FLIP DOWN SWIVEL (WXDXH) IN: 34.65 x 36.22 x 41.50 LBS: 65.00 MM: 880.11 x 919.99 x 1054.10 KG: 29.49 LOCATION: FLOOR SPEC COMMENTS: SPACE REQUIREMENT: 87" X 79" NOTES: 2025.09.09 ADDED PER SHD REVIEW CHANGE CODE: A CHANGE NOTES: 2025.09.09 ADDED PER SHD REVIEW	ARJOHUNTLEIGH NTB2000 SARA STEDY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		1023049	1	CHAIR, COMMODE, SHOWER HYBRID COMMODE AND SHOWER CHAIR; ALUMINUM FRAME; TILTING; ADJUSTABLE HEADREST AND FOOTREST; 20" SEAT WIDTH; FOUR LOCKING 5" (WXDXH) IN: 24.00 x 40.00 x 41.00 LBS: 0.00 MM: 609.60 x 1016.00 x 1041.40 KG: 0.00 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW	HEALTHLINE MEDICAL EZEE LIFE 2970009301	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		053947 CRI01	1	CART, ISOLATION 4 DRAWER;YELLOW;BASIC CART W/DRAWER; CONFIGURATION OF 1 EA.3" (WXDXH) IN: 33.50 x 25.00 x 46.25 LBS: 165.00 MM: 850.90 x 635.00 x 1174.75 KG: 74.85 LOCATION: FLOOR	ARMSTRONG MEDICAL INDUSTRIES AKL-Y-4 TWO TONE	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SKILLED NURSING
ROOM NAME: EQUIPMENT ALCOVE
ROOM NUMBER: 874

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1001245 MVS02	1	MONITOR, VITAL SIGNS MOBILE STAND;10.1" LCD TOUCHSCREEN DISPLAY;EWS (EARLY WARNING SYSTEM); MASIMO RAINBOW SET OPTIONS (SPO2, RRA, SPHB);THERMAL (WXDXH) IN: 21.00 x 21.00 x 49.85 LBS: 39.00 MM: 533.40 x 533.40 x 1266.19 KG: 17.70 LOCATION: FLOOR NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: NO WIRELESS: YES DATA NOTE: CONNECT DIRECTLY INTO EMR OR HIS VIA HL-7 VOLTS: 120 HERTZ: 60 WATS: 60 ELECT NOTE: RECHARGEABLE LITHIUM ION BATTERY 11.1V 7800MAH	PHILIPS MEDICAL SYSTEMS 863380/989803176601 EARLYVUE VS30 W/ROLLSTAND	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
6		091214 SCL04*	1	SCALE, WHEELCHAIR PORTABLE FOLD UP SCALE;DIGITAL COLUMN;WHEELCHAIR RAMP;1000LB (WXDXH) IN: 41.40 x 43.30 x 50.00 LBS: 91.10 MM: 1051.56 x 1099.82 x 1270.00 KG: 41.33 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: ADDED DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: WIRED OR WIRELESS AVAILABLE RS-232 SERIAL PORT, USB PORT VOLTS: 100-240 HERTZ: 50-60 AMPS: 1.25 ELECT NOTE: MEDICAL GRADE POWER SUPPLY (MV1PWR); 12VDC OR (6) AA ALKALINE, NICAD, OR NIMH BATTERIES	DETECTO SCALE COMPANY 6550	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1023063 SCL05	1	SCALE, STAND-ON HANDRAILS; LB/KG; DIGITAL; BATTERIES, ADAPTER AND COMM CABLE INCLUDED (WXDXH) IN: 20.50 x 20.50 x 0.00 LBS: 0.00 MM: 520.70 x 520.70 x 0.00 KG: 0.00 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: ADDED ELECT NOTE: (4) AAA BATTERIES; AC ADAPTER	SALTER BRECKNELL MS140-300	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

END OF DEPARTMENT

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: REGISTRATION
ROOM NUMBER: 107

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1005565 BIN01	1	BIN, SHREDDING SHREDDER COLLECTION BIN (WXDXH) IN: 19.00 x 16.00 x 27.00 LBS: 35.00 MM: 482.60 x 406.40 x 685.80 KG: 15.88 LOCATION: FLOOR SPEC COMMENTS: LEASED ITEM CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	RAPID SHRED SHORT CONSOLE	OF/OI	LEASED	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PRIVATE REGISTRATION
ROOM NUMBER: 108

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: EQUIPMENT ALCOVE
ROOM NUMBER: 605

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		093154 TBO01	1	TABLE, OVERBED HEIGHT ADJ FROM 29.25"-46.25";LOW PROFILE;TUNING FORK BASE DESIGN;SGL THERMOFOIL TOP TO INCLUDE: (WXDXH) IN: 19.50 x 37.70 x 46.70 LBS: 48.00 MM: 495.30 x 957.58 x 1186.18 KG: 21.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY COLOR PREFERENCE DURING PROCUREMENT	HILL-ROM OBT-635 ART OF CARE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 606

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXDH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXDH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		052851 HKP01	1	CART, HOUSEKEEPING SS;2 SHELF;FOLDING PLATFORM W/LOCKING DOOR;ZIP BAG;TRASH LID;TOOL BRACKET W/4 HANDLES & 3 HOOKS;TRAY TOP (WDXDH) IN: 23.00 x 36.00 x 42.50 LBS: 32.00 MM: 584.20 x 914.40 x 1079.50 KG: 14.52 LOCATION: FLOOR	ROYCE ROLLS RINGER CO. F36 FOLDING F36 FOLDING	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 3
ROOM NUMBER: 607

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON ELECTRIC, STANDARD MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1015744 PHY05	1	MONITOR, PHYSIOLOGICAL 9" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXH) IN: 10.70 x 7.00 x 11.30 LBS: 12.30 MM: 271.78 x 177.80 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866060_NAM INTELLIVUE MX400	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY05, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 3
ROOM NUMBER: 607

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		006656 MAY01	1	STAND, MAYO FOOT OPERATED; LARGE TRAY; STAINLESS STEEL; 2" CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY; REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ; HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 3
ROOM NUMBER: 607

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		093116	1	REGULATOR, SUCTION, CONTINUOUS TWO-MODE W/DIAMOND & LOCKING GLAND OVERFLOW SAFETY TRAP (WDXH) IN: 3.00 x 5.00 x 7.00 LBS: 1.10 MM: 76.20 x 127.00 x 177.80 KG: 0.50 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; ELECT NOTE: 2-3AA BATTERIES	OHIO MEDICAL CORPORATION 8701-1225-905	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 3
ROOM NUMBER: 607

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		074363 CRS04	1	CART, PROCEDURE 5-DRAWER (3-3", 1-6", 1-9");SURESEAL AUTO LOCKING;MASTER CODE FOR ALL PROGRAMMING;UP TO 250 USER CODES POSSIBLE AUTO-LOCK (WXDXH) IN: 22.00 x 21.00 x 35.00 LBS: 70.10 MM: 558.80 x 533.40 x 889.00 KG: 31.80 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT ADDED T INSTRUCTION OF JEREMY BRAUNSCHWEIG 4/16 EMAIL. SURGICAL TEAM APPROVED VIA EMAIL 4/17. ELECT NOTE: BATTERY OPERATED KEYPAD	ARMSTRONG MEDICAL INDUSTRIES AMC-4-COLOR MINI-CART ALUMINUM	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 2
ROOM NUMBER: 608

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON ELECTRIC, STANDARD MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1015744 PHY05	1	MONITOR, PHYSIOLOGICAL 9" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXH) IN: 10.70 x 7.00 x 11.30 LBS: 12.30 MM: 271.78 x 177.80 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866060_NAM INTELLIVUE MX400	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY05, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 2
ROOM NUMBER: 608

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		006656 MAY01	1	STAND, MAYO FOOT OPERATED; LARGE TRAY; STAINLESS STEEL; 2" CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY; REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ; HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 2
ROOM NUMBER: 608

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		093116	1	REGULATOR, SUCTION, CONTINUOUS TWO-MODE W/DIAMOND & LOCKING GLAND OVERFLOW SAFETY TRAP (WDXH) IN: 3.00 x 5.00 x 7.00 LBS: 1.10 MM: 76.20 x 127.00 x 177.80 KG: 0.50 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; ELECT NOTE: 2-3AA BATTERIES	OHIO MEDICAL CORPORATION 8701-1225-905	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 2
ROOM NUMBER: 608

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		074363 CRS04	1	CART, PROCEDURE 5-DRAWER (3-3", 1-6", 1-9");SURESEAL AUTO LOCKING;MASTER CODE FOR ALL PROGRAMMING;UP TO 250 USER CODES POSSIBLE AUTO-LOCK (WXDXH) IN: 22.00 x 21.00 x 35.00 LBS: 70.10 MM: 558.80 x 533.40 x 889.00 KG: 31.80 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT ADDED T INSTRUCTION OF JEREMY BRAUNSCHWEIG 4/16 EMAIL. SURGICAL TEAM APPROVED VIA EMAIL 4/17. ELECT NOTE: BATTERY OPERATED KEYPAD	ARMSTRONG MEDICAL INDUSTRIES AMC-4-COLOR MINI-CART ALUMINUM	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 1
ROOM NUMBER: 609

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON ELECTRIC, STANDARD MATTRESS PER SHD CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1015744 PHY05	1	MONITOR, PHYSIOLOGICAL 9" TOUCH SCREEN; SUPPORTS 4WAVES AND 3 INVASIVE PRESSURES; INCLUDES 12-LEAD ECG, SPO2, NIBP; WALL MOUNTED VIA 19" SEISMIC (WDXH) IN: 10.70 x 7.00 x 11.30 LBS: 12.30 MM: 271.78 x 177.80 x 287.02 KG: 5.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70 DEDICATED CIRCUIT: YES	PHILIPS MEDICAL SYSTEMS 866060_NAM INTELLIVUE MX400	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		1023582 ARM01*	1	MONITOR ACCESSORY, MOUNTING ASSEMBLY, WALL VARIABLE HEIGHT MOUNTING ARM FOR MX400-450 & MP20-50 TO INCLUDE: (WDXH) IN: 12.90 x 24.50 x 24.80 LBS: 21.10 MM: 327.66 x 622.30 x 629.92 KG: 9.58 LOCATION: WALL NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR INSTALL RESPONSIBILITY CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY05, ADDED ON SEPERATE LINE	PHILIPS MEDICAL SYSTEMS 989805710348 VHM-PL	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 1
ROOM NUMBER: 609

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		081850 WLC01*	1	WALL CHANNEL 19" SEISMIC WALL CHANNEL; ACCOMMODATES PIVOTING OR VARIABLE HEIGHT MONITOR ARMS; INCLUDES MOUNTING HARDWARE AND (WDXDH) IN: 3.50 x 1.00 x 19.00 LBS: 11.00 MM: 88.90 x 25.40 x 482.60 KG: 4.99 LOCATION: WALL SPEC COMMENTS: WALL CHANNEL IS PLACED 56" +/- 4" ABOVE THE FLOOR. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. WAS PART OF PHY04 (PHYS MON, ARM, AND WALL CHANNEL), ADDED ON SEPARATE LINE FOR OFCI CHANGE CODE: C CHANGE NOTES: WAS PART OF PHY04, ADDED ON SEPARATE LINE OF/CI	GCX CORPORATION POLYMOUNT 19 WC-0002-05	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY; SINGLE STEP TITRATION LIMIT; MULTIPLE MEDS ERROR PREVENTION (WDXDH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
6		006656 MAY01	1	STAND, MAYO FOOT OPERATED; LARGE TRAY; STAINLESS STEEL; 2" CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY; REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ; HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXDH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 1
ROOM NUMBER: 609

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1004559	1	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		093116	1	REGULATOR, SUCTION, CONTINUOUS TWO-MODE W/DIAMOND & LOCKING GLAND OVERFLOW SAFETY TRAP (WDXH) IN: 3.00 x 5.00 x 7.00 LBS: 1.10 MM: 76.20 x 127.00 x 177.80 KG: 0.50 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; ELECT NOTE: 2-3AA BATTERIES	OHIO MEDICAL CORPORATION 8701-1225-905	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023253	1	BRACKET, SUCTION CANISTER 1500CC SUCTION CANISTER O RING HOLDER (WDXH) IN: 8.00 x 8.00 x 4.00 LBS: 0.00 MM: 203.20 x 203.20 x 101.60 KG: 0.00 LOCATION: SPACE NOTES: 2025.09.24 UPDATED FROM CANISTER AND HOLDER TO HOLDER ONLY 1500CC PER SHD. UPDATED FROM OF/CI TO OF/OI CHANGE CODE: C CHANGE NOTES: UDPATED SPEC, UPDATED FROM OF/CI TO OFOI	ALLIED HEALTHCARE PRODUCTS INC. 422935 20-02-0150	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
12		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PACU 1
ROOM NUMBER: 609

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
13		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		094274	1	WASTE RECEPTACLE, STEP-ON BIOHAZARD; STEP-ON CONTAINER; RED; 8 GALLON; PLASTIC; W/LID (WXDXH) IN: 15.75 x 16.25 x 17.13 LBS: 7.90 MM: 400.05 x 412.75 x 435.11 KG: 3.59 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614300RED STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
15		094275	1	WASTE RECEPTACLE, STEP-ON 8 GALLON; PLASTIC; BEIGE; STEP-ON PEDAL; W/ SLOW CLOSE LID (WXDXH) IN: 16.25 x 15.75 x 17.13 LBS: 5.63 MM: 412.75 x 400.05 x 435.11 KG: 2.56 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG614300BEIG LEGACY STEP-ON	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		074363 CRS04	1	CART, PROCEDURE 5-DRAWER (3-3", 1-6", 1-9");SURESEAL AUTO LOCKING;MASTER CODE FOR ALL PROGRAMMING;UP TO 250 USER CODES POSSIBLE AUTO-LOCK (WXDXH) IN: 22.00 x 21.00 x 35.00 LBS: 70.10 MM: 558.80 x 533.40 x 889.00 KG: 31.80 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT ADDED T INSTRUCTION OF JEREMY BRAUNSCHWEIG 4/16 EMAIL. SURGICAL TEAM APPROVED VIA EMAIL 4/17. ELECT NOTE: BATTERY OPERATED KEYPAD	ARMSTRONG MEDICAL INDUSTRIES AMC-4-COLOR MINI-CART ALUMINUM	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: SUPERVISOR'S OFFICE
ROOM NUMBER: 610

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PATIENT CHANGE
ROOM NUMBER: 611

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: PATIENT TOILET
ROOM NUMBER: 612

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: CLEAN SUPPLY
ROOM NUMBER: 614

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	2	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		066552 PAN01	3	LOUVER PANEL, WALL MOUNTED STORAGE PANEL; 16 GAUGE STEEL W/ GRAY ENAMEL FINISH LOUVERED FOR BINS; 250LB LOAD CAPACITY ONCE FASTENED SECURELY TO (WDXH) IN: 48.00 x 1.00 x 19.00 LBS: 18.00 MM: 1219.20 x 25.40 x 482.60 KG: 8.17 LOCATION: WALL ANCHOR DETAIL REQ: YES	QUANTUM STORAGE SYSTEMS QLP-4819 ULTRA HANGING SYSTEMS	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3	No image available	506580	1	ALLOWANCE, BIN (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 SPEC COMMENTS: SPECIFICATIONS NEED TO BE DETERMINED. CANNOT BID/QUOTE. NOTES: 2025.09.18 SIZES/COLORS TO BE SELECTED DURING PROCUREMENT	ZZZ - GENERIC DESCRIPTIONS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		055768 HYP01	2	HYPERTHERMIA UNIT CONTAINS BLOWER;HEATING ELEMENT;HAND CONTROLLER ATTACHES TO PATIENT GOWN VIA HOSE;WALL MOUNT;ADJUSTABLE AIR TEMP & (WDXH) IN: 7.70 x 4.00 x 13.00 LBS: 7.00 MM: 195.58 x 101.60 x 330.20 KG: 3.18 LOCATION: WALL SPEC COMMENTS: ; VOLTS: 120 HERTZ: 60 AMPS: 4.6 ELECT NOTE: FUSE 6.0 A ;15-FOOT;SJT, 3COND., 15 A	ARIZANT HEALTHCARE, A 3M COMPANY 87500/90080 BAIR PAWS 875 BTU: 750	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
5		1001188 CBW01	1	CABINET, WARMING 2.5 CUBIC FT; 6 BLANKET CAPACITY; COUNTERTOP MODEL; STACKABLE; NO (WDXH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP	ENTHERMICS MEDICAL SYSTEMS TITAN SERIES EC250	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: CLEAN SUPPLY
ROOM NUMBER: 614

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
SPEC COMMENTS: CLEARANCES: TOP/LEFT/RIGHT: 1", BACK 3" NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO LOCATED IN THE ALCOVE NEXT TO THE CLEAN SUPPLY ROOM ENTRANCE DOOR ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 5 BTU: 2050 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 0.6KW;15A;125V PLUG;HOSPITAL GRADE PLUG								
6		105040 CBW02	1	CABINET, WARMING, WIPES FOR USE WITH COMFORT PERSONAL CLEANSING BRAND;ACCOMMODATES 12 INDIVIDUAL PACKAGES;125 DEGREES F TARGET PRODUCT (WXDXH) IN: 14.69 x 15.25 x 16.88 LBS: 30.00 MM: 373.13 x 387.35 x 428.76 KG: 13.61 LOCATION: COUNTERTOP NOTES: SITS ON TOP OF CBW01 ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 7 DEDICATED CIRCUIT: YES ELECT NOTE: SINGLE PHASE;DEDICATED CIRCUIT RECOMMENDED	SAGE PRODUCTS INC. 7937 12-SLOT	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
7		059223	3	PUMP, COMPRESSION LEG & FOOT COMPRESSION; COMPRESSION SLEEVE CONTAINS 3 BLADDERS TO DELIVER CIRCUMFERENTIAL COMPRESSION ALL-IN-ONE (WXDXH) IN: 7.70 x 7.30 x 6.80 LBS: 5.00 MM: 195.58 x 185.42 x 172.72 KG: 2.27 LOCATION: COUNTERTOP CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 WATS: 50 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 50VA;13' POWER CORD WITH HOSPITAL GRADE, REGION-SPECIFIC CORDAGE/PLUG OUTLET MUST BE GROUNDED RECHARGEABLE LITHIUM ION BATTERY 10.8V, 2200MAH	KENDALL HEALTHCARE, COVIDIEN 29525 700 SERIES	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015781 MTC02	1	MONITOR, CENTRAL STATION PIIC IX SURVEILLANCE CENTRAL MONITORING; 23" TOUCH; CPU TO INCLUDE: (WXDXH) IN: 16.00 x 8.20 x 16.00 LBS: 30.00 MM: 406.40 x 208.28 x 406.40 KG: 13.61 LOCATION: COUNTERTOP NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 ELECT NOTE: CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. EMERGENCY POWER	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		1015755 PHY08	1	MONITOR, PHYSIOLOGIC, TRANSPORT INTELLIVUE MULTI-MEASUREMENT TRANSPORT ACQUISITION MODULE; 5 HOUR BATTERY; CHARGES VIA DOCKING STATION, 120VAC TO INCLUDE: (WXDXH) IN: 9.80 x 3.80 x 4.40 LBS: 3.10 MM: 248.92 x 96.52 x 111.76 KG: 1.41 LOCATION: COUNTERTOP NOTES: 2025.09.08 SHD CONFIRMED KEEP ON MEQ LIST, UPDATED LOCATION FROM SPACE TO COUNTER MOUNTED CHANGE CODE: C CHANGE NOTES: LOCATION UPDATED DATA HARDWIRED: NO WIRELESS: YES VOLTS: 120 HERTZ: 60 AMPS: 1.3 ELECT NOTE: - REMOVABLE, RECHARGEABLE 5 HOUR BATTERY - 120 V AC-POWERED INTELLIVUE DOCK PROVIDES CHARGING AND NETWORK CONNECTIVITY - CAN BE ATTACHED TO A HOST MONITOR WHICH PROVIDES THE POWER TO THE X3	PHILIPS HEALTHCARE - MONITORING SYSTEMS 867030 INTELLIVUE X3	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
3		104161 MED03	1	DISPENSER, MEDICATION, AUTOMATED MAIN; 2 DRAWER; 14.75"LCD MONITOR; KEYBOARD;TOUCHPAD; BIO-ID (WXDXH) IN: 23.00 x 27.00 x 28.00 LBS: 100.70 MM: 584.20 x 685.80 x 711.20 KG: 45.68 LOCATION: COUNTERTOP ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: 802.11 B/G; BLUETOOTH VOLTS: 120 HERTZ: 60 AMPS: 3 WATS: 80 ELECT NOTE: 1A NOM;EMERGENCY POWER REQUIRED	BD CAREFUSION PYXIS MEDSTATION ES 323 2 DRAWER MAIN BTU: 409/HR	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		105843 REF04	1	REFRIGERATOR, UNDERCOUNTER, MEDICAL GRADE 5.3 CF;SOLID FIELD REVERSIBLE DOOR;TEMP RANGE 2-10 DEG C;7" I.C3 COLOR TOUCHSCREEN INTERFACE;AUTO DEFROST;R600A ISOBUTANE (WXDXH) IN: 25.00 x 29.00 x 34.00 LBS: 174.00 MM: 635.00 x 736.60 x 863.60 KG: 78.93 LOCATION: FLOOR NOTES: 2024-09 SHD CONFIRMED REMOT CONNECTIONS NOT REQUIRED ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CAHNGED FROM OF/OI TO OF/VI DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ45 ETHERNET VOLTS: 115 HERTZ: 60 AMPS: 1.4 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 8-10' POWER CORD;12V RECHARGEABLE LEAD ACID BATTERY FOR DISPLAY & ALARM BACKUP	HELMER LABS, INC. ILR105-GX I SERIES ADA BTU: 460 MECH NOTE: REAR ACCESS PORT FOR EXT MONITORING PROBES	OF/VI	LEASED	GROUP: 2 MAJOR MOVABLE
5		091485 ICE01	1	ICE MAKER W/WATER DISPENSER 12LB STORAGE;260LB PROD/24 HRS;SS EXTERIOR;EXTERNAL AIR FILTER;AGION ANTIMICROBIAL;R-134A REFRIGERANT;H2 NUGGET STYLE ICE (WXDXH) IN: 17.00 x 25.00 x 35.00 LBS: 183.00 MM: 431.80 x 635.00 x 889.00 KG: 83.01 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES VOLTS: 115 HERTZ: 60 AMPS: 7 ELECT NOTE: SINGLE PHASE;15A MAX FUSE SEE VENDOR TECHNICAL INFORMATION	SCOTSMAN ICE SYSTEMS HID312A-1 MERIDIAN AIR COOL BTU: 3200	OF/CI	NEW	GROUP: 1 FIXED
6		1021297	1	FILTER VALVE-IN-HEAD WATER FILTER SYSTEM. INSTALLS ON THE 3/4-INCH MAIN WATER LINE AT THE POINT OF ENTRY. FILTER HOUSING DESIGNED (WXDXH) IN: 5.00 x 5.00 x 13.25 LBS: 0.50 MM: 127.00 x 127.00 x 336.55 KG: 0.23 LOCATION: WALL CHANGE CODE: A CHANGE NOTES: ADDED PER SENECA, HGA HAS ALREADY ACCOUNTED FOR FI	CULLIGAN WATER SYSTEMS HF-360	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
						COLD WATER: YES PLUMB NOTE: SEE VENDOR SPECS		

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		105759 GLU01	1	LAB ANALYZER, GLUCOSE TOUCH SCREEN;QR BARCODE SCANNER;1,000 TEST STORAGE; 4,000 USER STORAGE; 500 TEST STORAGE;NUMERIC OR PASS/FAIL (WDXDH) IN: 3.10 x 5.80 x 1.18 LBS: 0.49 MM: 78.74 x 147.32 x 29.98 KG: 0.23 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: DATA REQUIRED: ETHERNET NETWORK DROP OR 802.11 A/B/G/N WIRELESS ACCESS. SEE CUTSHEET FOR FURTHER DETAILS. VOLTS: 120 HERTZ: 60 AMPS: .6 ELECT NOTE: STATSTRIP: 3.7V LI POLYMER RECHARGEABLE BATTERY ELECTRICAL INFO IS FOR DOCKING STATION	NOVA BIOMEDICAL 44993 STATSTRIP GLU 2	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
8		1023001 TMP02	2	THERMOMETER, ELECTRONIC, WALL MOUNTED TEMPORAL ARTERY PROFESSIONAL THERMOMETER; KEYLESS WALL MOUNT W/HIDDEN RELEASE LEVER; NO CABLE (WDXDH) IN: 3.23 x 0.00 x 6.20 LBS: 0.00 MM: 82.05 x 0.00 x 157.48 KG: 0.00 LOCATION: WALL NOTES: 2025.05.12 UPDATED MODEL FROM 5000 TO 5000S WITH NO CABLE KEYLESS WALL MOUNT PER SHD CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: 9-VOLT BATTERY	EXERGEN CORPORATION TAT-5000S 124275S/134305	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
9		083455 COF01	1	COFFEE BREWER, PLUMBED 4 CUP SIZES;HOT WATER SPOUT;LCD INTERFACE;FLIP UP DRIP (WDXDH) IN: 12.00 x 22.00 x 20.00 LBS: 35.00 MM: 304.80 x 558.80 x 508.00 KG: 15.88 LOCATION: COUNTERTOP SPEC COMMENTS: ;HEIGHT ACCOUNTS FOR AN OPEN LID DEPTH ACCOUNTS FOR FILTER NOTES: 08/30/2024 - UNIT AND FILTER TO BE INSTALLED BY SENECA AFTER SUBSTANTIAL. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ11 JACK VOLTS: 120 HERTZ: 60 AMPS: 11.6 WATS: 1400 PLUG TYPE: Type B (NEMA 5-15)	KEURIG, INC. K3000SE COMMERCIAL	OF/OI	NEW	GROUP: 1 FIXED
						COLD WATER: YES PLUMB NOTE: WATER PRESSURES: 40PSIG TO 125PSIG 1/4"FPT PORTS IN/OUT FILTER WATER CONNECTION-3/4" MALE GARDEN HOSE CONNECTION		

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10		1006012 COF04	1	COFFEE STATION COFFEE CUP & CONDIMENT ORGANIZER; BLACK; PLASTIC; CHROME HANDLES (WDXDH) IN: 5.40 x 11.00 x 12.60 LBS: 2.00 MM: 137.16 x 279.40 x 320.04 KG: 0.91 LOCATION: COUNTERTOP	MIND READER EMS-CAD01BLK CADDY ORGANIZER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
11		055770 OVM01	1	OVEN, MICROWAVE 1.0CF; 1000W; S.S. INTERIOR/EXTERIOR/TOUCH CONTROL; (WDXDH) IN: 20.50 x 18.03 x 12.17 LBS: 37.00 MM: 520.70 x 457.97 x 309.12 KG: 16.79 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: 14 WATS: 1000 PLUG TYPE: Type B (NEMA 5-15)	SHARP ELECTRONICS CORP. 074000616486 R-21LVF	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
12		1020123 MBOX01	1	BOX, RETURN, DRUGS LOCKABLE; WALL MOUNTED; BEIGE METAL ANTI-THEFT BAFFLE; BOTTOM IS (WDXDH) IN: 12.00 x 5.20 x 6.25 LBS: 9.00 MM: 304.80 x 132.08 x 158.75 KG: 4.09 LOCATION: WALL NOTES: 2024-09 SHD CONFIRMED WALL MOUNTED CHANGE CODE: C CHANGE NOTES: CHANGE FROM OF/OI COUNTER MNT TO OF/CI WALL MNT	HEALTH CARE LOGISTICS 3783 LOCKABLE RETURN DRUG BOX	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
13		1020550 DSPP04B	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL CONTROLLED SUBSTANCE WASTE MANAGEMENT SYSTEM. COMPLIANCE WITH REGULATORY GUIDELINES. REDUCES OPPORTUNITIES FOR DRUG (WDXDH) IN: 15.50 x 9.50 x 12.50 LBS: 17.00 MM: 393.70 x 241.30 x 317.50 KG: 7.72 LOCATION: SPACE SPEC COMMENTS: WALL/COUNTER BRACKET INSTALLED SEPARATELY BY CONTRACTOR CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY 0085-013-000 SMART SINK	OF/OI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
14		1020549 DSPP04A	1	BRACKET CACTUS SMART SINK MOUNTING BRACKET. TO INCLUDE: (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: WALL SPEC COMMENTS: BRACKET INSTALLED BY CONTRACTOR, SMART SINK SENT TO SITE FOR OWNER INSTALL CHANGE CODE: C CHANGE NOTES: SEPERATED DSPP04 INTO A AND B, CI VS OI	CACTUS, A STRYKER COMPANY CACTUS SMART SINK MOUNTING BRACKET 0085-000-003	OF/CI	NEW	GROUP: 1 FIXED
15		1003160 DSPP03	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL LOCKING WALL BRACKET;WHITE/BUE 3 GALLON PHARMASAFETY CONTAINER LEAK RESISTANT;TAMPER EVIDENT;SECURE HINGED LID;ABSORBENT PA (WXDXH) IN: 13.75 x 6.50 x 16.50 LBS: 5.00 MM: 349.25 x 165.10 x 419.10 KG: 2.27 LOCATION: WALL SPEC COMMENTS: DIMENSIONS OF BRACKET: 13.75" X 6.5" X 11.75" (W X D X H)	COVIDIEN 8528C/8836SA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WXDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
17		1004090	1	WASTE RECEPTACLE 13GAL;OPEN TOP;GRAY (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMMERCIAL PRODUCTS 2026695 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		1014420	1	WASTE RECEPTACLE 13GAL;OPEN TOP;BLUE (WXDXH) IN: 11.17 x 22.23 x 26.09 LBS: 6.73 MM: 283.72 x 564.65 x 662.69 KG: 3.06 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 2026699 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: NURSE STATION / NOUR / MEDS
ROOM NUMBER: 615

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
19		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
20		1003479 WRM03	1	WARMER, FLUID CAP. (16) 1L BAGS;OR 12L BOTTLES;RADIANT HEAT TECHNOLOGY;TEMP RANGE IRRIGATION: 90-150DEG F;TEMP RANGE INJECTION: 90-104DEG (WXDXH) IN: 18.50 x 23.50 x 22.00 LBS: 67.00 MM: 469.90 x 596.90 x 558.80 KG: 30.40 LOCATION: COUNTERTOP SPEC COMMENTS: ; ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: 2025.05.12 MOVED FROM 631 TO 615 PER SHD REVIEW 2 VOLTS: 120 HERTZ: 60 AMPS: 5 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: SINGLE PHASE; 9.1" IEC CORD INLET IN REAR; 6' LONG CORD	PEDIGO PRODUCTS P-2112 P-2112	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: LOCKER, MALE/FEMALE (QTY 2)
ROOM NUMBER: 616
ROOM COMMENT: RM 616, 618

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1003921 LHM01	2	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		089875	2	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE


END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: LOCKER, TOILET (QTY 2)
ROOM NUMBER: 617
ROOM COMMENT: RM 617, 619

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023055 CAB10*	2	CABINET, STORAGE MOBILE; STEEL; 2 ADJUSTABLE SHELVES; KEY LOCK; COLOR: SAND (WXDXH) IN: 24.00 x 36.00 x 48.00 LBS: 107.00 MM: 609.60 x 914.40 x 1219.20 KG: 48.54 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	TENNSCO ATM-320	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		004363	2	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: SCRUB
ROOM NUMBER: 621

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		094999 SNK02	1	SINK, SCRUB 2 STATION;SS;TIMER;MANUAL;KNEE OPERATED CONTROLS TO INCLUDE: (WDXDH) IN: 64.00 x 28.00 x 52.00 LBS: 280.00 MM: 1625.60 x 711.20 x 1320.80 KG: 127.01 LOCATION: WALL SPEC COMMENTS: ; <i>NOTES: 2024.09.24 SEPARATED CHAIR CARRIER FROM SINK, ORIGINALLY BOTH VENDOR INSTALL</i> ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: SEPERATED CHAIR CARRIER TO OF/CI ELECT NOTE: FOR TIMER 120 V, 60 HZ, 1 AMP, GROUNDED	STERIS CORPORATION CE12-230-1/ACC FLEXMATIC	OF/VI	NEW	GROUP: 1 FIXED
2		046964 SNK02 B*	1	SINK,SCRUB,ACCESSORY,CARRIER,CHAIR CHAIR CARRIER FOR DOUBLE SINK (WDXDH) IN: 68.25 x 4.00 x 54.00 LBS: 48.00 MM: 1733.55 x 101.60 x 1371.60 KG: 21.78 LOCATION: WALL SPEC COMMENTS: REFERENCE VENDORS INSTALLATION DRAWING <i>NOTES: 2024.09.24 SEPARATED CAHIR CARRIER TO ENSURE CONTRACTOR INSTALL, ORIGINALLY PART OF SNK02 INDICATING VENDOR INSTALL</i> CHANGE CODE: C CHANGE NOTES: SPERATED FROM SNK02, CHANGED TO OF/CI	STERIS CORPORATION CE00-006 DOUBLE	OF/CI	NEW	GROUP: 1 FIXED
3		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXDH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 04
ROOM NUMBER: 622

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		107260 SCP03	1	SCOPE, LARYNGOSCOPE 7" SCREEN SIZE;FOR DIFFICULT INTUBATION. (WXDXH) IN: 23.00 x 23.00 x 55.00 LBS: 24.00 MM: 584.20 x 584.20 x 1397.00 KG: 10.89 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY UNIT WITH ACCESSORIES. MAKE/MODEL MAY CHANGE DURING PROCUREMENT DATA NOTE: HDMI;USB 2.0;VIDEO OUTPUT TO EXTERNAL MONITOR VOLTS: 120 HERTZ: 60 AMPS: .5 ELECT NOTE: POWER CORD ADAPTER;RECHARGEABLE LI-ION BATTERIES	KARL STORZ ENDOSCOPY-AMERICA, INC. C-MAC/OPTIONS 8403ZXK	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
2		056002 CRC04	1	CART, RESUSCITATION, CARDIAC CART, RESUSCITATION, CARDIAC; 6 DRAWERS;4-3", 1-6",1-12",PVC BUMPER;LOCKING BREAKAWAY BAR, TWIN (WXDXH) IN: 43.00 x 25.00 x 47.00 LBS: 165.00 MM: 1092.20 x 635.00 x 1193.80 KG: 74.85 LOCATION: FLOOR SPEC COMMENTS: OPTIONS MAY ADD TO THE OVERALL DIMENSIONS BACKBOARD HOLDER, GLOVE BOX, SHARPS CONTAINER, O2 TANK HOLDER, SHELF FOR PORTABLE SUCTION PUMP; IV POLE;	ARMSTRONG MEDICAL INDUSTRIES AR-6 PREMIER W/ER PK	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		082533 DEF01	1	DEFIBRILLATOR R SERIES ALS DEFIBRILLATOR;AED W/MANUAL OVERRIDE;GUIDELINES 2010 COMPATIBLE;CODE READINESS TESTING SYSTEM;RECTILINEAR (WXDXH) IN: 10.50 x 12.50 x 8.20 LBS: 15.20 MM: 266.70 x 317.50 x 208.28 KG: 6.90 LOCATION: COUNTERTOP DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: OPTIONAL WIFI: 802.11B USB PORT 2.0 COMPATIBLE VOLTS: 120 HERTZ: 60 ELECT NOTE: SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER EMERGENCY POWER	ZOLL MEDICAL CORPORATION 30720005201310012 R SERIES ALS	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
4		094808 ASP01	1	PUMP, SUCTION/ASPIRATOR, GENERAL, PORTABLE PORTABLE; RECHARGEABLE BATTERY; LED DISPLAY; HEPA FILTER; BEMIS DISPOSABLE CANISTER SYSTEM; HIGH FLOW; HIGH VACUUM; (WXDXH) IN: 13.00 x 6.30 x 12.40 LBS: 8.90 MM: 330.20 x 160.02 x 314.96 KG: 4.04 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 ELECT NOTE: ELECTRICAL SHOWN FOR BATTERY CHARGER	LAERDAL MEDICAL CORP. LSU 4000 78002001	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 03
ROOM NUMBER: 623

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WXDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON- ELECTRIC, STANDARD MATTRESS PER SHD 2025.07.22 STRETCHER MAKE/MODEL TBC CHANGE CODE: C CHANGE NOTES: CHANGED MODEL	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 02
ROOM NUMBER: 624

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015397 CRL01	1	CART, LINEN 5" QUIET, NON-MARKING TENTE CASTERS AMONG THE BEST ON THE MARKET; REINFORCED INJECTION-MOLDED SHELVES WITH ½" LIP TO (WXDXH) IN: 40.00 x 20.00 x 63.00 LBS: 50.00 MM: 1016.00 x 508.00 x 1600.20 KG: 22.68 LOCATION: FLOOR	DIRECT SUPPLY, INC LC36-4-42 IPU PREMIUM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1002258 CMH01	1	CART, MALIGNANT HYPERTHERMIA TWO 6" DRAWERS AND 1 COMPARTMENT WITH PRE-INSTALLED FOLLETT REF 1 REFRIGERATOR; KEY LOCK FOR DRAWERS AND REFRIGERATOR; (WXDXH) IN: 37.50 x 26.00 x 45.50 LBS: 175.00 MM: 952.50 x 660.40 x 1155.70 KG: 79.38 LOCATION: FLOOR NOTES: 2025.05.12 UPDATED FROM EXISTING TO NEWSHD REVIEW 2 DOC CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC VOLTS: 115 HERTZ: 60 AMPS: 1.8 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: SJT 18 AWG 8' POWER CORD	HARLOFF COMPANY, INC. MH5216K MH5216K	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
3		008082 REF02	1	REFRIGERATOR, UNDERCOUNTER 115V; 3.3 CF; BLACK; AUTO DEFROST; ADA; 2.5 GLASS SHELVES; ACCEPTS PANEL; (WXDXH) IN: 20.88 x 23.00 x 28.50 LBS: 80.00 MM: 530.36 x 584.20 x 723.90 KG: 36.29 LOCATION: FLOOR SPEC COMMENTS: NO ADDITIONAL CLEARANCE AROUND SIDES, TOP OR REAR OF UNIT IS NEEDED FOR VENTILATION. NOTES: SITS ON HYPERTHERMIA CART VOLTS: 115 HERTZ: 60 AMPS: 1.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A	U-LINE CORPORATION 29R U-29RB-00A	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 02
ROOM NUMBER: 624

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1023182 RCK08*	2	RACK CUSTOM WALL MOUNTED 18U INTEGRATION RACKS, INTEGRATION PHONE HOLDER ACCESSORIES (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: WALL <i>NOTES: 2025.08.01 2 WALL MOUNTED INTEGRATION RACKS FOR STRYKER AV EQUIPMENT ADDED PER SHD</i> CHANGE CODE: A CHANGE NOTES: ADDED AV RACKS PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS OR RENOVATION	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
					SELLER: STRYKER COMMUNICATIONS			

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: HOUSEKEEPING
ROOM NUMBER: 625

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078704 RCK04	1	RACK, MOP/BROOM SS;SHELF W/3 MOP OR BROOM HOLDERS;4 HOOKS; TYPE-304 STAINLESS;ALL WELDED CONSTRUCTION;SATIN FINISH (WDXDH) IN: 34.00 x 8.00 x 13.00 LBS: 10.00 MM: 863.60 x 203.20 x 330.20 KG: 4.54 LOCATION: WALL	BOBRICK WASHROOM EQUIPMENT, INC. B-239	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXDH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		093627 VAC04	1	VACUUM, WET/DRY 19GAL TANK;FRONT MOUNT SQUEEGEE;ERGO TIP-&-POUR DESIGN;9GAL DUST BAG (WDXDH) IN: 24.00 x 26.00 x 41.00 LBS: 61.00 MM: 609.60 x 660.40 x 1041.40 KG: 27.67 LOCATION: FLOOR CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 WATS: 1100 ELECT NOTE: 50' POWER CORD	NILFISK ADVANCE 56383408 VL500-75	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 01
ROOM NUMBER: 626

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1002328 CRM02	1	IMAGING, TABLE, C-ARM MOTORIZED 5-MOVEMENTS INCLUDING 12" ELEVATION (FROM 26" TO 38"), TRENDLENBURG TILT, LATERAL TILT, LONGITUDINAL TRAVEL (WXDXH) IN: 22.00 x 82.00 x 38.00 LBS: 510.00 MM: 558.80 x 2082.80 x 965.20 KG: 231.34 LOCATION: FLOOR VOLTS: 120 HERTZ: 60 AMPS: 9 ELECT NOTE: 1 PHASE	MEDSTONE TABLES 0047-TM5(H)SCU	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
2		098605 RCK01	1	RACK, APRON, WALL MOUNTED 8 PEGS;4 APRONS;ELECTRO-MAGNETIC POWDER COATING;STAINLESS STEEL (WXDXH) IN: 19.00 x 5.00 x 3.00 LBS: 7.20 MM: 482.60 x 127.00 x 76.20 KG: 3.27 LOCATION: WALL	PHILLIPS SAFETY PRODUCTS RAR-P4	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
3		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4		106731 CRM01	1	IMAGING, X-RAY, C-ARM COMPACT;ALL-IN-ONE;ONE TOUCH CONTROL PANEL;DOSE CONTROL;EXTENDED REACH;STAND-BY POWER (WXDXH) IN: 30.70 x 89.80 x 88.60 LBS: 706.00 MM: 779.78 x 2280.92 x 2250.44 KG: 320.24 LOCATION: FLOOR SPEC COMMENTS: ; NOTES: *OEC - PREMIER (CONTRACT# PP-IM-268) POWER: TYPE B NEMA 5-20 RECEPTACLE NEEDED DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: USB 2.0; DISPLAY PORT; ETHERNET; WLAN 802.11 A/B/G/N; SUPPORTS WEP ENCRYPTION VOLTS: 120 HERTZ: 60 AMPS: 20	GE OEC MEDICAL SYSTEMS ONE CFD OEC ONE CFD	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 01
ROOM NUMBER: 626

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		1023144	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; PARTIAL OVERLAP DESIGN; FRONT HOOK-AND-LOOP CLOSURE W/ADJUSTABLE 2" BUCKLE CLOSUR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-KILT ECKILT-SB-A-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
6		1023142	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; FRONT AND BACK PROTECTION; FRONT BUCKLE CLOSURE; SIZE: MALE EXTRA LARGE (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-REGULAR VEST ECVEST-SB-XLM-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		1023140	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; FRONT AND BACK PROTECTION; FRONT BUCKLE CLOSURE; SIZE: MALE MEDIUM (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-REGULAR VEST ECVEST-SB-B-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 01
ROOM NUMBER: 626

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1023147	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; PARTIAL OVERLAP DESIGN; FRONT HOOK-AND-LOOP CLOSURE W/ADJUSTABLE 2" BUCKLE CLOSUR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-KILT ECKILT-SB-C-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		1023149	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; PARTIAL OVERLAP DESIGN; FRONT HOOK-AND-LOOP CLOSURE W/ADJUSTABLE 2" BUCKLE CLOSUR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-KILT ECKILT-SB-XLM-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
10		1023138	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; FRONT AND BACK PROTECTION; FRONT BUCKLE CLOSURE; SIZE: MALE SMALL (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFNIRMED CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-REGULAR VEST ECVEST-SB-A-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR EQUIP ALCOVE 01
ROOM NUMBER: 626

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		1023141	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; FRONT AND BACK PROTECTION; FRONT BUCKLE CLOSURE; SIZE: MALE LARGE (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE <i>NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFIRMED</i> CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-REGULAR VEST ECVEST-SB-C-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
12		1023145	1	APRON, LEAD FREE XENOLITE 600-B-0.35MM PROTECTIVE MATERIAL; PARTIAL OVERLAP DESIGN; FRONT HOOK-AND-LOOP CLOSURE W/ADJUSTABLE 2" BUCKLE CLOSUR (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE <i>NOTES: 2025.07.31 ADDED TO LIST PER SHD, MAKES/MODELS COFIRMED</i> CHANGE CODE: C CHANGE NOTES: 2025-05-12 INCREASED QTY PER SHD REVIEW 2 DOC	BURLINGTON MEDICAL SUPPLIES BASICS-KILT ECKILT-SB-B-043	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO EQUIP ALCOVE 01
ROOM NUMBER: 627

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		058850 CRC01	1	CART, CAST STAINLESS;8 DRAWER;S.S. FOLDING SIDE SHELF; DRAWERS: (1) 3"W/BLACK VINYL MAT, (5) 3"W/ 4 ADJ DIVIDERS PER DRAWER, (2) (WDXH) IN: 38.00 x 21.00 x 77.00 LBS: 202.00 MM: 965.20 x 533.40 x 1955.80 KG: 91.63 LOCATION: FLOOR	HARLOFF COMPANY, INC. DELUXE 6035	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO EQUIP ALCOVE 02
ROOM NUMBER: 628

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: SOILED UTILITY
ROOM NUMBER: 630

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		100858 TRK01	1	TRUCK/BIN, LINEN 500LB; 16 CUFT CAPACITY; MANUAL UNLOADING; BLUE (WDXH) IN: 30.38 x 44.00 x 35.50 LBS: 60.50 MM: 771.66 x 1117.60 x 901.70 KG: 27.45 LOCATION: FLOOR NOTES: INCLUDES LID	TOTER INC. MMC16-90701 CUBE HEAVY DUTY	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
2		1015414 TRK02	1	TRUCK, TILT, WASTE/LINEN 16 CUFT MANUAL CUBE TRUCK; INCLUDES LID (WDXH) IN: 30.40 x 44.00 x 35.50 LBS: 60.00 MM: 772.16 x 1117.60 x 901.70 KG: 27.22 LOCATION: FLOOR NOTES: BLACK IN COLOR; INCLUDES LID	TOTER INC. MMC16-10836 BLACK	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
3		100859 WST02	2	CONTAINER, BIOHAZARD RED; 64 GALLON; 2 WHEEL; HINGED LID; 2-10"WHEELS; 2-4"CASTERS; PLASTIC (WDXH) IN: 24.25 x 32.00 x 44.25 LBS: 45.30 MM: 615.95 x 812.80 x 1123.95 KG: 20.55 LOCATION: FLOOR	TOTER INC. RMC64-00RED RMC64	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		1008170 CRK01	1	CART, CYLINDER, D&E, MULTI 6 CYLINDER CAP; CHROME PLATED STEEL; STANDARD 5" (WDXH) IN: 11.00 x 16.00 x 39.00 LBS: 27.00 MM: 279.40 x 406.40 x 990.60 KG: 12.25 LOCATION: FLOOR	WT FARLEY WTFRCDC06E CR-DC06E	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
5		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: SOILED UTILITY
ROOM NUMBER: 630

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		004363		1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ANESTHESIA WORKROOM
ROOM NUMBER: 631

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015782 MTCR02	1	MONITOR, CENTRAL STATION, REMOTE MIRROR STATION; 23" DISPLAY; 120V AC (WXDXH) IN: 17.00 x 8.20 x 16.00 LBS: 30.00 MM: 431.80 x 208.28 x 406.40 KG: 13.61 LOCATION: COUNTERTOP SPEC COMMENTS: OVERALL DIMENSIONS INCLUDE CPU, UPS, AND RECORDER/ALARM SPEAKER MODULE SEE CUT SHEET FOR SPECIFIC DIMENSIONS. NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: MINIMUM ETHERNET 10/100 T, CAT 5E OR HIGHER VOLTS: 115 HERTZ: 60 AMPS: 10.8 DEDICATED CIRCUIT: YES ELECT NOTE: REQUIRES A 15AMP QUAD OUTLET WITHIN 3' OF UPS/CPU.	PHILIPS MEDICAL SYSTEMS MXU0295 DISPLAY: 23 BTU: 1595 BTU/HR	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
2		078823 SHW02	2	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WXDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		1008170 CRK01	1	CART, CYLINDER, D&E, MULTI 6 CYLINDER CAP; CHROME PLATED STEEL; STANDARD 5"	WT FARLEY WTFCD06E CR-DC06E	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WXDXH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - NULL	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ANESTHESIA WORKROOM
ROOM NUMBER: 631

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
5		092852		1	WASTE RECEPTACLE, RECYCLING 16 GALLON;RECYCLING;PLASTIC CONSTRUCTION; VENTED CHANNELS; HANDLES AT RIM (WDXH) IN: 22.00 x 11.00 x 25.00 LBS: 6.42 MM: 558.80 x 279.40 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971257 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023173 MGAA01	1	BOOM, ANESTHESIA MOUNTING PLATE: SINGLE COMMON PLATE; FRICTION BRAKE SYSTEM; CEILING COVER: STANDALONE MFR CONTROLS: REAR ONLY;FRONT MFR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING ANCHOR DETAIL REQ: YES DATA NOTE: LOW VOLTAGE SERVICES QUAD RJ-45 CAT 6: PLATE 1 1 EA. BLANK PLATE COR IP 4-IN-1 PLATE: 1 PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: 20A/125V DUPLEX (4 OUTLETS): 2; 20A/125V DUPLEX (6 OUTLETS): 1 PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL S-SERIES	OF/VI	NEW	GROUP: 1 FIXED
2		1023171	1	LIGHT, SURGICAL TC ONLY MOUNT;SURGICAL HIGH DEFINITION MODE;MICROSPOT AND SHD MODE;RANGE OF SPOT SIZES;NO CARDANIC CONTROL;TOUCH CONTROL UNI (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL OCULAN	OF/VI	NEW	GROUP: 1 FIXED
3		1023172	1	LIGHT, SURGICAL TC ONLY MOUNT;SURGICAL HIGH DEFINITION MODE;MICROSPOT AND SHD MODE;RANGE OF SPOT SIZES;NO CARDANIC CONTROL;TOUCH CONTROL (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL OCULAN	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1017713	2	MONITOR, VIDEO 32" 4K OLED SURGICAL DISPLAY (WXDXH) IN: 28.70 x 2.10 x 18.30 LBS: 19.25 MM: 728.98 x 53.34 x 464.82 KG: 8.74 LOCATION: SPACE DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 2.5 WATS: 115.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: HOSPITAL GRADE AC POWER CORD PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS 0240031300 32 INCH 4K OLED SYRGUCAK DISPLAY	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
5		1013822 MON01	1	MONITOR, VIDEO 55" LCD 4K DISPLAY (WXDXH) IN: 48.80 x 2.40 x 28.10 LBS: 60.80 MM: 1239.52 x 60.96 x 713.74 KG: 27.58 LOCATION: SPACE DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ-45 10 BASE-T/100 BASE-TX X 2 REFERENCE VENDOR'S TECHNICAL DATA VOLTS: 120 HERTZ: 60 AMPS: 3.3 WATS: 295 ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS P45952 P45952	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
6		1021686 MON01A	1	BRACKET, TELEVISION, WALL UNIVERSAL; DEPTH 2.73-11.18"; BLACK (WXDXH) IN: 24.38 x 11.18 x 16.78 LBS: 30.40 MM: 619.26 x 283.98 x 426.22 KG: 13.79 LOCATION: WALL NOTES: THE WALL LCD BRACKETS SHIP WITH THE CAPITAL EQUIPMENT. THE GC TYPICALLY INSTALLS THESE JUST BEFORE THE EQUIPMENT INSTALL OR EVEN THE SAME WEEK AS THE EQUIPMENT INSTALLATION. IF THE GC WANTS BEFOREHAND, CONFIRM WITH STRYKER ALTERNATE ARRANGEMENTS CHANGE CODE: C CHANGE NOTES: ON SEPERATE LINE AS OFCI PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	PEERLESS SALES COMPANY SP850-UNL SMARTMOUNT	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
7		1023174 BEQ01	1	BOOM, EQUIPMENT MOUNTING PLATE;SINGLE COMMON PLATE; SHELF RAIL TYPE: FAIRFIELD;3 SHELVES (750MM EACH);SHELF 2 W/CONTROLS;ELECTRIC BRAKE (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING ANCHOR DETAIL REQ: YES DATA NOTE: 1 EA: DATA PASS THRU PLATE 1 EA: DISTRIBUTION BD PLATE 7 EA. BLANK PLATE 2 EA. COR-IP 4-IN-1 PLATE PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: 20A/125V DUPLEX (6 OUTLETS): 4 PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL S-SERIES	OF/VI	NEW	GROUP: 1 FIXED
8		1023176 AUD01	1	AUDIO/VISUAL SYSTEM,INTEGRATED, SURGERY 4K AND HD VIDEO ROUTING; PC HD HDMI/DVI INPUT; 4K DISPLAYS (X2) WITH OPTIONAL 55IN 4K DISPLAY; 4-IN-1 BOOM PLATE (X2); (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA DATA NOTE: REFERENCE VENDOR'S TECHNICAL DATA WIRED; LAN; HDMI/DVI; ETHERNET; AUDIO SYSTEM ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL COR LITE	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		064557 ESU03	1	ELECTROSURGICAL UNIT CUTTING;COAGULATION;VOLTAGE REGULATION;BIPOLAR;ENDOCUT;AUTOMATIC POWER DOSING;AUTO START/STOP;9 STORABLE PROGRAMS TO (WXDXH) IN: 16.10 x 15.00 x 6.50 LBS: 20.00 MM: 408.94 x 381.00 x 165.10 KG: 9.08 LOCATION: COUNTERTOP NOTES: LOCATED ON BOOM CHANGE CODE: C CHANGE NOTES: 2025.05.12 MODEL AND STATUS CHANGE PER SHD REVIEW VOLTS: 120 HERTZ: 60 AMPS: 8 ELECT NOTE: 40W POWER INPUT DURING STAND-BY 920VA POWER INPUT DURING MAX. ELECTROSURGICAL OUTPUT	ERBE USA, INC VIO 200S 10140-400/ACC	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
10		067138 ENDOV01*	1	SCOPE ACCESSORY, PROCESSOR, ENDOSCOPIC HDTV SIGNAL OUTPUT;SDTV SIGNAL OUTPUT;AUTOMATIC WHITE BALANCE ADJUSTMENT; STANDARD COLOR CHART OUTPUT;AUTOMATIC GAIN (WXDXH) IN: 15.04 x 19.26 x 3.59 LBS: 23.59 MM: 382.02 x 489.21 x 91.19 KG: 10.71 LOCATION: SPACE NOTES: 2025.05.12 ADDED EXISTING UNIT PER SHD REVIEW 2 DOC. LIVES ON EQUIP BOOM CHANGE CODE: A CHANGE NOTES: ADDED DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: SEE CUT SHEET FOR DETAILS VOLTS: 120 HERTZ: 60 WATS: 150	OLYMPUS AMERICA INC. CV-190 EVIS EXERA III	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
11		067135 PMP01*	1	PUMP, IRRIGATION FLUSHING PUMP;USE W/OLYMPUS SCOPES; AUTOCLAVABLE 2L WATER CONTAINER WITH LID (WXDXH) IN: 8.00 x 16.00 x 7.00 LBS: 9.00 MM: 203.20 x 406.40 x 177.80 KG: 4.09 LOCATION: COUNTERTOP NOTES: 2025.05.12 ADDED EXISTING UNIT PER SHD REVIEW 2 DOC. LIVES ON EQUIP BOOM CHANGE CODE: A CHANGE NOTES: ADDED VOLTS: 120 HERTZ: 60 WATS: 100	OLYMPUS AMERICA INC. OFP-2 OFP-2	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
12		081027	1	TESTER LEAKAGE TESTER EVIS TO INCLUDE: (WXDXH) IN: 4.00 x 7.00 x 7.00 LBS: 5.00 MM: 101.60 x 177.80 x 177.80 KG: 2.27 LOCATION: COUNTERTOP NOTES: 2025.05.12 ADDED EXISTING UNIT PER SHD REVIEW 2 DOC. LIVES ON EQUIP BOOM CHANGE CODE: A CHANGE NOTES: ADDED VOLTS: 120 HERTZ: 60 AMPS: .2 ELECT NOTE: SINGLE PHASE	OLYMPUS AMERICA INC. MU-1/MB-155	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
13		096403 ANES01	1	ANESTHESIA MACHINE FABIUS GS PREMIUM; 3 GAS;VAPORIZER TO INCLUDE: (WXDXH) IN: 41.00 x 33.00 x 52.00 LBS: 296.00 MM: 1041.40 x 838.20 x 1320.80 KG: 134.27 LOCATION: FLOOR SPEC COMMENTS: VENDOR SET-UP REQUIRED; CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: (1) RS232 VOLTS: 120 HERTZ: 60 AMPS: 2.3 WATS: 70 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: NEMA 5-15R HOSPITAL GRADE	DRAGER MEDICAL, INC. 8607000/ACC FABIUS GS PREM	OF/VI	EXISTING	GROUP: 2 MAJOR MOVABLE
14		1015753 PHY07	1	MONITOR, PHYSIOLOGICAL W/ANESTHESIA MOUNT, 15" LCD TFT TOUCH DISPLAY; 6-WAVES; SIDE MOUNTS TO ANESTHESIA MACHINE; 120 VAC TO INCLUDE: (WXDXH) IN: 15.90 x 7.20 x 12.10 LBS: 22.30 MM: 403.86 x 182.88 x 307.34 KG: 10.12 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70	PHILIPS MEDICAL SYSTEMS 866066_NAM MX550	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
15		1004828 GAS01	1	ANESTHESIA MACHINE ACCESSORY INTELLIVUE ANESTHESIA GAS MODULE; GAS ANALYZER MODULE; PROVIDES A NON-DISPERSIVE INFRARED MEASUREMENT OF RESPIRATORY AND (WDXH) IN: 3.00 x 6.75 x 4.25 LBS: 3.00 MM: 76.20 x 171.45 x 107.95 KG: 1.37 LOCATION: SPACE NOTES: 2025-0908 CONFIRMED KEEP FOR ETCO2	PHILIPS MEDICAL SYSTEMS 866173 G7 MODULE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
16		1023694 STR07*	1	STRETCHER PRIME SERIES FIFTH WHEEL STRETCHER. 30" LITTER. MANUAL. 700 LB WEIGHTT CAPACITY (WDXH) IN: 38.00 x 85.00 x 34.00 LBS: 0.00 MM: 965.20 x 2159.00 x 863.60 KG: 0.00 LOCATION: FLOOR NOTES: 2025.09.18 STRYKER PRIME 5TH WHEEL, WITH SCALE, NON- ELECTRIC, STANDARD MATTRESS PER SHD 2025.07.22 STRETCHER MAKE/MODEL TBC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	STRYKER CORP/MEDICAL DIV 1105 PRIME FIFTH WHEEL	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
17		1023017 STL04*	2	STOOL, REVOLVING 360° SWIVEL; ARMREST; HEIGHT ADJUSTABLE; FOOTREST (WDXH) IN: 18.00 x 18.00 x 38.60 LBS: 40.00 MM: 457.20 x 457.20 x 980.44 KG: 18.15 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	OUBO FURNITURE (AMAZON) B0C2D253PB	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
18		089895 STL01	1	STOOL, REVOLVING ERGO;W/BACK;PNEUMATIC HGT ADJ;2" CASTERS 18X18 1/2" CONTOURED SEAT CUSHION W/ADJUSTMENT; (WDXH) IN: 25.00 x 25.00 x 50.00 LBS: 38.00 MM: 635.00 x 635.00 x 1270.00 KG: 17.24 LOCATION: FLOOR CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	PEDIGO PRODUCTS, INC. T-584 ANESTHESIA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
19		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY;SINGLE STEP TITRATION LIMIT;MULTIPLE MEDS ERROR PREVENTION (WXDXH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
20		080356 SCT01	1	STAND, SUCTION CANISTER, MOBILE COLLECTION CANISTER CARRIER; 4-CANISTER: 36" HIGH;ROLLING STAND (WXDXH) IN: 24.00 x 24.00 x 36.00 LBS: 10.00 MM: 609.60 x 609.60 x 914.40 KG: 4.54 LOCATION: FLOOR	ALLIED HEALTHCARE PRODUCTS, INC. 20-02-0023 4-CANISTER MODEL	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
21		051630 TBW02	1	TABLE, WORK S.S.;WITH UNDER-SHELF;MOBILE TO INCLUDE: (WXDXH) IN: 36.00 x 24.00 x 34.00 LBS: 58.00 MM: 914.40 x 609.60 x 863.60 KG: 26.31 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. SG-98-SS/ACC -	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
22		1006176 CRS03	1	CART, SUPPLY 3 3"-, 2 6"-, AND 1 9"-HIGH DRAWER CONFIGURATION;ACCOMMODATES A VARIETY OF INTERIOR AND EXTERIOR ACCESSORIES;LIGHTWEIGHT (WXDXH) IN: 29.25 x 21.25 x 42.50 LBS: 145.00 MM: 742.95 x 539.75 x 1079.50 KG: 65.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	INNERSPACE BY SOLAIRE SP30W6 PACE 30	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
23		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WDXH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
24		093116	3	REGULATOR, SUCTION, CONTINUOUS TWO-MODE W/DIAMOND & LOCKING GLAND OVERFLOW SAFETY TRAP (WDXH) IN: 3.00 x 5.00 x 7.00 LBS: 1.10 MM: 76.20 x 127.00 x 177.80 KG: 0.50 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: 2-3AA BATTERIES	OHIO MEDICAL CORPORATION 8701-1225-905	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
25		1004558	3	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
26		1004559	3	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
27		084252	2	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
28		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
29		1003160 DSPP03	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL LOCKING WALL BRACKET; WHITE/BUE 3 GALLON PHARMASAFETY CONTAINER LEAK RESISTANT; TAMPER EVIDENT; SECURE HINGED LID; ABSORBENT PA (WDXH) IN: 13.75 x 6.50 x 16.50 LBS: 5.00 MM: 349.25 x 165.10 x 419.10 KG: 2.27 LOCATION: WALL SPEC COMMENTS: DIMENSIONS OF BRACKET: 13.75" X 6.5" X 11.75" (W X D X H)	COVIDIEN 8528C/8836SA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
30		006656 MAY01	2	STAND, MAYO FOOT OPERATED; LARGE TRAY; STAINLESS STEEL; 2" CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY; REINFORCED; STAINLESS STEEL TRAY (WDXH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ; HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
31		021273 KCB01	2	KICKBUCKET SS; REMOVABLE 12 QT. SS PAIL; TIP-RESISTANT; ROUND (WDXH) IN: 13.75 x 13.75 x 13.25 LBS: 8.00 MM: 349.25 x 349.25 x 336.55 KG: 3.63 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. P-1020-SS P-1020-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
32		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
33		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
34		1005731 CLK01	1	CLOCK, ELAPSED TIME SURFACE MOUNT 6 DIGIT; RED; 4' DIGIT HEIGHT; 4" SLOPE (WXDXH) IN: 23.50 x 2.40 x 8.00 LBS: 4.60 MM: 596.90 x 60.96 x 203.20 KG: 2.09 LOCATION: WALL NOTES: 2025.06.10 CHANGED FROM POE TO 120V PER ELECTRICAL FUTURE ITEM; ROUGH IN ONLY AT THIS TIME CHANGE CODE: C CHANGE NOTES: CHANGED FROM POE TO 120V VOLTS: 120 HERTZ: 60 ELECT NOTE: 10' CORD WITH PLUG	PRIMEX XRA7A203E LEVO SERIES	OF/CI	FUTURE	GROUP: 1 FIXED
35		099447 STL03	1	STOOL, FOOT, WITH HANDRAIL TWO STEP; STEPS ARE 8" AND 16" HIGH; CHROME CORRUGATED RUBBER SURFACE; GREY FLOOR TIPS (WXDXH) IN: 16.00 x 12.00 x 47.00 LBS: 35.00 MM: 406.40 x 304.80 x 1193.80 KG: 15.88 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC	PEDIGO PRODUCTS, INC. P-18-A-K P-18-A-K	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
36		1023071 CAB09*	1	CABINET FULL GLAZING DOOR; PRIMED STEEL; SEMI-RECESSED 3" ROUND EDGE (WXDXH) IN: 13.88 x 6.00 x 27.38 LBS: 25.00 MM: 352.56 x 152.40 x 695.46 KG: 11.34 LOCATION: WALL SPEC COMMENTS: ROUGH OPENING: 11.5" X 3.375" X 25" NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC	JL INDUSTRIES 1017G10	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: ENDO PROCEDURE ROOM
ROOM NUMBER: 632

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
37		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: DECONTAMINATION
ROOM NUMBER: 633

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015523 SNK01	1	SINK, REPROCESSING REPROCESSING SINK 120" TRIPLE BAY;3 SINK;STAINLESS;30" BAYS;AIR & WATER PISTOLS;HEIGHT ADJUSTABLE 33-41" WORK SURFACE; 2 (WXDXH) IN: 120.00 x 28.00 x 53.00 LBS: 490.00 MM: 3048.00 x 711.20 x 1346.20 KG: 222.27 LOCATION: FLOOR NOTES: 2024.09.24 REQUIRES RO/DI, HGA CONFIRMED BUILDING RO PROVIDED ANCHOR DETAIL REQ: YES DATA NOTE: REFERENCE VENDOR'S TECHNICAL DATA VOLTS: 120 HERTZ: 60 AMPS: 15 ELECT NOTE: GFCI PROTECTED OUTLET SINGLE PHASE SEE VENDOR INFO FOR DETAILS	STERIS CORPORATION SINK50120 AMSCO 53	OF/VI	NEW	GROUP: 1 FIXED
2		1015522 CLR01	1	CLEANER, ULTRASONIC ULTRASONIC CLEANER. FEATURES: 20 GALLON. 16 LUMEN CAPACITY. AUTO-TUNING FREQUENCIES WHICH ENSURED BY CONSTANT POWER OUTPUT (WXDXH) IN: 34.80 x 25.80 x 58.90 LBS: 461.00 MM: 883.92 x 655.32 x 1496.06 KG: 209.11 LOCATION: FLOOR SPEC COMMENTS: HEIGHT REFLECTS WITH LID OPEN. CLOSED: 42.5 IN. MAX OPERATING WEIGHT: 430LBS. ANCHOR DETAIL REQ: YES DATA NOTE: USB FOR DOWNLOADING VOLTS: 208 HERTZ: 60 AMPS: 15 WATS: 3000 DEDICATED CIRCUIT: YES PLUG TYPE: Special ELECT NOTE: 15 AMP/PH DEDICATED CIRCUIT REQUIRED;3PH	STERIS CORPORATION IW523546NPR INNOWAVE UNITY ULTRASONIC IRRIGATOR (20 GAL) BTU: 700/HR	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

COLD WATER: YES HOT WATER: YES
TREATED WATER: YES MEDICAL GAS: YES
DRAIN REQUIRED: YES
PLUMB NOTE: TREATED WATER REQUIRED FOR FINAL RINSE CYCLE

COLD WATER: YES HOT WATER: YES
DRAIN REQUIRED: YES
PLUMB NOTE: REFERENCE VENDOR'S TECHNICAL DATA FOR DETAILS

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: DECONTAMINATION
ROOM NUMBER: 633

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1015533 WSHD01	1	WASHER/DISINFECTOR ELECTRIC;480V;DD;SINGLE CHAMBER;VERTICAL SLIDING POWER DOOR;TOUCH SCREEN CONTROL;USB CYCLE TAPE DOWNLOAD;DRAIN DISCHARGE (WDXDH) IN: 42.00 x 31.00 x 80.00 LBS: 457.00 MM: 1066.80 x 787.40 x 2032.00 KG: 207.30 LOCATION: FLOOR NOTES: 2025-0708 CONFIRMED INCLUDE VENTING PER MEP 2025-0924 CONFIRMED BUILDING RO/DI ANCHOR DETAIL REQ: YES VOLTS: 480 HERTZ: 60 ELECT NOTE: REFER TO VENDOR TECHNICAL INFORMATION	STERIS CORPORATION FH14043 AMSCO 5052	OF/VI	NEW	GROUP: 1 FIXED
4	No image available	098041 WRCK01*	1	RACK, WALL MOUNTED PROLYSTICA ACU-HOLD CONTAINER SYSTEM, WALL RACK FOR CHEMICAL CONTAINERS; (WDXDH) IN: 39.00 x 15.00 x 23.00 LBS: 100.00 MM: 990.60 x 381.00 x 584.20 KG: 45.36 LOCATION: WALL NOTES: 2025.08.19 ADDED PER STERIS CHANGE CODE: A CHANGE NOTES: ADDED PER STERIS DRAWINGS	STERIS CORPORATION ACU-HOLD PROLYSTICA	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE
5		1015520 CAB04	1	CABINET, DRYING 29" PASS THROUGH DRYING CABINET; 9-SHELF; TOUCH-SCREEN CONTROL; SCOPE HANGERS (WDXDH) IN: 29.00 x 31.00 x 78.00 LBS: 650.00 MM: 736.60 x 787.40 x 1981.20 KG: 294.84 LOCATION: FLOOR ANCHOR DETAIL REQ: YES VOLTS: 120 HERTZ: 60 WATS: 1800 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: 15A STANDARD DUPLEX RECEPTACLE	STERIS CORPORATION DCR100101 AMSCO	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: DECONTAMINATION
ROOM NUMBER: 633

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		098933 CRU03	1	CART, UTILITY 3 SHELF MOBILE CART; 1200 LB LOAD CAPACITY TO INCLUDE: (WDXH) IN: 36.00 x 24.00 x 37.50 LBS: 55.00 MM: 914.40 x 609.60 x 952.50 KG: 24.95 LOCATION: FLOOR	QUANTUM STORAGE SYSTEMS UTILITY CART WRC-2436-3CG	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
7		078823 SHW02	1	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
8		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9		081027 TST01	1	TESTER LEAKAGE TESTER EVIS TO INCLUDE: (WDXH) IN: 4.00 x 7.00 x 7.00 LBS: 5.00 MM: 101.60 x 177.80 x 177.80 KG: 2.27 LOCATION: COUNTERTOP VOLTS: 120 HERTZ: 60 AMPS: .2 ELECT NOTE: SINGLE PHASE	OLYMPUS AMERICA INC. MU-1/MB-155	OF/OI	EXISTING	GROUP: 2 MAJOR MOVABLE
10		1015402	1	SHARPS, DISPOSAL, FLOOR BIN 8 GAL SHARPS WITH SLIDING LID (WDXH) IN: 10.50 x 15.50 x 17.25 LBS: 1.00 MM: 266.70 x 393.70 x 438.15 KG: 0.46 LOCATION: FLOOR NOTES: 2024-0924 CHANGED FROM NEW TO INFO PER SHD	COVIDIEN 8980S 8 GAL	OF/OI	INFO	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: DECONTAMINATION
ROOM NUMBER: 633

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
11		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WXDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
12		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
13		100859 WST02	1	CONTAINER, BIOHAZARD RED; 64 GALLON; 2 WHEEL; HINGED LID; 2-10"WHEELS; 2-4"CASTERS; PLASTIC (WXDXH) IN: 24.25 x 32.00 x 44.25 LBS: 45.30 MM: 615.95 x 812.80 x 1123.95 KG: 20.55 LOCATION: FLOOR	TOTER INC. RMC64-00RED RMC64	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		100860 WST01	1	WASTE RECEPTACLE GRAY;64 GALLON;224LB CAPACITY;HINGED LID;PLASTIC;2-10"WHEELS;2-4"CASTERS (WXDXH) IN: 25.00 x 32.00 x 45.00 LBS: 30.50 MM: 635.00 x 812.80 x 1143.00 KG: 13.84 LOCATION: FLOOR	TOTER INC. ACC64-54647	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1015528 STE01	1	STERILIZER 16X16X26 CHAMBER;FREESTANDING;CABINET SINGLE DOOR W/VERT SLIDE;PREVAC W/LIQUID CYCLE;TOUCHSCREEN OPERATION;IMPACT PRINTER; (WDXDH) IN: 30.00 x 36.00 x 75.00 LBS: 750.00 MM: 762.00 x 914.40 x 1905.00 KG: 340.20 LOCATION: FLOOR SPEC COMMENTS: SERVICE CLEARANCE REQUIRED. REFERENCE VENDOR'S TECHNICAL DATA DETAILS NOTES: 2025.03.28 REDUCE TO 1 <i>DOES NOT REQUIRE RO/DI WATER</i> ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: TCP/IP NETWORK 10/100BASET ETHERNET CONNECTION FOR PROCONNECT RESPONSE CENTER (OPTIONAL) 5 GB OF AVAILABLE HARD DRIVE SPACE TO RUN THE SERVICE AGENT. REFERENCE VENDOR'S TECHNICAL DATA VOLTS: 480 HERTZ: 60 AMPS: 36 DEDICATED CIRCUIT: YES ELECT NOTE: THREE PHASE, 36A REQUIRED FOR STEAM GENERATOR; ADDITIONAL 120V 1-PHASE, 9.5A REQUIRED FOR CONTROLS	STERIS CORPORATION AMSCO 400 SR0101110011	OF/VI	NEW	GROUP: 1 FIXED
					BTU: 4300		COLD WATER: YES HOT WATER: YES DRAIN REQUIRED: YES PLUMB NOTE: REFERENCE VENDOR'S TECHNICAL DATA FOR DETAILS	

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
2		1001979 STE04*	1	STERILIZER 480V;ELECTRIC STEAM GENERATOR;FREESTANDING;VERTICAL SLIDING DOOR;TOUCH SCREEN DISPLAY;2-STAGE VAC. PUMP;THERMAL (WDXH) IN: 40.00 x 53.00 x 90.00 LBS: 2327.00 MM: 1016.00 x 1346.20 x 2286.00 KG: 1055.51 LOCATION: FLOOR SPEC COMMENTS: 2600LBS. OPERATING WEIGHT NOTES: 2025.03.28 APPROVED - 1 AMSCO 600 1 AMSCO 400 ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CHANGED FROM AMSCO 400 TO AMSCO 600 DATA NOTE: FOR OPTIONAL PROCONNECT VOLTS: 480 HERTZ: 60 AMPS: 7 ELECT NOTE: SEE VENDOR TECHNICAL INFORMATION STERILIZER: 3 PH, 7A CURRENT, MAX PROTECTION: 20A, WIRE SIZE #12AWG STEAM GENERATOR: 480V;57A;MAX PROTECTION: 80A, WIRE SIZE #6AWG GENERATOR CONTROL: 120V;60HZ;10A;MAX PROTECTION 20A, WIRE SIZE #12AWG	STERIS CORPORATION AMSCO 600 939V-1	OF/VI	NEW	GROUP: 1 FIXED
3		1002472 STA01*	1	STERILIZER ACCESSORY,LOADING CAR/TRANSFER CARRIAGE FOR 39" LENGTH CHAMBER; 9 TRAYS MAX TO INCLUDE: (WDXH) IN: 26.44 x 49.75 x 58.75 LBS: 700.00 MM: 671.58 x 1263.65 x 1492.25 KG: 317.52 LOCATION: FLOOR NOTES: 2025.0128 ADDED PER SHD CHANGE CODE: A CHANGE NOTES: 2025.0128 ADDED PER SHD	STERIS CORPORATION AY1596602/AY1598102 AMSCO 600 39	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
4		1015530 STE02	1	STERILIZER, LOW TEMP LOW TEMP; SINGLE DOOR;VAPROX HC STERILANT;19 MINUTE FAST CYCLE;PRE-PROGRAMMED LUMEN CYCLE;NON LUMEN CYCLE;FAST CYCLE;FLEXIBL (WXDXH) IN: 22.50 x 33.00 x 61.50 LBS: 540.00 MM: 571.50 x 838.20 x 1562.10 KG: 244.94 LOCATION: FLOOR ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: 10/100 BASE T SUPPORTING TCP/IP PROTOCOL RJ45F CONNECTION VOLTS: 120 HERTZ: 60 AMPS: 20 DEDICATED CIRCUIT: YES ELECT NOTE: SINGLE PHASE	STERIS CORPORATION VP60002401 V-PRO S2	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
5		1000106 STE03	1	CLEANER, SCOPE AUTOMATED W/COMPRESSOR;DUAL BASIN;RAPICIDE PA DISINFECTANT;INDIVIDUAL CHANNEL MONITORING (WXDXH) IN: 49.00 x 31.00 x 62.00 LBS: 400.00 MM: 1244.60 x 787.40 x 1574.80 KG: 181.44 LOCATION: FLOOR SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA FOR INSTALLATION INFORMATION AND SITE READINESS GUIDE NOTES: POWER REQUIREMENTS: 2X NEMA 5-15 RECEPTACLES REQUIRE (X1 FOR UNIT, X1 FOR COMPRESSOR) ANCHOR DETAIL REQ: YES DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: WIRELESS AVAILABLE WITH CUSTOMER SUPPLIED WIRELESS ADAPTER VOLTS: 120 HERTZ: 60 WATS: 1000 DEDICATED CIRCUIT: YES PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: NEMA 5-15R; UPS WITH A MINIMUM RATING OF 1500 VA/865 WATTS IS REQUIRED	MEDIVATORS REPROCESSING SYSTEMS ADV-1007 ADVANTAGE PLUS SS	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
					BTU: 2000/HR MECH NOTE: REFERENCE VENDOR'S TECHNICAL DATA FOR FULL VENTILATION INFORMATION HOUSE AIR 58 TO 125 PSI WITH A MINIMUM FLOW VOLUME OF 1.1 CFM, MAX PARTICLE SIZE OF 5 MICRONS, MAXIMUM DEW POINT OF 7°C (45°F), MAXIMUM OIL CONCENTRATION OF 5 MG/M3; IDLE: 500 BTU/HR		COLD WATER: YES TREATED WATER: YES DRAIN REQUIRED: YES PLUMB NOTE: REFERENCE VENDOR'S TECHNICAL DATA FOR FULL DETAILS MIN FLOW RATE1.32 GPM AT 35-87 PSI; HARDNESS, LESS THAN 12 GPG (200 PPM); 3.17 GPM, WITH THE OPENING NO HIGHER THAN 18" FROM THE FLOOR, VENTED	

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		104990 SCP01	1	CLEANER, ACCESSORY, FLUSHING AID ENDOCOPE FLUSHING AID;MANUAL CLEANING (WXDXH) IN: 7.50 x 11.50 x 10.30 LBS: 10.00 MM: 190.50 x 292.10 x 261.62 KG: 4.54 LOCATION: COUNTERTOP SPEC COMMENTS: ;MUST BE LOCATED WITHIN 3' ABOVE THE FLUID BASIN. THE DISTANCE BETWEEN THE SINK SOLUTION & SCOPE BUDDY MUST NOT EXCEED 3' VOLTS: 115 HERTZ: 60 AMPS: 1 WATS: 115 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: SINGLE PHASE	MEDIVATORS REPROCESSING SYSTEMS SBP-1000 SCOPE BUDDY PLUS	OF/CI	NEW	GROUP: 1 FIXED
7		107241 TWA01	1	TABLE, WORK, ADJUSTABLE AMSCO PREP & PACK DELUXE ELEC HT ADJ WORKSTATION; MOBILE; SS SURFACE; 500LB WT CAP (WHEN EVENLY DISTRIBUTED); 8' POWER (WXDXH) IN: 60.00 x 33.75 x 78.50 LBS: 390.00 MM: 1524.00 x 857.25 x 1993.90 KG: 176.91 LOCATION: FLOOR SPEC COMMENTS: SPACE REQUIREMENT: 61L X 36W X 86H IN INCHES MAX HEIGHT: 86" IN INCHES ANCHOR DETAIL REQ: TBD CHANGE CODE: C CHANGE NOTES: CHANGED FROM OF/OI TO OF/VI VOLTS: 120 HERTZ: 60 AMPS: 15 WATS: 1800 ELECT NOTE: 15A;REFERENCE VENDOR TECHNICAL INFORMATION (2) OUTLETS WITHIN 4' OF UNIT REQUIRED	STERIS CORPORATION PREP63 DELUXE ACCESS	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
8	No image available	1023696	1	FREIGHT FREIGHT FOR STERIS AND STERIS ENDODRY EQUIPMENT (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE	STERIS CORPORATION	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
9	No image available	1023697	1	INSTALLATION INSTALLATION COSTS FOR STERIS EQUIPMENT AND STERIS ENDO EQUIPMENT (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE	STERIS CORPORATION	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
10	No image available	1023698	1	SERVICES TO INCLUDE: PROJECT DESIGN FEES PER STERIS QUOTE (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE	STERIS CORPORATION	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
11		097455 SHW05	1	SHELVING, WIRE 4 SHELVES;SOLID BOTTOM;CASTERS TO INCLUDE: (WDXH) IN: 30.00 x 18.00 x 68.00 LBS: 65.50 MM: 762.00 x 457.20 x 1727.20 KG: 29.72 LOCATION: FLOOR NOTES: NO CASTERS ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 63UP/A1830NC SUPER ADJUST	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
12		098933 CRU03	1	CART, UTILITY 3 SHELF MOBILE CART; 1200 LB LOAD CAPACITY TO INCLUDE: (WDXH) IN: 36.00 x 24.00 x 37.50 LBS: 55.00 MM: 914.40 x 609.60 x 952.50 KG: 24.95 LOCATION: FLOOR	QUANTUM STORAGE SYSTEMS UTILITY CART WRC-2436-3CG	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
13		106653 STL02	1	STOOL, REVOLVING 300 LB WEIGHT CAPACITY, HEIGHT ADJ, W/GLIDES UPHOLSTERY TO BE (WDXH) IN: 23.00 x 23.00 x 24.50 LBS: 24.00 MM: 584.20 x 584.20 x 622.30 KG: 10.89 LOCATION: FLOOR	MIDMARK CORPORATION 272-001-XXX VALUE SERIES 272	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
14		1003538 PEG01	1	BOARD WITH PEGS, WALL MOUNTED 4'X4'; 16 GAUGE STAINLESS STEEL; 1/4" DIAMETER HOLES ON 1' (WDXH) IN: 48.00 x 0.88 x 48.00 LBS: 46.00 MM: 1219.20 x 22.36 x 1219.20 KG: 20.87 LOCATION: WALL	LOGIQUIP SSPEG4848	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STORAGE / PREP & PACK / STERILIZER
ROOM NUMBER: 634

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
15		100745 RCK02	1	RACK STERILE WRAP RACK;CHROME;5" CASTERS;ADJUSTABLE POST HEIGHT;4 BARS;WIRE BOTTOM SHELF TO INCLUDE: (WXDXH) IN: 60.00 x 24.00 x 68.00 LBS: 94.00 MM: 1524.00 x 609.60 x 1727.20 KG: 42.64 LOCATION: FLOOR	INTERMETRO INDUSTRIES CORP. SWR60S/2460FG/ACC STERILE WRAP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
16		074525 IND01	1	STERILIZER, INDICATOR INDICATOR READER/INCUBATOR;USE W/STEAM STERILIZERS;READS THE FLUORESCENCE PRODUCED BY BIOLOGICAL INDICATORS TO INDICATE (WXDXH) IN: 8.75 x 9.00 x 3.25 LBS: 8.00 MM: 222.25 x 228.60 x 82.55 KG: 3.63 LOCATION: COUNTERTOP NOTES: 2025.05.12 UPDATED FROM EXISTING TO NEW PER SHD CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 AMPS: 0.6 ELECT NOTE: TO BE USED ONLY WITH THE SUPPLIED POWER MODULE	3M ATTEST AUTO-READER 390	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
17		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WXDXH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STERILE & CLEAN SUPPLY
ROOM NUMBER: 635

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		078823 SHW02	4	SHELVING, WIRE 5 SHELF;MOBILE;CHROME;CLEAR INLAY BOTTOM SHELF TO CONSIST OF: (WDXDH) IN: 48.00 x 18.00 x 80.00 LBS: 84.75 MM: 1219.20 x 457.20 x 2032.00 KG: 38.45 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1848NC/74P SUPER ERECTA	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
2		1015581 SHW04	1	SHELVING, WIRE 4 SHELF,STATIONARY, CHROME; SEISMIC TO CONSIST OF THE FOLLOWING: (WDXDH) IN: 36.00 x 18.00 x 63.00 LBS: 53.00 MM: 914.40 x 457.20 x 1600.20 KG: 24.05 LOCATION: FLOOR ANCHOR DETAIL REQ: YES	INTERMETRO INDUSTRIES CORP. 1836NC/63P SUPER ERECTA SEISMIC	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
3		098933 CRU03	1	CART, UTILITY 3 SHELF MOBILE CART; 1200 LB LOAD CAPACITY TO INCLUDE: (WDXDH) IN: 36.00 x 24.00 x 37.50 LBS: 55.00 MM: 914.40 x 609.60 x 952.50 KG: 24.95 LOCATION: FLOOR	QUANTUM STORAGE SYSTEMS UTILITY CART WRC-2436-3CG	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
4		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXDH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
5		004363	1	WASTE RECEPTACLE CAP 28 QT; RECTANGULAR; FIRE RESISTANT; FIBERGLASS; CA FIRE MARSHAL APPROVED; UL LISTED; 6/CARTON; CLIENT TO CHOOSE COLOR (WDXDH) IN: 14.50 x 10.50 x 15.30 LBS: 5.00 MM: 368.30 x 266.70 x 388.62 KG: 2.27 LOCATION: FLOOR HOSPITAL TAG: - GRAY	RUBBERMAID COMM. PRODUCTS FG254300 2543	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STERILE & CLEAN SUPPLY
ROOM NUMBER: 635

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
6		104233 CAB08*	1	CABINET, SCOPE 8-SCOPE DRYING CABINET;SS;HINGE LEFT LOAD AND UNLOAD;BARCODE SCANNER,SCOPE TRACKING (WXDXH) IN: 24.00 x 22.00 x 75.00 LBS: 375.00 MM: 609.60 x 558.80 x 1905.00 KG: 170.10 LOCATION: FLOOR NOTES: 2024-11-27 CHANGED CABINET FROM MASS MEDICAL TO STERIS PER BRADY T ANCHOR DETAIL REQ: YES CHANGE CODE: C CHANGE NOTES: CHANGED CABINET TO STERIS DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ45 INTERFACE NOTE NETWORK CONNECTION LOCATION ABOVE TOP OF UNIT VOLTS: 120 HERTZ: 60 AMPS: .62 WATS: 75 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: FUSE:T2A (INTERNAL DEVICE) 15A (CIRCUIT) NOTE ELECTRICAL CONNECTION LOCATION ABOUT TOP OF UNIT.	STERIS CORPORATION ENDODRY CABINET DRY-1001	OF/VI-CC	NEW	GROUP: 3 MINOR MOVABLE
7		1018690 SCRT01*	1	CART, SCOPE, TRANSPORT 6 TRAY ENDO TRANSPORT CART (WXDXH) IN: 25.10 x 22.40 x 41.70 LBS: 91.00 MM: 637.54 x 568.96 x 1059.18 KG: 41.28 LOCATION: FLOOR NOTES: 2025.0128 ADDED PER USER MEETING AND CHANGE TO ENDODRY CABINET CHANGE CODE: A CHANGE NOTES: 2025.0128 ADDED PER SHD	MEDIVATORS REPROCESSING SYSTEMS 103466 CLEANASCOPE ADVANTAGE	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: STERILE & CLEAN SUPPLY
ROOM NUMBER: 635

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
8		1003252 CBW03	1	CABINET, WARMING, DUAL, FREESTANDING DUAL FREESTANDING FLUID AND BLANKET WARMING CABINET WITH WINDOW GLASS DOORS. FEATURES 18 INCH DEPTH PERFECT FOR BUILT-IN OR (WDXDH) IN: 30.90 x 21.70 x 67.25 LBS: 358.00 MM: 784.86 x 551.18 x 1708.15 KG: 162.39 LOCATION: FLOOR SPEC COMMENTS: 50" DEEP WITH DOOR OPEN NOTES: 2024.09.16 SHD CONFIRMED CBW01 IS ENTHERMICS CBW03 IS PEDIGO VOLTS: 120 HERTZ: 60 AMPS: 14.2 WATS: 1.7KW PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: SINGLE PHASE.;8FT CORD	PEDIGO PRODUCTS, INC. P-2148	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE

END OF ROOM

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
1		1023173 MGAA01	1	BOOM, ANESTHESIA MOUNTING PLATE: SINGLE COMMON PLATE; FRICTION BRAKE SYSTEM; CEILING COVER: STANDALONE MFR CONTROLS: REAR ONLY;FRONT MFR (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING ANCHOR DETAIL REQ: YES DATA NOTE: LOW VOLTAGE SERVICES QUAD RJ-45 CAT 6: PLATE 1 1 EA. BLANK PLATE COR IP 4-IN-1 PLATE: 1 PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: 20A/125V DUPLEX (4 OUTLETS): 2; 20A/125V DUPLEX (6 OUTLETS): 1 PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL S-SERIES	OF/VI	NEW	GROUP: 1 FIXED
2		1023175 BEQ02	1	BOOM, EQUIPMENT SINGLE COMMON MOUNTING PLATE; 4 SHELVES, SHELF 1: 1000MM, SHELF 2: 750MM W/CONTROLS, SHELF 3: 750MM, SHELF 4: 750MM WITH (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING ANCHOR DETAIL REQ: YES DATA NOTE: 1 EA. DATA PASS THRU PLATE 1 EA. DISTRIBUTION BD PLATE 7 EA. BLANK PLATE 2 EA. COR-IP 4-IN-1 PLATE PLUG TYPE: Type B (NEMA 5-20) ELECT NOTE: 4 EA. 20A; 125V DUPLEX OUTLET (6 OUTLETS) PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL S-SERIES	OF/VI	NEW	GROUP: 1 FIXED

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
3		1017713	1	MONITOR, VIDEO 32" 4K OLED SURGICAL DISPLAY (WXDXH) IN: 28.70 x 2.10 x 18.30 LBS: 19.25 MM: 728.98 x 53.34 x 464.82 KG: 8.74 LOCATION: SPACE DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 2.5 WATS: 115.2 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: HOSPITAL GRADE AC POWER CORD PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS 0240031300 32 INCH 4K OLED SYRGUCAF DISPLAY	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
4	No image available	1023185	1	BASE HUB CONNECTED OR HUB BASE SYSTEM W/SDP1000 PRINTER (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE SPEC COMMENTS: ; PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL 0240200800	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
5		1013822 MON01	1	MONITOR, VIDEO 55" LCD 4K DISPLAY (WXDXH) IN: 48.80 x 2.40 x 28.10 LBS: 60.80 MM: 1239.52 x 60.96 x 713.74 KG: 27.58 LOCATION: SPACE DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: RJ-45 10 BASE-T/100 BASE-TX X 2 REFERENCE VENDOR'S TECHNICAL DATA VOLTS: 120 HERTZ: 60 AMPS: 3.3 WATS: 295 ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS P45952 P45952	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
6		1021686 MON01A	1	BRACKET, TELEVISION, WALL UNIVERSAL; DEPTH 2.73-11.18"; BLACK (WXDXH) IN: 24.38 x 11.18 x 16.78 LBS: 30.40 MM: 619.26 x 283.98 x 426.22 KG: 13.79 LOCATION: WALL	PEERLESS SALES COMPANY SP850-UNL SMARTMOUNT	OF/CI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS
				<i>NOTES: THE WALL LCD BRACKETS SHIP WITH THE CAPITAL EQUIPMENT. THE GC TYPICALLY INSTALLS THESE JUST BEFORE THE EQUIPMENT INSTALL OR EVEN THE SAME WEEK AS THE EQUIPMENT INSTALLATION. IF THE GC WANTS BEFOREHAND, CONFIRM WITH STRYKER ALTERNATE ARRANGEMENTS</i> CHANGE CODE: C CHANGE NOTES: ON SEPERATE LINE AS OFCI PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910			
7		1023171	1	LIGHT, SURGICAL TC ONLY MOUNT;SURGICAL HIGH DEFINITION MODE;MICROSPOT AND SHD MODE;RANGE OF SPOT SIZES;NO CARDANIC CONTROL;TOUCH CONTROL UNI (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL OCULAN	OF/VI	NEW
				GROUP: 1 FIXED MECH NOTE: REFERENCE VENDOR'S TECHNICAL DATA SELLER: STRYKER COMMUNICATIONS			
8		1023170	1	LIGHT, SURGICAL TC ONLY MOUNT;SURGICAL HIGH DEFINITION MODE;MICROSPOT AND SHD MODE;RANGE OF SPOT SIZES;NO CARDANIC CONTROL;TOUCH CONTROL UNI (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL OCULAN	OF/VI	NEW
				GROUP: 1 FIXED MECH NOTE: REFERENCE VENDOR'S TECHNICAL DATA SELLER: STRYKER COMMUNICATIONS			

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
9		1023176 AUD01	1	AUDIO/VISUAL SYSTEM, INTEGRATED, SURGERY 4K AND HD VIDEO ROUTING; PC HD HDMI/DVI INPUT; 4K DISPLAYS (X2) WITH OPTIONAL 55IN 4K DISPLAY; 4-IN-1 BOOM PLATE (X2); (WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: CEILING SPEC COMMENTS: REFERENCE VENDOR'S TECHNICAL DATA DATA NOTE: REFERENCE VENDOR'S TECHNICAL DATA WIRED; LAN; HDMI/DVI; ETHERNET; AUDIO SYSTEM ELECT NOTE: REFERENCE VENDOR'S TECHNICAL DATA PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL COR LITE	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
10		1019423 ORC01	1	SCOPE, VIDEO SYSTEM CONNECTED OR CART - MOBILE, MEDICAL GRADE CART TO ACCOMMODATE STRYKER ADVANCED IMAGING SYSTEM, ENDOSCOPIC FLUORESCENCE (WXDXH) IN: 30.00 x 30.00 x 90.00 LBS: 570.00 MM: 762.00 x 762.00 x 2286.00 KG: 258.55 LOCATION: FLOOR SPEC COMMENTS: 1688 4K CAMERA SYSTEM: 13X16.63X4.46 WXDXH IN INCHES, 12LBS. PRINTER: 12.49X16.74X8.15 WXDXH IN INCHES, 25.36LBS. SURGICAL DISPLAY: 29.7X3.0X17.8 WXDXH IN INCHES, 23.1LBS. L11 LED LIGHT: 12.5X16.8X12.5 WXDXH IN INCHES, 16 LBS. PHASE: EARLY-PO/LEASED NOTES: EQUIPMENT WILL LIVE ON OR BOOM DATA HARDWIRED: YES WIRELESS: YES DATA NOTE: ETHERNET RJ45; USB; HDMI; DVI VOLTS: 120 HERTZ: 60 AMPS: 3.4 WATS: 1700 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: ELECTRICAL SHOWN FOR CART SPD1000 PRINTER: 100- 240VAC; 50/60HZ MONITOR: 100-240VAC; 50/60HZ; 2.5A PO REQ #: 944A-00006 CLIENT PO #: SNCA972	STRYKER ENDOSCOPY 1688010000 1688 AIM 4K	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
11		096403 ANES01	1	ANESTHESIA MACHINE FABIUS GS PREMIUM; 3 GAS; VAPORIZER TO INCLUDE: (WXDXH) IN: 41.00 x 33.00 x 52.00 LBS: 296.00 MM: 1041.40 x 838.20 x 1320.80 KG: 134.27 LOCATION: FLOOR	DRAGER MEDICAL, INC. 8607000/ACC FABIUS GS PREM	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
				SPEC COMMENTS: VENDOR SET-UP REQUIRED; NOTES: ENSURE VENDOR UPS PROCURED 2025.05.12 UPDATED FROM EXISTING TO NEW CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC DATA HARDWIRED: YES WIRELESS: NO DATA NOTE: (1) RS232 VOLTS: 120 HERTZ: 60 AMPS: 2.3 WATS: 70 PLUG TYPE: Type B (NEMA 5-15) ELECT NOTE: NEMA 5-15R HOSPITAL GRADE				
12		1023058 ACRT01*	1	CART, ANESTHESIA 6 DRAWERS: (3) 3"; (2) 6"; (1) 9" MD30-ANS ACCESSORIES INCLUDE: (WXDXH) IN: 37.50 x 22.00 x 66.75 LBS: 198.00 MM: 952.50 x 558.80 x 1695.45 KG: 89.82 LOCATION: FLOOR SPEC COMMENTS: ; NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	HARLOFF COMPANY, INC. MDS3030K06+MD30-ANS M-SERIES	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
13		084123 ESU05	1	ELECTROSURGICAL UNIT TWO GENERATORS ON A CART TO INCLUDE: (WXDXH) IN: 21.00 x 29.00 x 48.00 LBS: 131.20 MM: 533.40 x 736.60 x 1219.20 KG: 59.52 LOCATION: FLOOR NOTES: LOCATION ON BOOM CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS AND MODEL CHANGE PER SHD REVIEW VOLTS: 120 HERTZ: 60 WATS: 320 ELECT NOTE: 100-240V;50/60HZ;180VA	OLYMPUS AMERICA INC. ESG-400/USG-400/ACC THUNDERBEAT	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
					MECH NOTE: BTU INFORMATION UNAVAILABLE FROM MANUFACTURER			

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
14		1023112 SMK01*	1	SMOKE EVACUATION, SURGICAL SURGICAL SMOKE EVACUATOR, FEATURES INCLUDE: AUTOMATIC ACTIVATION DEVICE PORT - CAN BE COUPLED TO ESG-400 AND (WXDXH) IN: 11.00 x 15.50 x 6.00 LBS: 12.00 MM: 279.40 x 393.70 x 152.40 KG: 5.45 LOCATION: COUNTERTOP NOTES: 2025-0729 ADDED PER SHD, LOCATED ON BOOM CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER VOLTS: 120 HERTZ: 60	OLYMPUS AMERICA HYVV120 OR-VAC	OF/OI	NEW	GROUP: 2 MAJOR MOVABLE
15		1015753 PHY07	1	MONITOR, PHYSIOLOGICAL W/ANESTHESIA MOUNT, 15" LCD TFT TOUCH DISPLAY; 6-WAVES; SIDE MOUNTS TO ANESTHESIA MACHINE; 120 VAC TO INCLUDE: (WXDXH) IN: 15.90 x 7.20 x 12.10 LBS: 22.30 MM: 403.86 x 182.88 x 307.34 KG: 10.12 LOCATION: SPACE NOTES: 2025.07.22 ALL PHILIPS MONITORING TO BE REVIEWED IN SEPTEMBER 2025. DATA HARDWIRED: YES WIRELESS: NO VOLTS: 120 HERTZ: 60 AMPS: 1.2 WATS: 70	PHILIPS MEDICAL SYSTEMS 866066_NAM MX550	OF/VI	NEW	GROUP: 2 MAJOR MOVABLE
16		1004828 GAS01	1	ANESTHESIA MACHINE ACCESSORY INTELLIVUE ANESTHESIA GAS MODULE; GAS ANALYZER MODULE; PROVIDES A NON-DISPERSIVE INFRARED MEASUREMENT OF RESPIRATORY AND (WXDXH) IN: 3.00 x 6.75 x 4.25 LBS: 3.00 MM: 76.20 x 171.45 x 107.95 KG: 1.37 LOCATION: SPACE NOTES: 2025-0908 CONFIRMED KEEP FOR ETCO2	PHILIPS MEDICAL SYSTEMS 866173 G7 MODULE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
17		1023179 TBS01	1	TABLE, SURGICAL 1000LBS CAP NORMAL POSITION;500LBS FOR REVERSE POSITION; HEIGHT ADJ. (WXDXH) IN: 23.78 x 94.60 x 42.60 LBS: 725.00 MM: 604.02 x 2402.84 x 1082.04 KG: 328.86 LOCATION: FLOOR ANCHOR DETAIL REQ: NO VOLTS: 120 HERTZ: 60 AMPS: 3.0 PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL D770	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
18		082570 IVP01	1	I.V. INFUSION PUMP SMART PUMP TECHNOLOGY;SINGLE STEP TITRATION LIMIT;MULTIPLE MEDS ERROR PREVENTION (WXDXH) IN: 4.20 x 2.50 x 5.80 LBS: 2.00 MM: 106.68 x 63.50 x 147.32 KG: 0.91 LOCATION: COUNTERTOP SPEC COMMENTS: ; VOLTS: 115 HERTZ: 60 AMPS: .3 ELECT NOTE: • INPUT: 115 VAC ±15%, 50 - 60 HZ / 300 MA MAX • OUTPUT: 9 VDC/1200 MA, SHORT CIRCUIT PROTECTED	BAXTER HEALTHCARE/I.V. DIVISION 35700BAX2 SIGMA SPECTRUM	OF/OI	LEASED	GROUP: 2 MAJOR MOVABLE
19		051488 IVP02	1	I.V. POLE STAINLESS STEEL; 4 HOOK; 5 CASTER BASE; HAND OPERATED (WXDXH) IN: 21.25 x 21.25 x 118.00 LBS: 20.00 MM: 539.75 x 539.75 x 2997.20 KG: 9.08 LOCATION: FLOOR SPEC COMMENTS: HEIGHT ADJUSTABLE FROM 69" TO 118" CHANGE CODE: C CHANGE NOTES: 2025.05.12 STATUS CHANGE PER SHD REVIEW 2 DOC	PEDIGO PRODUCTS, INC. P-1572-4 FOUR HOOK	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
20		093116	3	REGULATOR, SUCTION, CONTINUOUS TWO-MODE W/DIAMOND & LOCKING GLAND OVERFLOW SAFETY TRAP (WXDXH) IN: 3.00 x 5.00 x 7.00 LBS: 1.10 MM: 76.20 x 127.00 x 177.80 KG: 0.50 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE; CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC ELECT NOTE: 2-3AA BATTERIES	OHIO MEDICAL CORPORATION 8701-1225-905	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
21		1004558	3	FLOWMETER, AIR 15 LPM;OHIO DIAMOND ADAPTER OHMEDA W/50 PSI PORT USA; PB (WXDXH) IN: 1.25 x 3.50 x 5.50 LBS: 0.40 MM: 31.75 x 88.90 x 139.70 KG: 0.19 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UA-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
22		1004559	3	FLOWMETER, OXYGEN FLOWMETER, OXYGEN 15 LPM W/50 PSI TAKE OFF (WXDXH) IN: 1.10 x 3.40 x 5.50 LBS: 0.35 MM: 27.94 x 86.36 x 139.70 KG: 0.16 LOCATION: SPACE SPEC COMMENTS: CONFIRM CONNECTION TYPE;	OHIO MEDICAL CORPORATION FM-15UO-OHPT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
23		1003921 LHM01	1	HAMPER, LINEN RECTANGULAR; W/ LID; FOOT PEDAL; POWDER-COATED WHITE GLOS (WXDXH) IN: 18.50 x 19.50 x 32.75 LBS: 25.00 MM: 469.90 x 495.30 x 831.85 KG: 11.34 LOCATION: FLOOR	MEDLINE INDUSTRIES, INC. TRICX302 29 INCH	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
24		028909 TBW01	1	TABLE, WORK STAINLESS;W/UNDERSHELF;3 IN CONDUCTIVE CASTERS (WXDXH) IN: 60.00 x 24.00 x 34.00 LBS: 88.00 MM: 1524.00 x 609.60 x 863.60 KG: 39.92 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. STERIL-GARD SG-94-SS	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE
25		051630 TBW02	1	TABLE, WORK S.S.;WITH UNDER-SHELF;MOBILE TO INCLUDE: (WXDXH) IN: 36.00 x 24.00 x 34.00 LBS: 58.00 MM: 914.40 x 609.60 x 863.60 KG: 26.31 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. SG-98-SS/ACC -	OF/OI	EXISTING	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
26		1006224 CRS02	1	CART, SUPPLY NARROW ALUMINUM CART;24"DRAWER SPACE;5" SWIVEL CASTERS (WXDXH) IN: 17.50 x 21.25 x 36.00 LBS: 85.00 MM: 444.50 x 539.75 x 914.40 KG: 38.56 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	INNERSPACE BY SOLAIRE MEDICAL SPN24W5 NARROW PACE 24	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
27		1006176 CRS03	1	CART, SUPPLY 3 3"-, 2 6"-, AND 1 9"-HIGH DRAWER CONFIGURATION;ACCOMMODATES A VARIETY OF INTERIOR AND EXTERIOR ACCESSORIES;LIGHTWEIGHT (WXDXH) IN: 29.25 x 21.25 x 42.50 LBS: 145.00 MM: 742.95 x 539.75 x 1079.50 KG: 65.78 LOCATION: FLOOR NOTES: 2025.09.18 SHD TO SPECIFY DRAWER CONFIG AND COLOR DURING PROCUREMENT 2025.09.09 ADDED PER SHD FINAL REVIEW DOC CHANGE CODE: A CHANGE NOTES: 2025.09.09 ADDED PER SHD	INNERSPACE BY SOLAIRE SP30W6 PACE 30	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
28		089895 STL01	1	STOOL, REVOLVING ERGO;W/BACK;PNEUMATIC HGT ADJ;2" CASTERS 18X18 1/2" CONTOURED SEAT CUSHION W/ADJUSTMENT; (WXDXH) IN: 25.00 x 25.00 x 50.00 LBS: 38.00 MM: 635.00 x 635.00 x 1270.00 KG: 17.24 LOCATION: FLOOR CHANGE CODE: C CHANGE NOTES: 2025.05.12 QTY CHANGE PER SHD REVIEW 2 DOC	PEDIGO PRODUCTS, INC. T-584 ANESTHESIA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
29		1023017 STL04*	2	STOOL, REVOLVING 360° SWIVEL; ARMREST; HEIGHT ADJUSTABLE; FOOTREST (WXDXH) IN: 18.00 x 18.00 x 38.60 LBS: 40.00 MM: 457.20 x 457.20 x 980.44 KG: 18.15 LOCATION: FLOOR NOTES: 2025.05.12 ADDED PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	OUBO FURNITURE (AMAZON) BOC2D253PB	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
30		080356 SCT01	1	STAND, SUCTION CANISTER, MOBILE COLLECTION CANISTER CARRIER; 4-CANISTER: 36" HIGH;ROLLING STAND (WDXH) IN: 24.00 x 24.00 x 36.00 LBS: 10.00 MM: 609.60 x 609.60 x 914.40 KG: 4.54 LOCATION: FLOOR	ALLIED HEALTHCARE PRODUCTS, INC. 20-02-0023 4-CANISTER MODEL	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
31		105431 SUTR01*	1	RACK, SUTURE DISPENSER RACK;INCLUDES 20 MODULES/10 BACK PLATES W/4 FASTENERS/20 MODULE (WDXH) IN: 25.75 x 5.75 x 27.00 LBS: 17.40 MM: 654.05 x 146.05 x 685.80 KG: 7.90 LOCATION: WALL NOTES: 2025.09.16 ETHICON MR20 CONFIRMED PER SHD	ETHICON ENDO-SURGERY, INC. MR20	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
32		1010995 CAB02	1	CABINET, MEDICATIONS, NARCOTICS, WALL-MOUNTED HEAVY DUTY 20 GAUGE STEEL CABINET, HIGH SECURITY PICK RESISTANT LOCK, ONE FIXED AND ONE ADJUSTABLE SHELF, FULL LENGTH (WDXH) IN: 23.50 x 10.50 x 29.50 LBS: 35.00 MM: 596.90 x 266.70 x 749.30 KG: 15.88 LOCATION: WALL ANCHOR DETAIL REQ: YES	HARLOFF COMPANY, INC. 2830AQ	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
33		084252	2	WASTE RECEPTACLE, STEP-ON 23 GAL;FRONT STEP;LEAK-PROOF BASE;TIGHT FITTING LID;CALIF FIRE MARSHALL APP'D;CLIENT TO CHOOSE COLOR IN OPTIONS AT (WDXH) IN: 19.80 x 16.50 x 32.60 LBS: 18.00 MM: 502.92 x 419.10 x 828.04 KG: 8.17 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS FG614600[X] LEGACY FRONT STEP	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
34		087035 DSP01	1	BRACKET, SHARPS CONTAINER LOCKING WALL BRACKET FOR 5.4QT. B.D. CONTAINERS (WDXH) IN: 10.00 x 5.00 x 10.00 LBS: 2.00 MM: 254.00 x 127.00 x 254.00 KG: 0.91 LOCATION: WALL CHANGE CODE: C CHANGE NOTES: UPDATED SPEC	BECTON DICKINSON 305447	OF/CI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
35		1003160 DSPP03	1	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL LOCKING WALL BRACKET;WHITE/BUE 3 GALLON PHARMASAFETY CONTAINER LEAK RESISTANT;TAMPER EVIDENT;SECURE HINGED LID;ABSORBENT PA (WDXDH) IN: 13.75 x 6.50 x 16.50 LBS: 5.00 MM: 349.25 x 165.10 x 419.10 KG: 2.27 LOCATION: WALL SPEC COMMENTS: DIMENSIONS OF BRACKET: 13.75" X 6.5" X 11.75" (W X D X H)	COVIDIEN 8528C/8836SA	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
36		006656 MAY01	2	STAND, MAYO FOOT OPERATED;LARGE TRAY;STAINLESS STEEL;2"CASTERS; 16 1/4 X 21 1/4 IN REMOVABLE TRAY;REINFORCED; STAINLESS STEEL TRAY (WDXDH) IN: 16.25 x 21.25 x 62.00 LBS: 33.00 MM: 412.75 x 539.75 x 1574.80 KG: 14.97 LOCATION: FLOOR SPEC COMMENTS: ;HEIGHT ADJUSTS FROM 36" TO 62"	PEDIGO PRODUCTS, INC. P-1066-SS P-1066-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
37		021273 KCB01	2	KICKBUCKET SS;REMOVABLE 12 QT. SS PAIL;TIP-RESISTANT;ROUND (WDXDH) IN: 13.75 x 13.75 x 13.25 LBS: 8.00 MM: 349.25 x 349.25 x 336.55 KG: 3.63 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. P-1020-SS P-1020-SS	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
38		089875	1	WASTE RECEPTACLE GRAY; SLIM JIM; VENTING CHANNELS; 16GAL; 4/CASE (WDXDH) IN: 11.00 x 22.00 x 25.00 LBS: 6.42 MM: 279.40 x 558.80 x 635.00 KG: 2.92 LOCATION: FLOOR	RUBBERMAID COMM. PRODUCTS 1971258 SLIM JIM	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
39		1005731 CLK01	1	CLOCK, ELAPSED TIME SURFACE MOUNT 6 DIGIT; RED; 4' DIGIT HEIGHT; 4" SLOPE (WDXH) IN: 23.50 x 2.40 x 8.00 LBS: 4.60 MM: 596.90 x 60.96 x 203.20 KG: 2.09 LOCATION: WALL NOTES: 2025.06.10 CHANGED FROM POE TO 120V PER ELECTRICAL FUTURE ITEM; ROUGH IN ONLY AT THIS TIME CHANGE CODE: C CHANGE NOTES: CHANGED FROM POE TO 120V VOLTS: 120 HERTZ: 60 ELECT NOTE: 10' CORD WITH PLUG	PRIMEX XRA7A203E LEVO SERIES	OF/CI	FUTURE	GROUP: 1 FIXED
40		099447 STL03	1	STOOL, FOOT, WITH HANDRAIL TWO STEP; STEPS ARE 8" AND 16" HIGH; CHROME CORRUGATED RUBBER SURFACE; GREY FLOOR TIPS (WDXH) IN: 16.00 x 12.00 x 47.00 LBS: 35.00 MM: 406.40 x 304.80 x 1193.80 KG: 15.88 LOCATION: FLOOR	PEDIGO PRODUCTS, INC. P-18-A-K P-18-A-K	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
41		057206 GLV01	1	DISPENSER, GLOVE, QUADRUPLE BOX HIGH GRADE CLEAR PETG PLASTIC; HOLDS 4; KEYHOLE MOUNT; (WDXH) IN: 22.45 x 3.88 x 10.00 LBS: 2.00 MM: 570.23 x 98.56 x 254.00 KG: 0.91 LOCATION: WALL	BOWMAN MANUFACTURING COMPANY QUAD GLOVE DISPENSER GP-061	OF/CI	NEW	GROUP: 3 MINOR MOVABLE
42		1023059	1	I.V. INFUSION PUMP ACCESSORY 1000ML (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.05.12 ADDED METPAK 5275 PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	RIESTER 5275	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
43		1023061	1	I.V. INFUSION PUMP ACCESSORY 3000ML (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE NOTES: 2025.05.12 ADDED METPAK #5270-536 PER SHD REVIEW 2 DOC CHANGE CODE: A CHANGE NOTES: 2025.05.12 PER SHD REVIEW 2 DOC	RIESTER 5270-536	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
44		008693 BWMR01*	1	WARMER, BLOOD/FLUID FLOW RATE UP TO 30 L/HR;SET POINT 41 DEGREES C; AUDIBLE/VISUAL ALARM IF FLUID EXCEEDS 42 DEGREES C; AUDIBLE/VISUAL UNDER (WDXH) IN: 7.50 x 10.00 x 4.50 LBS: 7.44 MM: 190.50 x 254.00 x 114.30 KG: 3.38 LOCATION: COUNTERTOP SPEC COMMENTS: ; NOTES: 2025.05.12 ADDED UNIT PER SHD REVIEW 2 DOC. SHD STILL VERIFYING PREFERRED MODEL TBC CHANGE CODE: A CHANGE NOTES: 2025.05.12 ADDED UNIT PER SHD REVIEW 2 DOC VOLTS: 120 HERTZ: 60 AMPS: 7.5 PLUG TYPE: Type B (NEMA 5-15)	ARIZANT HEALTHCARE, A 3M COMPANY 245-00/245 RANGER	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
45	No image available	1023183	1	SERVICES TO INCLUDE: TO INCLUDE 12 EACH OF THE BELOW DATAMEDIATOR 1 EMR; DATAMEDIATOR CONNECTION DEVICE; (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE DATA NOTE: WIRED; LAN; ETHERNET PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL PROCARE	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
46	No image available	1023187	1	FREIGHT FREIGHT FOR LIGHTS BOOMS/INTEGRATION PROPOSAL: 10312261 V.1 (WDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL	OF/OI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID# EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
47	No image available	1023186	1	INSTALLATION ISUITE INSTALLATION PROPOSAL: 10312261 V.1 (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL ISUTE INSTALLATION	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
48	No image available	1023330	1	DISCOUNT DISCOUNT FOR LIGHTS BOOM AND INTEGRATION (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER COMMUNICATIONS DISCOUNT	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
49	No image available	1023181	1	ANCILLARY PARTS 7 MOUNTING CHANNEL FOR PIVOT; CPU CLAMP ADJUSTABLE 85-215MM WIDTH; FOLDABLE ARM SUPPORTS 16-22KG; POLE-MOUNTED ENDOSCOPE (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL ANCILLARY PARTS	OF/VI	NEW	GROUP: 3 MINOR MOVABLE
50		1023180	1	EQUIPMENT, SOFTWARE MEDIA ASSET MGMT;SECURELY SHARE, EDIT AND ARCHIVE MEDIA (WDXDH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00 MM: 0.00 x 0.00 x 0.00 KG: 0.00 LOCATION: SPACE PO REQ #: 944A-00009 - LIGHTS/BOOMS CLIENT PO #: SNCA1910	STRYKER MEDICAL DATA MEDIAOR	OF/VI	NEW	GROUP: 3 MINOR MOVABLE

FLOOR: 1
DEPARTMENT: SURGERY
ROOM NAME: OR
ROOM NUMBER: 636

LINE	IMAGE	ID#	EQUIP#	QTY	ITEM DESCRIPTION	MANUFACTURER	RESP	STATUS	
51	No image available	100004		1	DISCOUNT	ZZZ ADDITIONAL COST	OF/OI	NEW	GROUP: 3 MINOR MOVABLE
					(WXDXH) IN: 0.00 x 0.00 x 0.00 LBS: 0.00				
					MM: 0.00 x 0.00 x 0.00 KG: 0.00				
					LOCATION: SPACE				
					NOTES: DISCOUNT FOR STRYKER VIDEO SCOPE SYSTEM				
					PO REQ #: 944A-00006	SELLER: STRYKER ENDOSCOPY			
					CLIENT PO #: SNCA972				

END OF ROOM

END OF DEPARTMENT

END OF REPORT